FDLE
Florida Department of Law Enforcement

INSTRUMENT PROCESSING SHEET

Agency Sarasota County SO S/N 80-005076

Date In 05/28/2019 DI Completion Date 06/04/2019 ☑Ship ☐P/U ☐H/D ☐CMI ☐EE

Intake Performed By KP

- ☑ Annual
- ☐ Registration
- ☑ Return from CMI / EE

Visual Inspection:

☑ Case	☑ Handle
☑ Keyboard	☑ Dry Gas Shelf
☑ Feet	☑ Breath Tube
☑ Ports	☑ Screws Tight

Other Equipment/ Accessories:

☐ Power cord	☐ Printer Cable
☑ Static Bag	☐ 12V DC Cable

Notes: _____

Final Release Date

FDLE

JUN 0 4 2019

Alcohol Testing Program

Quality Checks Performed By JD

- ☑ Breath Tube Screen
- ☑ Replace External O-Rings
- ☑ Instrument Set Up Verified
- ☑ R-Value 239
- ☑ Flow Verification (L/s)

Flow Column # ATP-105

32 mm	.164	(.139 - .169)
36 mm	.175	(.156 - .190)
53 mm	.250	(.228 - .278)
103 mm	.531	(.447 - .547)

- ☑ Barometric Pressure Check

Gauge ID # 28427

- ☑ Stability Checks

Simulator	Serial #	Lot #/Exp
0.050	SD1018	201707D
		07/25/2019
0.080	SD3962	201707E
		07/25/2019
0.200	G2078	201707C
		07/24/2019
0.080 DGS	N/A	AG831804
		11/14/2020

Flow Calibration Performed By _____

Flow Column # _____

- ☐ 5L/min – 17mm
- ☐ 15L/min – 53mm
- ☐ 30L/min – 103mm
- ☐ R-Value _____
- ☐ Post Calibration Verification (L/s)

Flow Column # _____

32 mm		(.139 - .169)
36 mm		(.156 - .190)
53 mm		(.228 - .278)
103 mm		(.447 - .547)

Maintenance Performed By _____

- ☐ Battery Replacement
- ☐ Dry Gas Regulator Replacement
- ☐ Breath Tube Replacement
- ☐ Other _____

Temperature Checks Performed By JD

- ☑ Lab Temp °C 21.7

External Digital Therm. ID#: 300503

- ☑ 34°C +-.2 Serial #: SD1018
- ☑ 34°C +-.2 Serial #: SD3962
- ☑ 34°C +-.2 Serial #: G2078

Calibration Adjustment Performed By _____

Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

☐ Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

Department Inspection Performed By JA

Barometric Pressure ID# 28662

Gauge 1015 Instrument 1013

Mouth Alcohol Solution Lot # 2018-B

Acetone Stock Solution Lot # 2018-A

Simulator	Serial Number
0.000	G4444
Interferent	G6621
0.050	SD1018
0.080	SD3962
0.200	G2078

Attachments

☑ Form 41	☐ Post-Stability Checks
☑ Stability Checks	☐ Flow Calibration
☑ Calibration Certificate	☐ Form 40
☐ Calibration Adjustment	☐ Other _____

Notes/Suggested Service: Please change the level 2 password to something unique. JD

- ☑ **Instrument Complies with Chapter 11D-8, FAC**
- ☐ **Instrument Does Not Comply with Chapter 11D-8, FAC**
- ☑ **Return to/Place into Evidentiary Use**
- ☐ **Remain Out of Evidentiary Use**
- ☑ **Conduct an Agency Inspection Before Evidentiary Use**

SP 6/4/19 Brett Kirkland 6/4/19

Tech Review / Date Admin Review / Date

Florida Department of Law Enforcement
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: SARASOTA COUNTY SO
Time of Inspection: 10:25
Date of Inspection: 06/04/2019
Serial Number: 80-005076
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG831804 Exp: 11/14/2020
0.000	0.050	0.082	0.202	0.079
0.000	0.050	0.082	0.202	0.079
0.000	0.050	0.082	0.202	0.078
0.000	0.050	0.082	0.202	0.079
0.000	0.050	0.082	0.202	0.079
0.000	0.050	0.082	0.202	0.078
0.000	0.050	0.082	0.202	0.078
0.000	0.050	0.082	0.201	0.079
0.000	0.050	0.082	0.201	0.079
0.000	0.050	0.082	0.202	0.078

Standard Deviations	0.0000	0.0000	0.0004	0.0005

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

SP BK 6/4/19

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

TJ Graham — THOMAS J GRAHAM
Signature and Printed Name

06/04/2019
Date

FDLE/ATP Form 41 –Revised August 2005

80-005076
6/4/19
JD

SARASOTA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005076
06/04/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:27
Control Test	0.050	08:27
Air Blank	0.000	08:28
Control Test	0.050	08:28
Air Blank	0.000	08:29
Control Test	0.050	08:30
Air Blank	0.000	08:30
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

JD
Operator's Signature

SARASOTA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005076
06/04/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:31
Control Test	0.080	08:32
Air Blank	0.000	08:32
Control Test	0.080	08:33
Air Blank	0.000	08:34
Control Test	0.081	08:34
Air Blank	0.000	08:35
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

JD
Operator's Signature

SARASOTA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005076
06/04/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:36
Control Test	0.201	08:36
Air Blank	0.000	08:37
Control Test	0.200	08:38
Air Blank	0.000	08:38
Control Test	0.200	08:39
Air Blank	0.000	08:39
Control Test Stats		
Average	0.2003	
Std Dev	0.0006	
Rel Std Dev(%)	0.2882	

JD
Operator's Signature

SARASOTA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005076
06/04/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:42
Control Test	0.079	08:42
Air Blank	0.000	08:42
Control Test	0.080	08:43
Air Blank	0.000	08:43
Control Test	0.079	08:44
Air Blank	0.000	08:44
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	DGS

JD
Operator's Signature

SP BK 6/4/19

Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-005076 , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	80-005076
Owning Agency:	SARASOTA COUNTY SO
Calibration Date:	06/04/2019
Calibration Time:	10:25

	UNCERTAINTY* ±
0.050 g/ 210 L	0.004
0.080 g/ 210 L	0.004
0.200 g/ 210 L	0.007
0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

TRACEABILITY INFORMATION
This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

06/04/2019
Date

THOMAS J GRAHAM,
Department Inspector

SP BK 6/4/19

Return Material Authorization

Ship to: ☒ CMI, Inc.

☐ Enforcement Electronics

Shipment to repair facility authorized by: Mike Feltovic on 3/6/2019

Items Returned: Instrument ☑ Supplies ☐ Other ☐ Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-005076

Bill To Address:	Ship to Address:
Sarasota County SO	FL. Dept of Law Enforcement
	Alcohol Testing Program
	813 B Lake Bradford Road
	Tallahassee, FL 32304

Reason for Return:

Speaker not working.

Please choose one of the following options:

☐ 1. I _____, authorize all repairs.

☐ 2. I _____, authorize repairs up to $_____.

☑ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Mike Feltovic

Phone #: _____ Email: Michael.Feltovic@sarasotasheriff.or

ATP Contact Name: Shayla Platt ATP Email: shaylaplatt@fdle.state.fl.us

FDLE
Florida Department of Law Enforcement

INSTRUMENT PROCESSING SHEET

Agency Sarasota County SO S/N 80-005076

Date In 03/01/2019 DI Completion Date _____ ☐Ship ☐P/U ☐H/D ☒CMI ☐EE

Intake Performed By JG

- ☑ Annual
- ☐ Registration
- ☐ Return from CMI / EE

Visual Inspection:

- ☑ Case ☑ Handle
- ☑ Keyboard ☑ Dry Gas Shelf
- ☑ Feet ☑ Breath Tube
- ☑ Ports ☑ Screws Tight

Other Equipment/ Accessories:

- ☐ Power cord ☐ Printer Cable
- ☑ Static Bag ☐ 12V DC Cable

Notes: See note shipped with instrument.

Final Release Date

FDLE

JUN 04 2019

Alcohol Testing Program

Quality Checks Performed By _____

- ☐ Breath Tube Screen
- ☐ Replace External O-Rings
- ☐ Instrument Set Up Verified
- ☐ R-Value _____
- ☐ Flow Verification (L/s)

Flow Column # _____

32 mm _____ (.139 - .169)
36 mm _____ (.156 - .190)
53 mm _____ (.228 - .278)
103 mm _____ (.447 - .547)

- ☐ Barometric Pressure Check

Gauge ID # _____

- ☐ Stability Checks

Simulator	Serial #	Lot #/Exp
0.050		201707D
		07/25/2019
0.080		201707E
		07/25/2019
0.200		201707C
		07/24/2019
0.080 DGS	N/A	

Flow Calibration Performed By _____

Flow Column # _____

- ☐ 5L/min – 17mm
- ☐ 15L/min – 53mm
- ☐ 30L/min – 103mm
- ☐ R-Value _____
- ☐ Post Calibration Verification (L/s)

Flow Column # _____

32 mm _____ (.139 - .169)
36 mm _____ (.156 - .190)
53 mm _____ (.228 - .278)
103 mm _____ (.447 - .547)

Maintenance Performed By _____

- ☐ Battery Replacement
- ☐ Dry Gas Regulator Replacement
- ☐ Breath Tube Replacement
- ☐ Other _____

Temperature Checks Performed By _____

- ☐ Lab Temp °C _____

External Digital Therm. ID#: _____

- ☐ 34°C +-.2 Serial #: _____
- ☐ 34°C +-.2 Serial #: _____
- ☐ 34°C +-.2 Serial #: _____

Calibration Adjustment Performed By _____

Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

- ☐ Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

Department Inspection Performed By _____

Barometric Pressure ID# _____

Gauge _____ Instrument _____

Mouth Alcohol Solution Lot # _____

Acetone Stock Solution Lot # _____

Simulator	Serial Number
0.000	
Interferent	
0.050	
0.080	
0.200	

Attachments

- ☐ Form 41
- ☐ Stability Checks
- ☐ Calibration Certificate
- ☐ Calibration Adjustment
- ☐ Post-Stability Checks
- ☐ Flow Calibration
- ☐ Form 40
- ☐ Other _____

Notes/Suggested Service:

Sending to Repair for inoperable speaker. SP

- ☐ **Instrument Complies with Chapter 11D-8, FAC**
- ☐ **Instrument Does Not Comply with Chapter 11D-8, FAC**
- ☐ **Return to/Place into Evidentiary Use**
- ☐ **Remain Out of Evidentiary Use**
- ☐ **Conduct an Agency Inspection Before Evidentiary Use**

Tech Review / Date _____ Admin Review / Date _____

80-005076

Rec'd
3/1/19
JD

SPEAKER NOT WORKING

Return Material Authorization

Ship to: ☒ CMI, Inc.

☐ Enforcement Electronics

Shipment to repair facility authorized by: Mike Feltovic on 3/6/2019

Items Returned: Instrument ☑ Supplies ☐ Other ☐ Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-005076

Bill To Address:
Sarasota County SO

Ship to Address:
FL. Dept of Law Enforcement
Alcohol Testing Program
813 B Lake Bradford Road
Tallahassee, FL 32304

Reason for Return:
Speaker not working.

Please choose one of the following options:

☐ 1. I _____, authorize all repairs.

☐ 2. I _____, authorize repairs up to $_____.

☑ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Mike Feltovic

Phone #: _____ Email: Michael.Feltovic@sarasotasheriff.or

ATP Contact Name: Shayla Platt ATP Email: shaylaplatt@fdle.state.fl.us