



INSTRUMENT PROCESSING SHEET

Agency North Miami Police DepartmentS/N 80-005077

Florida Department of Law Enforcement

Date In 01/11/2019 DI Completion Date 01/14/2019 Ship P/U H/D CMI EE

Intake Performed By <u>Dee</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	Quality Checks Performed By <u>Dee</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>218</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP 106</u> 32 mm <u>.140</u> (.139 - .169) 36 mm <u>.156</u> (.156 - .190) 53 mm <u>.226</u> (.228 - .278) 103 mm <u>.500</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28663</u> <input checked="" type="checkbox"/> Stability Checks	Flow Calibration Performed By <u>Dee</u> Flow Column # <u>ATP104</u> <input checked="" type="checkbox"/> 5L/min - 17mm <input checked="" type="checkbox"/> 15L/min - 53mm <input checked="" type="checkbox"/> 30L/min - 103mm <input checked="" type="checkbox"/> R-Value <u>219</u> <input checked="" type="checkbox"/> Post Calibration Verification (L/s) Flow Column # <u>ATP106</u> 32 mm <u>.144</u> (.139 - .169) 36 mm <u>.164</u> (.156 - .190) 53 mm <u>.238</u> (.228 - .278) 103 mm <u>.500</u> (.447 - .547)															
Final Release Date <div style="text-align: center; font-size: 1.2em; font-weight: bold;">FDLE</div> <div style="text-align: center; font-size: 1.2em; font-weight: bold;">JAN 23 2019</div> <div style="text-align: center; font-weight: bold;">Alcohol Testing Program</div>	<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD3967</td> <td>201707D 07/25/2019</td> </tr> <tr> <td>0.080</td> <td>SD3968</td> <td>201707E 07/25/2019</td> </tr> <tr> <td>0.200</td> <td>SD3969</td> <td>201707C 07/24/2019</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG805701 02/26/2020</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	SD3967	201707D 07/25/2019	0.080	SD3968	201707E 07/25/2019	0.200	SD3969	201707C 07/24/2019	0.080 DGS	N/A	AG805701 02/26/2020	Maintenance Performed By <u>Dee</u> <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input checked="" type="checkbox"/> Other <u>Form load/Changed pass</u>
Simulator	Serial #	Lot #/Exp															
0.050	SD3967	201707D 07/25/2019															
0.080	SD3968	201707E 07/25/2019															
0.200	SD3969	201707C 07/24/2019															
0.080 DGS	N/A	AG805701 02/26/2020															
		Temperature Checks Performed By <u>Dee</u> <input checked="" type="checkbox"/> Lab Temp °C <u>21.30C</u> External Digital Therm. ID#: <u>300949</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD3967</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD3968</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD3969</u>															

Calibration Adjustment Performed By <u>Dee</u> Barometric Pressure Gauge <u>1021</u> ID # <u>28663</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>2235</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>MP4864</td> <td>17410</td> <td>12/06/2019</td> </tr> <tr> <td>0.100</td> <td>MP4863</td> <td>18070</td> <td>02/26/2020</td> </tr> <tr> <td>0.200</td> <td>SD3963</td> <td>17340</td> <td>10/09/2019</td> </tr> <tr> <td>0.300</td> <td>2108</td> <td>18110</td> <td>04/02/2020</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>34416080A2</td> <td>02/05/2019</td> </tr> </tbody> </table> <input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD3967</td> <td>201707D</td> <td>07/25/2019</td> </tr> <tr> <td>0.080</td> <td>SD3968</td> <td>201707E</td> <td>07/25/2019</td> </tr> <tr> <td>0.200</td> <td>SD3969</td> <td>201707C</td> <td>07/24/2019</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG805701</td> <td>02/26/2020</td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000	2235	N/A	N/A	0.040	MP4864	17410	12/06/2019	0.100	MP4863	18070	02/26/2020	0.200	SD3963	17340	10/09/2019	0.300	2108	18110	04/02/2020	0.080 DGS	N/A	34416080A2	02/05/2019	Simulator	Serial Number	Lot Number	Expiration	0.050	SD3967	201707D	07/25/2019	0.080	SD3968	201707E	07/25/2019	0.200	SD3969	201707C	07/24/2019	0.080 DGS	N/A	AG805701	02/26/2020	Department Inspection Performed By <u>Dee</u> Barometric Pressure ID# <u>68639</u> Gauge <u>1020</u> Instrument <u>1022</u> Mouth Alcohol Solution Lot # <u>2017-B</u> Acetone Stock Solution Lot # <u>2018-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>SD3965</td> </tr> <tr> <td>Interferent</td> <td>SD3966</td> </tr> <tr> <td>0.050</td> <td>SD3967</td> </tr> <tr> <td>0.080</td> <td>SD3968</td> </tr> <tr> <td>0.200</td> <td>SD3969</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	SD3965	Interferent	SD3966	0.050	SD3967	0.080	SD3968	0.200	SD3969
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000	2235	N/A	N/A																																																										
0.040	MP4864	17410	12/06/2019																																																										
0.100	MP4863	18070	02/26/2020																																																										
0.200	SD3963	17340	10/09/2019																																																										
0.300	2108	18110	04/02/2020																																																										
0.080 DGS	N/A	34416080A2	02/05/2019																																																										
Simulator	Serial Number	Lot Number	Expiration																																																										
0.050	SD3967	201707D	07/25/2019																																																										
0.080	SD3968	201707E	07/25/2019																																																										
0.200	SD3969	201707C	07/24/2019																																																										
0.080 DGS	N/A	AG805701	02/26/2020																																																										
Simulator	Serial Number																																																												
0.000	SD3965																																																												
Interferent	SD3966																																																												
0.050	SD3967																																																												
0.080	SD3968																																																												
0.200	SD3969																																																												
	Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input checked="" type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____																																																												

Notes/Suggested Service: <u>E-mailed</u> <input checked="" type="checkbox"/> APPROVED Performed optical calibration to adjust the barometric pressure. _____ _____ _____	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use <div style="margin-top: 10px;"> <u>Pgm 1/23/19</u> <u>JJ Deane 1/23/19</u> Tech Review / Date Admin Review / Date </div>
--	--

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: NORTH MIAMI PD
Time of Inspection: 09:01

Date of Inspection: 01/14/2019

Serial Number: 80-005077
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG805701 Exp: 02/26/2020
0.000	0.049	0.081	0.201	0.078
0.000	0.049	0.082	0.202	0.079
0.000	0.050	0.082	0.203	0.079
0.000	0.050	0.081	0.203	0.079
0.000	0.050	0.082	0.204	0.079
0.000	0.051	0.081	0.204	0.078
0.000	0.051	0.082	0.204	0.079
0.000	0.051	0.082	0.203	0.078
0.000	0.051	0.082	0.204	0.079
0.000	0.051	0.082	0.204	0.079

Standard Deviations	0.0008	0.0004	0.0010	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0006 Number of Simulators Used: 5

Remarks:

QJM

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David E Reyes Rivera DAVID E REYES-RIVERA
Signature and Printed Name

01/14/2019
Date

*1/23/19
JR*

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-005077	North Miami Police Department	01/14/2019	<i>Deer</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
SN: SD3967 Temp: 34.10c 0.047 to 0.053 <input checked="" type="checkbox"/>	SN: SD3968 Temp: 34.09c 0.077 to 0.083 <input checked="" type="checkbox"/>	SN: SD3969 Temp: 34.10c 0.194 to 0.206 <input checked="" type="checkbox"/>	Lot AG805701 0.077 to 0.083 <input checked="" type="checkbox"/>																																																																																																																																																
<p>NORTH MIAMI PD Intoxilyzer - Alcotest Analyzer Model: 8000 01/14/2019 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>06:13</td></tr> <tr><td>Control Test</td><td>0.050</td><td>06:14</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:14</td></tr> <tr><td>Control Test</td><td>0.050</td><td>06:15</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:16</td></tr> <tr><td>Control Test</td><td>0.050</td><td>06:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:17</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0500</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	06:13	Control Test	0.050	06:14	Air Blank	0.000	06:14	Control Test	0.050	06:15	Air Blank	0.000	06:16	Control Test	0.050	06:16	Air Blank	0.000	06:17	Control Test Stats			Average	0.0500		Std Dev	0.0000		Rel. Std Dev(%)	0.0000		<p>NORTH MIAMI PD Intoxilyzer - Alcotest Analyzer Model: 8000 01/14/2019 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>06:18</td></tr> <tr><td>Control Test</td><td>0.081</td><td>06:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:19</td></tr> <tr><td>Control Test</td><td>0.081</td><td>06:20</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:20</td></tr> <tr><td>Control Test</td><td>0.081</td><td>06:21</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:21</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0810</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	06:18	Control Test	0.081	06:18	Air Blank	0.000	06:19	Control Test	0.081	06:20	Air Blank	0.000	06:20	Control Test	0.081	06:21	Air Blank	0.000	06:21	Control Test Stats			Average	0.0810		Std Dev	0.0000		Rel. Std Dev(%)	0.0000		<p>NORTH MIAMI PD Intoxilyzer - Alcotest Analyzer Model: 8000 01/14/2019 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>06:22</td></tr> <tr><td>Control Test</td><td>0.198</td><td>06:23</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:24</td></tr> <tr><td>Control Test</td><td>0.199</td><td>06:24</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:25</td></tr> <tr><td>Control Test</td><td>0.200</td><td>06:25</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:26</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1990</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.5025</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	06:22	Control Test	0.198	06:23	Air Blank	0.000	06:24	Control Test	0.199	06:24	Air Blank	0.000	06:25	Control Test	0.200	06:25	Air Blank	0.000	06:26	Control Test Stats			Average	0.1990		Std Dev	0.0010		Rel. Std Dev(%)	0.5025		<p>NORTH MIAMI PD Intoxilyzer - Alcotest Analyzer Model: 8000 01/14/2019 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>06:27</td></tr> <tr><td>Control Test</td><td>0.079</td><td>06:27</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:28</td></tr> <tr><td>Control Test</td><td>0.079</td><td>06:28</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:29</td></tr> <tr><td>Control Test</td><td>0.079</td><td>06:29</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:30</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0790</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	06:27	Control Test	0.079	06:27	Air Blank	0.000	06:28	Control Test	0.079	06:28	Air Blank	0.000	06:29	Control Test	0.079	06:29	Air Blank	0.000	06:30	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel. Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:13																																																																																																																																																	
Control Test	0.050	06:14																																																																																																																																																	
Air Blank	0.000	06:14																																																																																																																																																	
Control Test	0.050	06:15																																																																																																																																																	
Air Blank	0.000	06:16																																																																																																																																																	
Control Test	0.050	06:16																																																																																																																																																	
Air Blank	0.000	06:17																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0500																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel. Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:18																																																																																																																																																	
Control Test	0.081	06:18																																																																																																																																																	
Air Blank	0.000	06:19																																																																																																																																																	
Control Test	0.081	06:20																																																																																																																																																	
Air Blank	0.000	06:20																																																																																																																																																	
Control Test	0.081	06:21																																																																																																																																																	
Air Blank	0.000	06:21																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0810																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel. Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:22																																																																																																																																																	
Control Test	0.198	06:23																																																																																																																																																	
Air Blank	0.000	06:24																																																																																																																																																	
Control Test	0.199	06:24																																																																																																																																																	
Air Blank	0.000	06:25																																																																																																																																																	
Control Test	0.200	06:25																																																																																																																																																	
Air Blank	0.000	06:26																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1990																																																																																																																																																		
Std Dev	0.0010																																																																																																																																																		
Rel. Std Dev(%)	0.5025																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:27																																																																																																																																																	
Control Test	0.079	06:27																																																																																																																																																	
Air Blank	0.000	06:28																																																																																																																																																	
Control Test	0.079	06:28																																																																																																																																																	
Air Blank	0.000	06:29																																																																																																																																																	
Control Test	0.079	06:29																																																																																																																																																	
Air Blank	0.000	06:30																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0790																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel. Std Dev(%)	0.0000																																																																																																																																																		
<p><i>Deer</i></p> <p>Operator's Signature</p> <p><i>6/12/19</i></p>	<p><i>Deer</i></p> <p>Operator's Signature</p>	<p><i>Deer</i></p> <p>Operator's Signature</p>	<p><i>Deer</i></p> <p>Operator's Signature</p>																																																																																																																																																



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-005077, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-005077</u>	UNCERTAINTY* ±	
Owning Agency:	<u>NORTH MIAMI PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>01/14/2019</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>09:01</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

01/14/2019

Date

David Reyes-Rivera
DAVID E REYES-RIVERA,

Department Inspector

FDLE/ATP Form 69 July 2018

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

1/23/19

189M

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.5640 (-0.0040)
Sample #2 = 1.5490 (0.0060)
Sample #3 = 1.5830 (0.0000)
Sample #4 = 1.5730 (0.0170)
Avg % Abs = 1.5683 (0.0077)
STD DEV = 0.0175 (0.0086)
REL STD DEV = 1.114 (112.457)

NORTH MIAMI PD
Intoxilyzer - Alconol Analyzer
Model 8000
SN 80-005977
06:31:38
01/14/2019

Auto Calibration

Max Power Res Value = 36
Auto Range Res Value = 25

Sol Value = 0.000 g/210L ***
Fit Value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12660, Sum Io = 13193
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.8880 (-0.0160)
Sample #2 = 1.9190 (0.0000)
Sample #3 = 1.9050 (0.0060)
Sample #4 = 1.9030 (0.0790)
Avg % Abs = 1.9030 (0.0283)
STD DEV = 0.0087 (0.0440)
REL STD DEV = 0.457 (155.227)

Sol Value = 0.000 g/210L ***
Fit Value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12678, Sum Io = 13199
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.0330 (-0.0100)
Sample #2 = 0.0270 (0.0200)
Sample #3 = 0.0420 (0.0360)
Sample #4 = 0.0050 (0.0650)
Avg % Abs = 0.0247 (0.0403)
STD DEV = 0.0186 (0.0228)
REL STD DEV = 75.446 (56.556)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.6260 (-0.0110)
Sample #2 = 3.6590 (0.0020)
Sample #3 = 3.6540 (0.0010)
Sample #4 = 3.6450 (0.0330)
Avg % Abs = 3.6527 (0.0120)
STD DEV = 0.0071 (0.0182)
REL STD DEV = 0.194 (151.612)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.0970 (-0.0140)
Sample #2 = 0.0810 (-0.0040)
Sample #3 = 0.0990 (-0.0010)
Sample #4 = 0.0800 (0.0000)
Avg % Abs = 0.0867 (-0.0017)
STD DEV = 0.0107 (0.0021)
REL STD DEV = 12.338 (124.900)

Sol Value = 0.200 g/210L ***
Fit Value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12646, Sum Io = 13186
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.6100 (0.0000)
Sample #2 = 3.6590 (-0.0060)
Sample #3 = 3.6690 (0.0000)
Sample #4 = 3.6350 (0.0260)
Avg % Abs = 3.6543 (0.0067)
STD DEV = 0.0175 (0.0170)
REL STD DEV = 0.478 (255.147)

Sol Value = 0.040 g/210L ***
Fit Value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12668, Sum Io = 13197
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.7950 (-0.0080)
Sample #2 = 0.7630 (0.0280)
Sample #3 = 0.8210 (0.0150)
Sample #4 = 0.7950 (0.0430)
Avg % Abs = 0.7897 (0.0287)
STD DEV = 0.0293 (0.0140)
REL STD DEV = 3.708 (48.879)

***** AUTO CAL DATA *****
<<<<< CHANNEL 1 >>>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.025
Std Dev = 0.02 Rel Std Dev = 75.45
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.790
Std Dev = 0.03 Rel Std Dev = 3.71
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.909
Std Dev = 0.01 Rel Std Dev = 0.46
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.654
Std Dev = 0.02 Rel Std Dev = 0.48
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.310
Std Dev = 0.02 Rel Std Dev = 0.29
Zero Order Coef = -59.14
First Order Coef = 2437.74
Second Order Coef = 49.74
Standard Deviation = 9.651946

<<<<< CHANNEL 2 >>>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.087
Std Dev = 0.01 Rel Std Dev = 12.34
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.568
Std Dev = 0.02 Rel Std Dev = 1.11
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.653
Std Dev = 0.01 Rel Std Dev = 0.19
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.878
Std Dev = 0.02 Rel Std Dev = 0.23
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.852
Std Dev = 0.02 Rel Std Dev = 0.21
Zero Order Coef = -113.84
First Order Coef = 1259.61
Second Order Coef = 20.51
Standard Deviation = 4.698000

***** Solution Stats Quadratic Fit Chan 2 *****
Act Fit Residual
g/210L g/210L g/210L
0.000 -0.000 0.0001
0.048 0.040 -0.0002
0.100 0.100 0.0000
0.200 0.200 0.0001
0.300 0.300 -0.0000

Sol Value = 0.080 g/210L ***
Fit Value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1
***** CHANNEL 1 *****
Sample #1 = 3530.00
Sample #2 = 3538.00
Sample #3 = 3559.00
Sample #4 = 3528.00
Average Result = 3541.6667
STD DEV = 15.8219
REL STD DEV = 0.447

***** CHANNEL 2 *****
Sample #1 = 3374.00
Sample #2 = 3401.00
Sample #3 = 3415.00
Sample #4 = 3419.00
Average Result = 3411.6667
STD DEV = 9.4516
REL STD DEV = 0.277

***** CHANNEL 1 *****
Dry Gas H2O Adjust Results *****
Barometric Pressure = 1021
3 um H2O Adjust (mg/l*10,000) = 268
9 um H2O Adjust (mg/l*10,000) = 398
***** AUTO CAL PASS *****

Optical Calibration	
SN: 80-005077	Residual
Agency: North Miami P.D.	g/210L
Date: 01/14/2019	0.000 -0.0000
Quadratic Fit: +/-0.002g/210L	0.040 0.0032
By: <i>DR</i>	0.100 -0.0003
	0.200 0.0002
	0.300 -0.0001

***** Solution Stats Quadratic Fit Chan 1 *****
Act Fit Residual
g/210L g/210L
0.000 0.000 -0.0000
0.040 0.040 0.0032
0.100 0.100 -0.0003
0.200 0.200 0.0002
0.300 0.300 -0.0001

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Post Stabilities	80-005077	North Miami Police Department	01/14/2019	<i>Dee</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
SN: SD3967 Temp: 34.10c 0.047 to 0.053 <input checked="" type="checkbox"/>	SN: SD3968 Temp: 34.09c 0.077 to 0.083 <input checked="" type="checkbox"/>	SN: SD3969 Temp: 34.10c 0.194 to 0.206 <input checked="" type="checkbox"/>	Lot AG805701 0.077 to 0.083 <input checked="" type="checkbox"/>																																																																																																																																																
NORTH MIAMI PD Intoxilyzer - Alconol Analyzer Model: 8000 SN 80-005077 01/14/2019 Software: 8100.27	NORTH MIAMI PD Intoxilyzer - Alconol Analyzer Model: 8000 SN 80-005077 01/14/2019 Software: 8100.27	NORTH MIAMI PD Intoxilyzer - Alconol Analyzer Model: 8000 SN 80-005077 01/14/2019 Software: 8100.27	NORTH MIAMI PD Intoxilyzer - Alconol Analyzer Model: 8000 SN 80-005077 01/14/2019 Software: 8100.27																																																																																																																																																
<table border="1"> <tr><th>Test</th><th>g/210L</th><th>Time</th></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:11</td></tr> <tr><td>Control Test</td><td>0.049</td><td>07:12</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:13</td></tr> <tr><td>Control Test</td><td>0.050</td><td>07:13</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:14</td></tr> <tr><td>Control Test</td><td>0.050</td><td>07:14</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:15</td></tr> <tr><td>Control Test</td><td>0.000</td><td>07:15</td></tr> <tr><td>Average</td><td>0.0497</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>1.1625</td><td></td></tr> </table>	Test	g/210L	Time	Air Blank	0.000	07:11	Control Test	0.049	07:12	Air Blank	0.000	07:13	Control Test	0.050	07:13	Air Blank	0.000	07:14	Control Test	0.050	07:14	Air Blank	0.000	07:15	Control Test	0.000	07:15	Average	0.0497		Std Dev	0.0006		Rel. Std Dev(%)	1.1625		<table border="1"> <tr><th>Test</th><th>g/210L</th><th>Time</th></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:16</td></tr> <tr><td>Control Test</td><td>0.081</td><td>07:17</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:17</td></tr> <tr><td>Control Test</td><td>0.081</td><td>07:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:18</td></tr> <tr><td>Control Test</td><td>0.081</td><td>07:19</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:20</td></tr> <tr><td>Control Test</td><td>0.000</td><td>07:20</td></tr> <tr><td>Average</td><td>0.0810</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.0000</td><td></td></tr> </table>	Test	g/210L	Time	Air Blank	0.000	07:16	Control Test	0.081	07:17	Air Blank	0.000	07:17	Control Test	0.081	07:18	Air Blank	0.000	07:18	Control Test	0.081	07:19	Air Blank	0.000	07:20	Control Test	0.000	07:20	Average	0.0810		Std Dev	0.0000		Rel. Std Dev(%)	0.0000		<table border="1"> <tr><th>Test</th><th>g/210L</th><th>Time</th></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:21</td></tr> <tr><td>Control Test</td><td>0.202</td><td>07:21</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:22</td></tr> <tr><td>Control Test</td><td>0.203</td><td>07:22</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:23</td></tr> <tr><td>Control Test</td><td>0.203</td><td>07:24</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:24</td></tr> <tr><td>Control Test</td><td>0.000</td><td>07:24</td></tr> <tr><td>Average</td><td>0.2027</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.2849</td><td></td></tr> </table>	Test	g/210L	Time	Air Blank	0.000	07:21	Control Test	0.202	07:21	Air Blank	0.000	07:22	Control Test	0.203	07:22	Air Blank	0.000	07:23	Control Test	0.203	07:24	Air Blank	0.000	07:24	Control Test	0.000	07:24	Average	0.2027		Std Dev	0.0006		Rel. Std Dev(%)	0.2849		<table border="1"> <tr><th>Test</th><th>g/210L</th><th>Time</th></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:26</td></tr> <tr><td>Control Test</td><td>0.079</td><td>07:27</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:27</td></tr> <tr><td>Control Test</td><td>0.079</td><td>07:28</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:28</td></tr> <tr><td>Control Test</td><td>0.078</td><td>07:28</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:29</td></tr> <tr><td>Control Test</td><td>0.000</td><td>07:29</td></tr> <tr><td>Average</td><td>0.0787</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.7359</td><td></td></tr> </table>	Test	g/210L	Time	Air Blank	0.000	07:26	Control Test	0.079	07:27	Air Blank	0.000	07:27	Control Test	0.079	07:28	Air Blank	0.000	07:28	Control Test	0.078	07:28	Air Blank	0.000	07:29	Control Test	0.000	07:29	Average	0.0787		Std Dev	0.0006		Rel. Std Dev(%)	0.7359	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	07:11																																																																																																																																																	
Control Test	0.049	07:12																																																																																																																																																	
Air Blank	0.000	07:13																																																																																																																																																	
Control Test	0.050	07:13																																																																																																																																																	
Air Blank	0.000	07:14																																																																																																																																																	
Control Test	0.050	07:14																																																																																																																																																	
Air Blank	0.000	07:15																																																																																																																																																	
Control Test	0.000	07:15																																																																																																																																																	
Average	0.0497																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel. Std Dev(%)	1.1625																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	07:16																																																																																																																																																	
Control Test	0.081	07:17																																																																																																																																																	
Air Blank	0.000	07:17																																																																																																																																																	
Control Test	0.081	07:18																																																																																																																																																	
Air Blank	0.000	07:18																																																																																																																																																	
Control Test	0.081	07:19																																																																																																																																																	
Air Blank	0.000	07:20																																																																																																																																																	
Control Test	0.000	07:20																																																																																																																																																	
Average	0.0810																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel. Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	07:21																																																																																																																																																	
Control Test	0.202	07:21																																																																																																																																																	
Air Blank	0.000	07:22																																																																																																																																																	
Control Test	0.203	07:22																																																																																																																																																	
Air Blank	0.000	07:23																																																																																																																																																	
Control Test	0.203	07:24																																																																																																																																																	
Air Blank	0.000	07:24																																																																																																																																																	
Control Test	0.000	07:24																																																																																																																																																	
Average	0.2027																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel. Std Dev(%)	0.2849																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	07:26																																																																																																																																																	
Control Test	0.079	07:27																																																																																																																																																	
Air Blank	0.000	07:27																																																																																																																																																	
Control Test	0.079	07:28																																																																																																																																																	
Air Blank	0.000	07:28																																																																																																																																																	
Control Test	0.078	07:28																																																																																																																																																	
Air Blank	0.000	07:29																																																																																																																																																	
Control Test	0.000	07:29																																																																																																																																																	
Average	0.0787																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel. Std Dev(%)	0.7359																																																																																																																																																		
<i>Dee</i> Operator's Signature	<i>Dee</i> Operator's Signature	<i>Dee</i> Operator's Signature	<i>Dee</i> Operator's Signature																																																																																																																																																

Dee

1/23/19
JD

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Flow calibration	80-005077	North Miami Police Department	01/14/2019	JEK

NORTH MIAMI PD
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-005077
 01/14/2019
 Software: 8100.27

Flow Rate Calibration*****
 1: Rate (Liters/min) = 5
 SORT(Diff)) = 7.000
 2: Rate (Liters/min) = 15
 SORT(Diff)) = 11.617
 3: Rate (Liters/min) = 30
 SORT(Diff)) = 20.809

Dependent Data Scale Factor = 100000 L/min
 Independent Data Scale Factor = 256
 Rounded Slope = 697
 Rounded Intercept = -579367
 Correlation = 0.99726

1/23/19
 JEK
 Room