



INSTRUMENT PROCESSING SHEET

Agency Pinellas County SO

S/N 80-005290

Florida Department of Law Enforcement

Date In 02/08/2019 DI Completion Date 02/14/2019

Ship P/U H/D CMI EE

Intake Performed By <u>JD</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	Quality Checks Performed By <u>JD</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value 120 <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # ATP-105 32 mm .148 (.139 - .169) 36 mm .167 (.156 - .190) 53 mm .238 (.228 - .278) 103 mm .507 (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # 28421 <input checked="" type="checkbox"/> Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD1012</td> <td>201707D 07/25/2019</td> </tr> <tr> <td>0.080</td> <td>DR1279</td> <td>201707E 07/25/2019</td> </tr> <tr> <td>0.200</td> <td>DR3856</td> <td>201707C 07/24/2019</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG805701 02/26/2020</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	SD1012	201707D 07/25/2019	0.080	DR1279	201707E 07/25/2019	0.200	DR3856	201707C 07/24/2019	0.080 DGS	N/A	AG805701 02/26/2020	Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Temperature Checks Performed By <u>JD</u> <input checked="" type="checkbox"/> Lab Temp °C 21.0 External Digital Therm. ID#: 300503 <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: SD1012 <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: DR1279 <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: DR3856
Simulator	Serial #	Lot #/Exp															
0.050	SD1012	201707D 07/25/2019															
0.080	DR1279	201707E 07/25/2019															
0.200	DR3856	201707C 07/24/2019															
0.080 DGS	N/A	AG805701 02/26/2020															
Final Release Date FDLE FEB 14 2019 Alcohol Testing Program																	

Calibration Adjustment Performed By <u>JD</u> Barometric Pressure Gauge 1019 ID # 28427 <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>G2834</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>SD1022</td> <td>17410</td> <td>12/06/2019</td> </tr> <tr> <td>0.100</td> <td>SD3964</td> <td>18070</td> <td>02/26/2020</td> </tr> <tr> <td>0.200</td> <td>SD1025</td> <td>17340</td> <td>10/09/2019</td> </tr> <tr> <td>0.300</td> <td>SD1024</td> <td>18110</td> <td>04/02/2020</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>17817080A2</td> <td>08/05/2019</td> </tr> </tbody> </table> <input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD1012</td> <td>201707D</td> <td>07/25/2019</td> </tr> <tr> <td>0.080</td> <td>DR1279</td> <td>201707E</td> <td>07/25/2019</td> </tr> <tr> <td>0.200</td> <td>DR3856</td> <td>201707C</td> <td>07/24/2019</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG805701</td> <td>02/26/2020</td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000	G2834	N/A	N/A	0.040	SD1022	17410	12/06/2019	0.100	SD3964	18070	02/26/2020	0.200	SD1025	17340	10/09/2019	0.300	SD1024	18110	04/02/2020	0.080 DGS	N/A	17817080A2	08/05/2019	Simulator	Serial Number	Lot Number	Expiration	0.050	SD1012	201707D	07/25/2019	0.080	DR1279	201707E	07/25/2019	0.200	DR3856	201707C	07/24/2019	0.080 DGS	N/A	AG805701	02/26/2020	Department Inspection Performed By <u>JD</u> Barometric Pressure ID# 28662 Gauge 1021 Instrument 1020 Mouth Alcohol Solution Lot # 2018-B Acetone Stock Solution Lot # 2018-A <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>G2408</td> </tr> <tr> <td>Interferent</td> <td>G2882</td> </tr> <tr> <td>0.050</td> <td>SD1012</td> </tr> <tr> <td>0.080</td> <td>DR1279</td> </tr> <tr> <td>0.200</td> <td>DR3856</td> </tr> </tbody> </table> Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input checked="" type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____	Simulator	Serial Number	0.000	G2408	Interferent	G2882	0.050	SD1012	0.080	DR1279	0.200	DR3856
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000	G2834	N/A	N/A																																																										
0.040	SD1022	17410	12/06/2019																																																										
0.100	SD3964	18070	02/26/2020																																																										
0.200	SD1025	17340	10/09/2019																																																										
0.300	SD1024	18110	04/02/2020																																																										
0.080 DGS	N/A	17817080A2	08/05/2019																																																										
Simulator	Serial Number	Lot Number	Expiration																																																										
0.050	SD1012	201707D	07/25/2019																																																										
0.080	DR1279	201707E	07/25/2019																																																										
0.200	DR3856	201707C	07/24/2019																																																										
0.080 DGS	N/A	AG805701	02/26/2020																																																										
Simulator	Serial Number																																																												
0.000	G2408																																																												
Interferent	G2882																																																												
0.050	SD1012																																																												
0.080	DR1279																																																												
0.200	DR3856																																																												

Notes/Suggested Service: Calibration Adjustment needed due to barometer being outside of 1% (instrument 1003 / barometer 1021) and 0.08 g/210 L ARS / DGS agreement. <u>JD</u> _____ _____ _____	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use <u>SP 2/14/19</u> <u>Beth Kirkland 2/14/19</u> Tech Review / Date Admin Review / Date
---	---

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PINELLAS COUNTY SO
Time of Inspection: 10:21

Date of Inspection: 02/14/2019

Serial Number: 80-005290
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG805701 Exp: 02/26/2020
0.000	0.047	0.080	0.199	0.079
0.000	0.047	0.080	0.200	0.079
0.000	0.047	0.079	0.199	0.079
0.000	0.047	0.080	0.198	0.079
0.000	0.047	0.080	0.199	0.079
0.000	0.047	0.080	0.198	0.079
0.000	0.047	0.080	0.198	0.079
0.000	0.047	0.080	0.198	0.079
0.000	0.047	0.080	0.198	0.078
0.000	0.047	0.080	0.197	0.078

Standard Deviations	0.0000	0.0003	0.0008	0.0004
---------------------	--------	--------	--------	--------


Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

SP BK
2/14/19

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



 Signature and Printed Name THOMAS J GRAHAM

02/14/2019
 Date

80-005290

2/8/19
JD

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005290
02/08/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:49
Control Test	0.050	12:49
Air Blank	0.000	12:50
Control Test	0.049	12:50
Air Blank	0.000	12:51
Control Test	0.049	12:52
Air Blank	0.000	12:52
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel. Std Dev(%)	1.1703	

JD

Operator's Signature

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005290
02/08/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:54
Control Test	0.078	12:55
Air Blank	0.000	12:55
Control Test	0.078	12:56
Air Blank	0.000	12:56
Control Test	0.077	12:57
Air Blank	0.000	12:58
Control Test Stats		
Average	0.0777	
Std Dev	0.0006	
Rel. Std Dev(%)	0.7434	

JD

Operator's Signature

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005290
02/08/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:59
Control Test	0.195	13:00
Air Blank	0.000	13:00
Control Test	0.195	13:01
Air Blank	0.000	13:02
Control Test	0.194	13:02
Air Blank	0.000	13:03
Control Test Stats		
Average	0.1947	
Std Dev	0.0006	
Rel. Std Dev(%)	0.2966	

JD

Operator's Signature

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005290
02/08/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:45
Control Test	0.082	12:45
Air Blank	0.000	12:45
Control Test	0.081	12:46
Air Blank	0.000	12:46
Control Test	0.080	12:47
Air Blank	0.000	12:47
Control Test Stats		
Average	0.0810	
Std Dev	0.0010	
Rel. Std Dev(%)	1.2346	

DCS

8P BK
2/14/19

JD

Operator's Signature



ANAB
ACCREDITED
FORENSIC CALIBRATION
LABORATORY

Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-005290, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-005290</u>	UNCERTAINTY* ±
Owning Agency:	<u>PINELLAS COUNTY SO</u>	0.050 g/ 210 L
Calibration Date:	<u>02/14/2019</u>	0.080 g/ 210 L
Calibration Time:	<u>10:21</u>	0.200 g/ 210 L
		0.080 g/ 210 L Dry Gas Control
		0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.


THOMAS J GRAHAM,
 Department Inspector

02/14/2019

Date

SP

BK
2/14/19

80-005290

Calibration Adjustment

<<<<< CHANNEL 2 >>>>>

Sample	% Abs	(% Abs Ref)
Sample #1	1.5850	(-0.0040)
Sample #2	1.5080	(0.0030)
Sample #3	1.5570	(0.0020)
Sample #4	1.6010	(0.0200)
Avg % Abs	1.6020	(0.0083)
STD DEV	0.0055	(0.0101)
REL STD DEV	0.348	(121.392)

Sol Value = 0.100 g/210L ***
 Fit Value = 0.4762 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12732, Sum Io = 13436

<<<<< CHANNEL 1 >>>>>

Sample	% Abs	(% Abs Ref)
Sample #1	1.9110	(-0.0240)
Sample #2	1.9060	(0.0020)
Sample #3	1.8510	(0.0410)
Sample #4	1.8510	(0.0540)
Avg % Abs	1.8683	(0.0323)
STD DEV	0.0318	(0.0271)
REL STD DEV	1.699	(83.696)

<<<<< CHANNEL 2 >>>>>

Sample	% Abs	(% Abs Ref)
Sample #1	3.7010	(-0.0250)
Sample #2	3.6790	(0.0070)
Sample #3	3.6760	(0.0020)
Sample #4	3.6820	(0.0060)
Avg % Abs	3.6790	(0.0050)
STD DEV	0.0030	(0.0026)
REL STD DEV	0.082	(32.915)

Sol Value = 0.200 g/210L ***
 Fit Value = 0.9524 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12723, Sum Io = 13433

<<<<< CHANNEL 1 >>>>>

Sample	% Abs	(% Abs Ref)
Sample #1	3.5480	(-0.0040)
Sample #2	3.5600	(0.0160)
Sample #3	3.5870	(0.0120)
Sample #4	3.6230	(0.0160)
Avg % Abs	3.5900	(0.0147)
STD DEV	0.0315	(0.0023)
REL STD DEV	0.880	(15.746)

<<<<< CHANNEL 2 >>>>>

Sample	% Abs	(% Abs Ref)
Sample #1	0.1960	(0.0040)
Sample #2	0.2080	(0.0090)
Sample #3	0.2060	(0.0190)
Sample #4	0.2200	(0.0250)
Avg % Abs	0.2113	(0.0177)
STD DEV	0.0076	(0.0081)
REL STD DEV	3.583	(45.752)

Sol Value = 0.040 g/210L ***
 Fit Value = 0.1905 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12739, Sum Io = 13441

<<<<< CHANNEL 1 >>>>>

Sample	% Abs	(% Abs Ref)
Sample #1	0.7830	(-0.0220)
Sample #2	0.8260	(-0.0200)
Sample #3	0.7920	(0.0000)
Sample #4	0.7950	(0.0440)
Avg % Abs	0.8050	(0.0180)
STD DEV	0.0200	(0.0227)
REL STD DEV	2.481	(409.268)

PINELLAS COUNTY, SO
 Intoxilyzer - Alconal Analyzer
 Model 8000
 02/14/2019
 SN 80-005290
 07:31:06

Auto Calibration
 Max Power Res Value = 45
 Auto Range Res Value = 28

Sol Value = 0.000 g/210L ***
 Fit Value = 0.0000 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12762, Sum Io = 13448

<<<<< CHANNEL 1 >>>>>

Sample	% Abs	(% Abs Ref)
Sample #1	0.0700	(0.0000)
Sample #2	0.0570	(0.0520)
Sample #3	0.1010	(0.0710)
Sample #4	0.1110	(0.1000)
Avg % Abs	0.0897	(0.0743)
STD DEV	0.0287	(0.0242)
REL STD DEV	32.039	(32.520)

***** AUTO CAL DATA *****
 <<<<< CHANNEL 1 >>>>>

Sol Val	= 0.0000 mg/l or 0.000 g/210L
% Abs	= 0.090
Std Dev	= 0.03 Rel Std Dev = 32.04
Sol Val	= 0.1905 mg/l or 0.040 g/210L
% Abs	= 0.805
Std Dev	= 0.02 Rel Std Dev = 2.48
Sol Val	= 0.4762 mg/l or 0.100 g/210L
% Abs	= 1.869
Std Dev	= 0.03 Rel Std Dev = 1.70
Sol Val	= 0.9524 mg/l or 0.200 g/210L
% Abs	= 3.590
Std Dev	= 0.03 Rel Std Dev = 0.88
Sol Val	= 1.4286 mg/l or 0.300 g/210L
% Abs	= 5.272
Std Dev	= 0.00 Rel Std Dev = 0.09
Zero Order Coef	= -236.10
First Order Coef	= 2635.72
Second Order Coef	= 22.63
Standard Deviation	= 5.587340

<<<<< CHANNEL 2 >>>>>

Sample	% Abs	(% Abs Ref)
Sample #1	6.8080	(-0.0020)
Sample #2	6.8570	(0.0030)
Sample #3	6.9030	(0.0050)
Sample #4	6.9030	(0.0120)
Avg % Abs	6.8877	(0.0067)
STD DEV	0.0266	(0.0047)
REL STD DEV	0.386	(70.887)

Sol Value = 0.300 g/210L ***
 Fit Value = 1.4286 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12721, Sum Io = 13433

<<<<< CHANNEL 1 >>>>>

Sample	% Abs	(% Abs Ref)
Sample #1	5.2860	(-0.0040)
Sample #2	5.2740	(0.0470)
Sample #3	5.2660	(0.0550)
Sample #4	5.2750	(0.0620)
Avg % Abs	5.2717	(0.0547)
STD DEV	0.0049	(0.0075)
REL STD DEV	0.094	(13.730)

<<<<< CHANNEL 2 >>>>>

Sol Val	= 0.0000 mg/l or 0.000 g/210L
% Abs	= 0.211
Std Dev	= 0.01 Rel Std Dev = 3.58
Sol Val	= 0.1905 mg/l or 0.040 g/210L
% Abs	= 1.602
Std Dev	= 0.01 Rel Std Dev = 0.35
Sol Val	= 0.4762 mg/l or 0.100 g/210L
% Abs	= 3.679
Std Dev	= 0.00 Rel Std Dev = 0.08
Sol Val	= 0.9524 mg/l or 0.200 g/210L
% Abs	= 6.888
Std Dev	= 0.03 Rel Std Dev = 0.39
Sol Val	= 1.4286 mg/l or 0.300 g/210L
% Abs	= 9.986
Std Dev	= 0.00 Rel Std Dev = 0.00
Zero Order Coef	= -280.04
First Order Coef	= 1332.40
Second Order Coef	= 12.70
Standard Deviation	= 22.145943

Solution Stats Quadratic Fit Chan 1

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0000
0.040	0.040	0.0001
0.100	0.100	-0.0002
0.200	0.200	0.0001
0.300	0.300	-0.0000

Solution Stats Quadratic Fit Chan 2

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0000
0.040	0.040	0.0004
0.100	0.101	-0.0007
0.200	0.199	0.0005
0.300	0.300	-0.0001

Sol Value = 0.080 g/210L ***
 Fit Value = 0.3810 mg/l %%%
 Samples Taken = 4, Discarded = 1

***** CHANNEL 1 *****

Sample #1	= 3422.00
Sample #2	= 3352.00
Sample #3	= 3382.00
Sample #4	= 3393.00
Average Result	= 3375.6667
STD DEV	= 21.2211
REL STD DEV	= 0.629

***** CHANNEL 2 *****

Sample #1	= 3298.00
Sample #2	= 3280.00
Sample #3	= 3204.00
Sample #4	= 3311.00
Average Result	= 3296.3333
STD DEV	= 16.2583
REL STD DEV	= 0.493

***** CHANNEL 1 *****

Dry Gas H2O Adjust Results *****
 Barometric Pressure = 1019
 3 um H2O Adjust (mg/l*10.000) = 494
 9 um H2O Adjust (mg/l*10.000) = 511
 ***** AUTO CAL PASS *****

SP BK
 2/14/19

80-005290

2/14/19 JD

Post Calibration Adjustment Stability Checks.

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005290
02/14/2019
Software: 8105.27

Test	g/210L	Time
Air Blank	0.000	08:13
Control Test	0.050	08:13
Air Blank	0.000	08:20
Control Test	0.050	08:21
Air Blank	0.000	08:21
Control Test	0.049	08:22
Air Blank	0.000	08:22
Control Test Stats		
Average	0.0497	
Std Dev	0.0006	
Rel Std Dev(%)	1.1625	

JD

Operator's Signature

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005290
02/14/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:14
Control Test	0.061	08:15
Air Blank	0.000	08:15
Control Test	0.061	08:16
Air Blank	0.000	08:16
Control Test	0.060	08:17
Air Blank	0.000	08:18
Control Test Stats		
Average	0.0607	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

JD

Operator's Signature

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005290
02/14/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:28
Control Test	0.200	08:29
Air Blank	0.000	08:29
Control Test	0.199	08:30
Air Blank	0.000	08:30
Control Test	0.198	08:31
Air Blank	0.000	08:32
Control Test Stats		
Average	0.1990	
Std Dev	0.0010	
Rel Std Dev(%)	0.5025	

JD

Operator's Signature

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005290
02/14/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:24
Control Test	0.079	08:24
Air Blank	0.000	08:25
Control Test	0.079	08:25
Air Blank	0.000	08:25
Control Test	0.079	08:26
Air Blank	0.000	08:26
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

JD

Operator's Signature

SP BK 2/14/19