



INSTRUMENT PROCESSING SHEET

Agency Ocala PD

S/N 80-005454

Florida Department of Law Enforcement

Date In 01/07/19

DI Completion Date 1/9/19

Ship  P/U  H/D  CMI  EE

<b>Intake</b> Performed By <u>SQC</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____		<b>Quality Checks</b> Performed By <u>JD</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>175</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP-105</u> 32 mm <u>.156</u> (.139 - .169) 36 mm <u>.175</u> (.156 - .190) 53 mm <u>.253</u> (.228 - .278) 103 mm <u>.527</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28421</u> <input checked="" type="checkbox"/> Stability Checks		<b>Flow Calibration</b> Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																							
<b>Final Release Date</b> <b>FDLE</b> <b>JAN 09 2019</b> <b>Alcohol Testing Program</b>		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD1012</td> <td>201707 D 7/25/19</td> </tr> <tr> <td>0.080</td> <td>DR1279</td> <td>201707 E 7/25/19</td> </tr> <tr> <td>0.200</td> <td>DR3856</td> <td>201707 C 7/24/19</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG805701 2/26/20</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #/Exp	0.050	SD1012	201707 D 7/25/19	0.080	DR1279	201707 E 7/25/19	0.200	DR3856	201707 C 7/24/19	0.080 DGS	N/A	AG805701 2/26/20	<b>Maintenance</b> Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____																								
Simulator	Serial #	Lot #/Exp																																									
0.050	SD1012	201707 D 7/25/19																																									
0.080	DR1279	201707 E 7/25/19																																									
0.200	DR3856	201707 C 7/24/19																																									
0.080 DGS	N/A	AG805701 2/26/20																																									
<b>Calibration Adjustment</b> Performed By <u>JD</u> Barometric Pressure Gauge <u>1021</u> ID # <u>26932</u>		<b>Department Inspection</b> Performed By <u>JD</u> Barometric Pressure ID# <u>28421</u> Gauge <u>1020</u> Instrument <u>1021</u> Mouth Alcohol Solution Lot # <u>2018-B</u> Acetone Stock Solution Lot # <u>2018-A</u>																																									
<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>G2834</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>SD1022</td> <td>17410</td> <td>12/6/19</td> </tr> <tr> <td>0.100</td> <td>SD3964</td> <td>18070</td> <td>2/26/20</td> </tr> <tr> <td>0.200</td> <td>SD1025</td> <td>17340</td> <td>10/9/19</td> </tr> <tr> <td>0.300</td> <td>SD1024</td> <td>18110</td> <td>4/2/20</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>17817080A2</td> <td>8/5/19</td> </tr> </tbody> </table>		Simulator	Serial Number	Lot Number	Expiration	0.000	G2834	N/A	N/A	0.040	SD1022	17410	12/6/19	0.100	SD3964	18070	2/26/20	0.200	SD1025	17340	10/9/19	0.300	SD1024	18110	4/2/20	0.080 DGS	N/A	17817080A2	8/5/19	<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>G2408</td> </tr> <tr> <td>Interferent</td> <td>G2882</td> </tr> <tr> <td>0.050</td> <td>SD1012</td> </tr> <tr> <td>0.080</td> <td>DR1279</td> </tr> <tr> <td>0.200</td> <td>DR3856</td> </tr> </tbody> </table>		Simulator	Serial Number	0.000	G2408	Interferent	G2882	0.050	SD1012	0.080	DR1279	0.200	DR3856
Simulator	Serial Number	Lot Number	Expiration																																								
0.000	G2834	N/A	N/A																																								
0.040	SD1022	17410	12/6/19																																								
0.100	SD3964	18070	2/26/20																																								
0.200	SD1025	17340	10/9/19																																								
0.300	SD1024	18110	4/2/20																																								
0.080 DGS	N/A	17817080A2	8/5/19																																								
Simulator	Serial Number																																										
0.000	G2408																																										
Interferent	G2882																																										
0.050	SD1012																																										
0.080	DR1279																																										
0.200	DR3856																																										
<input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD1012</td> <td>201707 D</td> <td>7/25/19</td> </tr> <tr> <td>0.080</td> <td>DR1279</td> <td>201707 E</td> <td>7/25/19</td> </tr> <tr> <td>0.200</td> <td>DR3856</td> <td>201707 C</td> <td>7/24/19</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG805701</td> <td>2/26/20</td> </tr> </tbody> </table>		Simulator	Serial Number	Lot Number	Expiration	0.050	SD1012	201707 D	7/25/19	0.080	DR1279	201707 E	7/25/19	0.200	DR3856	201707 C	7/24/19	0.080 DGS	N/A	AG805701	2/26/20	<b>Attachments</b> <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input checked="" type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____																					
Simulator	Serial Number	Lot Number	Expiration																																								
0.050	SD1012	201707 D	7/25/19																																								
0.080	DR1279	201707 E	7/25/19																																								
0.200	DR3856	201707 C	7/24/19																																								
0.080 DGS	N/A	AG805701	2/26/20																																								
Notes/Suggested Service: _____ _____ _____ _____ _____		<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use  <u>JD</u> 1/9/19 <u>Brett Kirkland</u> 1/9/19 Tech Review / Date Admin Review / Date																																									

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: OCALA POLICE DEPT  
Time of Inspection: 12:16

Date of Inspection: 01/09/2019

Serial Number: 80-005454  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG805701 Exp: 02/26/2020
0.000	0.050	0.081	0.200	0.080
0.000	0.049	0.081	0.200	0.080
0.000	0.050	0.080	0.200	0.080
0.000	0.049	0.080	0.200	0.080
0.000	0.050	0.080	0.199	0.080
0.000	0.050	0.081	0.199	0.081
0.000	0.050	0.080	0.199	0.080
0.000	0.051	0.080	0.199	0.080
0.000	0.050	0.080	0.199	0.080
0.000	0.050	0.080	0.199	0.080

Standard Deviations	0.0005	0.0004	0.0005	0.0003
---------------------	--------	--------	--------	--------


Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

TJG  
BK  
1/9/19

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

  
 \_\_\_\_\_  
 Signature and Printed Name

THOMAS J GRAHAM

01/09/2019  
 Date

80-005454

1/9/19  
JD

Stability  
Checks

OCALA POLICE DEPT  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-005454  
01/09/2019  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:25
Control Test	0.082	08:25
Air Blank	0.000	08:26
Control Test	0.081	08:27
Air Blank	0.000	08:27
Control Test	0.081	08:28
Air Blank	0.000	08:28
Control Test Stats	0.0813	
Average	0.0006	
Std Dev	0.7099	
Rel Std Dev(%)		

JD

Operator's Signature

OCALA POLICE DEPT  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-005454  
01/09/2019  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:38
Control Test	0.049	08:38
Air Blank	0.000	08:39
Control Test	0.049	08:40
Air Blank	0.000	08:40
Control Test	0.049	08:41
Air Blank	0.000	08:41
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

JD

Operator's Signature

OCALA POLICE DEPT  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-005454  
01/09/2019  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:29
Control Test	0.197	08:30
Air Blank	0.000	08:31
Control Test	0.198	08:31
Air Blank	0.000	08:32
Control Test	0.199	08:33
Air Blank	0.000	08:33
Control Test Stats		
Average	0.1980	
Std Dev	0.0010	
Rel Std Dev(%)	0.5051	

JD

Operator's Signature

OCALA POLICE DEPT  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-005454  
01/09/2019  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:34
Control Test	0.077	08:34
Air Blank	0.000	08:35
Control Test	0.077	08:35
Air Blank	0.000	08:36
Control Test	0.076	08:36
Air Blank	0.000	08:37
Control Test Stats		
Average	0.0767	
Std Dev	0.0006	
Rel Std Dev(%)	0.7531	

DGS

JD

Operator's Signature

JD  
1/9/19



# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
2729 Fort Knox Blvd.  
Bldg. 2, Suite 1300  
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-005454, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-005454</u>	UNCERTAINTY* ±	
Owning Agency:	<u>OCALA POLICE DEPT</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>01/09/2019</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>12:16</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration. \*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

01/09/2019

Date

THOMAS J GRAHAM,  
Department Inspector

FDLE/ATP Form 69 July 2018

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

*Room BK 1/9/19*

80-005454  
 1/9/19  
 Calibration Adjustment

<<<< CHANNEL 2 >>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 1.5830 (-0.0070)  
 Sample #2 = 1.5790 (-0.0110)  
 Sample #3 = 1.5640 (0.0170)  
 Sample #4 = 1.5450 (0.0100)  
 Avg % Abs = 1.5627 (0.0053)  
 STD DEV = 0.0170 (0.0146)  
 REL STD DEV = 1.090 (273.219)

OCALA POLICE DEPT  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-005454  
 01/09/2019 19:23:59

Auto Calibration  
 Max Power Res Value = 37  
 Auto Range Res Value = 16

Sol Val = 0.100 g/210L \*\*\*  
 Fit value = 0.4762 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 12545, Sum Io = 12617  
 <<<< CHANNEL 1 >>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 1.8190 (-0.0180)  
 Sample #2 = 1.7750 (0.0260)  
 Sample #3 = 1.8380 (0.0000)  
 Sample #4 = 1.8350 (0.0140)  
 Avg % Abs = 1.8160 (0.0133)  
 STD DEV = 0.0355 (0.0130)  
 REL STD DEV = 1.957 (97.596)

Sol Val = 0.000 g/210L \*\*\*  
 Fit value = 0.0000 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 12573, Sum Io = 12628  
 <<<< CHANNEL 1 >>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 0.0580 (0.0130)  
 Sample #2 = 0.0410 (0.0800)  
 Sample #3 = 0.0900 (0.0840)  
 Sample #4 = 0.0390 (0.1350)  
 Avg % Abs = 0.0567 (0.0997)  
 STD DEV = 0.0289 (0.0307)  
 REL STD DEV = 50.973 (30.767)

<<<< CHANNEL 2 >>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 3.5550 (-0.0050)  
 Sample #2 = 3.5620 (-0.0120)  
 Sample #3 = 3.5600 (0.0000)  
 Sample #4 = 3.5770 (-0.0030)  
 Avg % Abs = 3.5663 (-0.0050)  
 STD DEV = 0.0093 (0.0062)  
 REL STD DEV = 0.261 (124.900)

Sol Val = 0.200 g/210L \*\*\*  
 Fit value = 0.9524 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 12539, Sum Io = 12616  
 <<<< CHANNEL 1 >>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 3.4450 (0.0000)  
 Sample #2 = 3.4710 (0.0140)  
 Sample #3 = 3.5090 (0.0080)  
 Sample #4 = 3.4650 (0.0360)  
 Avg % Abs = 3.4817 (0.0193)  
 STD DEV = 0.0239 (0.0147)  
 REL STD DEV = 0.685 (76.253)

<<<< CHANNEL 2 >>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 0.2080 (0.0050)  
 Sample #2 = 0.1870 (0.0340)  
 Sample #3 = 0.2240 (0.0260)  
 Sample #4 = 0.1990 (0.0400)  
 Avg % Abs = 0.2033 (0.0333)  
 STD DEV = 0.0189 (0.0070)  
 REL STD DEV = 9.284 (21.071)

Sol Val = 0.040 g/210L \*\*\*  
 Fit value = 0.1905 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 12552, Sum Io = 12620  
 <<<< CHANNEL 1 >>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 0.7760 (-0.0100)  
 Sample #2 = 0.7820 (0.0060)  
 Sample #3 = 0.7290 (0.0450)  
 Sample #4 = 0.7560 (0.0320)  
 Avg % Abs = 0.7557 (0.0277)  
 STD DEV = 0.0265 (0.0199)  
 REL STD DEV = 3.507 (71.775)

DOAN  
 BK/1/9/19

\*\*\*\*\* AUTO CAL DATA \*\*\*\*\*  
 <<<< CHANNEL 1 >>>>  
 Sol Val = 0.0000 mg/l or 0.000 g/210L  
 % Abs = 0.057  
 Std Dev = 0.03 Rel Std Dev = 50.97  
 Sol Val = 0.1905 mg/l or 0.040 g/210L  
 % Abs = 0.756  
 Std Dev = 0.03 Rel Std Dev = 3.51  
 Sol Val = 0.4762 mg/l or 0.100 g/210L  
 % Abs = 1.816  
 Std Dev = 0.04 Rel Std Dev = 1.96  
 Sol Val = 0.9524 mg/l or 0.200 g/210L  
 % Abs = 3.482  
 Std Dev = 0.02 Rel Std Dev = 0.69  
 Sol Val = 1.4286 mg/l or 0.300 g/210L  
 % Abs = 5.102  
 Std Dev = 0.02 Rel Std Dev = 0.32  
 Zero Order Coef = -142.16  
 First Order Coef = 2651.21  
 Second Order Coef = 34.80  
 Standard Deviation = 19.097908

<<<< CHANNEL 2 >>>>  
 Sol Val = 0.0000 mg/l or 0.000 g/210L  
 % Abs = 0.203  
 Std Dev = 0.02 Rel Std Dev = 9.28  
 Sol Val = 0.1905 mg/l or 0.040 g/210L  
 % Abs = 1.563  
 Std Dev = 0.02 Rel Std Dev = 1.09  
 Sol Val = 0.4762 mg/l or 0.100 g/210L  
 % Abs = 3.566  
 Std Dev = 0.01 Rel Std Dev = 0.26  
 Sol Val = 0.9524 mg/l or 0.200 g/210L  
 % Abs = 6.712  
 Std Dev = 0.01 Rel Std Dev = 0.20  
 Sol Val = 1.4286 mg/l or 0.300 g/210L  
 % Abs = 9.692  
 Std Dev = 0.03 Rel Std Dev = 0.28  
 Zero Order Coef = -271.78  
 First Order Coef = 1362.38  
 Second Order Coef = 14.40  
 Standard Deviation = 8.051412

Solution Stats Quadratic Fit Chan 1

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0002
0.040	0.040	0.0005
0.100	0.101	-0.0005
0.200	0.200	0.0003
0.300	0.300	-0.0001

Solution Stats Quadratic Fit Chan 2

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0001
0.040	0.040	0.0003
0.100	0.100	-0.0002
0.200	0.200	0.0000
0.300	0.300	0.0000

Sol Value = 0.080 g/210L \*\*\*  
 Fit value = 0.3810 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 \*\*\*\*\* CHANNEL 1 \*\*\*\*\*  
 Sample #1 = 3323.00  
 Sample #2 = 3301.00  
 Sample #3 = 3274.00  
 Sample #4 = 3374.00  
 Average Result = 3316.3333  
 STD DEV = 51.7333  
 REL STD DEV = 1.560  
 \*\*\*\*\* CHANNEL 2 \*\*\*\*\*  
 Sample #1 = 3139.00  
 Sample #2 = 3086.00  
 Sample #3 = 3118.00  
 Sample #4 = 3130.00  
 Average Result = 3111.3333  
 STD DEV = 22.7450  
 REL STD DEV = 0.731  
 \*\*\*\*\*  
 Dry Gas H2O Adjust Results \*\*\*\*\*  
 Barometric Pressure = 1021  
 3 um H2O Adjust (mg/l\*10,000) = 493  
 9 um H2O Adjust (mg/l\*10,000) = 698  
 \*\*\*\*\* AUTO CAL PASS \*\*\*\*\*

80-005454

Post Stabilities

1/9/19 JSD

OCALA POLICE DEPT  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-005454  
01/09/2019  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:08
Control Test	0.048	10:09
Air Blank	0.000	10:09
Control Test	0.049	10:10
Air Blank	0.000	10:10
Control Test	0.049	10:11
Air Blank	0.000	10:12
Control Test Stats		
Average	0.0487	
Std Dev	0.0006	
Rel Std Dev(%)	1.1863	

JSD

Operator's Signature

OCALA POLICE DEPT  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-005454  
01/09/2019  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:20
Control Test	0.079	10:21
Air Blank	0.000	10:21
Control Test	0.079	10:22
Air Blank	0.000	10:23
Control Test	0.079	10:23
Air Blank	0.000	10:24
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

JSD

Operator's Signature

OCALA POLICE DEPT  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-005454  
01/09/2019  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:35
Control Test	0.198	10:35
Air Blank	0.000	10:36
Control Test	0.199	10:36
Air Blank	0.000	10:37
Control Test	0.198	10:38
Air Blank	0.000	10:38
Control Test Stats		
Average	0.1983	
Std Dev	0.0006	
Rel Std Dev(%)	0.2911	

JSD

Operator's Signature

OCALA POLICE DEPT  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-005454  
01/09/2019  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:13
Control Test	0.080	10:14
Air Blank	0.000	10:14
Control Test	0.079	10:14
Air Blank	0.000	10:15
Control Test	0.080	10:15
Air Blank	0.000	10:16
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

DGS

JSD

Operator's Signature

JSD  
1/9/19