



INSTRUMENT PROCESSING SHEET

Agency Miami-Dade Police Department

S/N 80-005551

Florida Department of Law Enforcement

Date In 01/15/2019 DI Completion Date 1/15/2019

Ship P/U H/D CMI EE

<b>Intake</b> Performed By <u>DLL</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____		<b>Quality Checks</b> Performed By <u>DLL</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>141</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP 106</u> 32 mm <u>.148</u> (.139 - .169) 36 mm <u>.171</u> (.156 - .190) 53 mm <u>.234</u> (.228 - .278) 103 mm <u>.511</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28663</u> <input checked="" type="checkbox"/> Stability Checks		<b>Flow Calibration</b> Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																							
<b>Final Release Date</b> <p style="text-align: center;">FDLE</p> <p style="text-align: center;">JAN 23 2019</p> <p style="text-align: center;">Alcohol Testing Program</p>		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD3967</td> <td>201707D 07/25/2019</td> </tr> <tr> <td>0.080</td> <td>SD3968</td> <td>201707E 07/25/2019</td> </tr> <tr> <td>0.200</td> <td>SD3969</td> <td>201707C 07/24/2019</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG805701 02/26/2020</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #/Exp	0.050	SD3967	201707D 07/25/2019	0.080	SD3968	201707E 07/25/2019	0.200	SD3969	201707C 07/24/2019	0.080 DGS	N/A	AG805701 02/26/2020	<b>Maintenance</b> Performed By <u>DLL</u> <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input checked="" type="checkbox"/> Other <u>Form load/Changed pass</u>																								
Simulator	Serial #	Lot #/Exp																																									
0.050	SD3967	201707D 07/25/2019																																									
0.080	SD3968	201707E 07/25/2019																																									
0.200	SD3969	201707C 07/24/2019																																									
0.080 DGS	N/A	AG805701 02/26/2020																																									
<b>Calibration Adjustment</b> Performed By _____ Barometric Pressure Gauge _____ ID # _____		<b>Department Inspection</b> Performed By <u>DLL</u> Barometric Pressure ID# <u>68639</u> Gauge <u>1021</u> Instrument <u>1022</u> Mouth Alcohol Solution Lot # <u>2017-B</u> Acetone Stock Solution Lot # <u>2018-A</u>																																									
<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>		Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>SD3965</td> </tr> <tr> <td>Interferent</td> <td>SD3966</td> </tr> <tr> <td>0.050</td> <td>SD3967</td> </tr> <tr> <td>0.080</td> <td>SD3968</td> </tr> <tr> <td>0.200</td> <td>SD3969</td> </tr> </tbody> </table>		Simulator	Serial Number	0.000	SD3965	Interferent	SD3966	0.050	SD3967	0.080	SD3968	0.200	SD3969
Simulator	Serial Number	Lot Number	Expiration																																								
0.000		N/A	N/A																																								
0.040																																											
0.100																																											
0.200																																											
0.300																																											
0.080 DGS	N/A																																										
Simulator	Serial Number																																										
0.000	SD3965																																										
Interferent	SD3966																																										
0.050	SD3967																																										
0.080	SD3968																																										
0.200	SD3969																																										
<input type="checkbox"/> Post Calibration Adjustment Stability Checks		<b>Attachments</b> <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____																																									
Notes/Suggested Service: <u>E-mailed</u> <input checked="" type="checkbox"/> <span style="color: red; font-weight: bold;">APPROVED</span>		<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																																									
_____ _____ _____ _____		<u>Peggy</u> 1/23/19 <u>J. P. [Signature]</u> 1/23/19 Tech Review / Date                   Admin Review / Date																																									

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI DADE PD  
Time of Inspection: 16:14

Date of Inspection: 01/15/2019

Serial Number: 80-005551  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG805701 Exp: 02/26/2020
0.000	0.049	0.081	0.201	0.079
0.000	0.050	0.082	0.203	0.079
0.000	0.050	0.082	0.204	0.078
0.000	0.050	0.082	0.203	0.079
0.000	0.050	0.082	0.204	0.078
0.000	0.050	0.082	0.204	0.078
0.000	0.050	0.082	0.204	0.078
0.000	0.050	0.081	0.204	0.078
0.000	0.050	0.082	0.206	0.078
0.000	0.050	0.082	0.205	0.079

Standard Deviations	0.0003	0.0004	0.0013	0.0005
---------------------	--------	--------	--------	--------


Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0006 Number of Simulators Used: 5

**Remarks:**

A F / M A: RAN OUT OF BREATH.

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


DAVID E REYES-RIVERA  
 \_\_\_\_\_  
 Signature and Printed Name

01/15/2019  
 Date

1/23/19  
22



<b>TYPE OF TEST</b>	<b>SERIAL NUMBER</b>	<b>AGENCY</b>	<b>DATE</b>	<b>PERFORMED BY</b>
Stabilities	80-005551	Miami-Dade Police Department	01/15/2019	DELL

<b>0.05g/210L</b>	<b>0.08g/210L</b>	<b>0.20g/210L</b>	<b>DGS 0.08g/210L</b>
<b>SN: SD3967 Temp: 34.10c</b>	<b>SN: SD3968 Temp: 34.09c</b>	<b>SN: SD3969 Temp: 34.10c</b>	<b>Lot AG805701</b>
<b>0.047 to 0.053</b> <input checked="" type="checkbox"/>	<b>0.077 to 0.083</b> <input checked="" type="checkbox"/>	<b>0.194 to 0.206</b> <input checked="" type="checkbox"/>	<b>0.077 to 0.083</b> <input checked="" type="checkbox"/>

<p>MIAMI DADE PD Intoxilyzer - Alconol Analyzer Model: 8000 01/15/2019 Software: 8100.27</p> <p>MIAMI DADE PD Intoxilyzer - Alconol Analyzer Model: 8000 01/15/2019 Software: 8100.27</p> <p>MIAMI DADE PD Intoxilyzer - Alconol Analyzer Model: 8000 01/15/2019 Software: 8100.27</p> <p>MIAMI DADE PD Intoxilyzer - Alconol Analyzer Model: 8000 01/15/2019 Software: 8100.27</p>	<p>MIAMI DADE PD Intoxilyzer - Alconol Analyzer Model: 8000 01/15/2019 Software: 8100.27</p> <p>MIAMI DADE PD Intoxilyzer - Alconol Analyzer Model: 8000 01/15/2019 Software: 8100.27</p> <p>MIAMI DADE PD Intoxilyzer - Alconol Analyzer Model: 8000 01/15/2019 Software: 8100.27</p> <p>MIAMI DADE PD Intoxilyzer - Alconol Analyzer Model: 8000 01/15/2019 Software: 8100.27</p>	<p>MIAMI DADE PD Intoxilyzer - Alconol Analyzer Model: 8000 01/15/2019 Software: 8100.27</p> <p>MIAMI DADE PD Intoxilyzer - Alconol Analyzer Model: 8000 01/15/2019 Software: 8100.27</p> <p>MIAMI DADE PD Intoxilyzer - Alconol Analyzer Model: 8000 01/15/2019 Software: 8100.27</p> <p>MIAMI DADE PD Intoxilyzer - Alconol Analyzer Model: 8000 01/15/2019 Software: 8100.27</p>	<p>MIAMI DADE PD Intoxilyzer - Alconol Analyzer Model: 8000 01/15/2019 Software: 8100.27</p> <p>MIAMI DADE PD Intoxilyzer - Alconol Analyzer Model: 8000 01/15/2019 Software: 8100.27</p> <p>MIAMI DADE PD Intoxilyzer - Alconol Analyzer Model: 8000 01/15/2019 Software: 8100.27</p> <p>MIAMI DADE PD Intoxilyzer - Alconol Analyzer Model: 8000 01/15/2019 Software: 8100.27</p>																																																																																																																																																
<table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:49</td></tr> <tr><td>Control Test</td><td>0.050</td><td>11:50</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:50</td></tr> <tr><td>Control Test</td><td>0.050</td><td>11:51</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:52</td></tr> <tr><td>Control Test</td><td>0.050</td><td>11:52</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:53</td></tr> <tr><td>Control Test</td><td>0.000</td><td>11:53</td></tr> <tr><td>Average</td><td>0.0500</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	11:49	Control Test	0.050	11:50	Air Blank	0.000	11:50	Control Test	0.050	11:51	Air Blank	0.000	11:52	Control Test	0.050	11:52	Air Blank	0.000	11:53	Control Test	0.000	11:53	Average	0.0500		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:54</td></tr> <tr><td>Control Test</td><td>0.082</td><td>11:55</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:56</td></tr> <tr><td>Control Test</td><td>0.082</td><td>11:56</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:57</td></tr> <tr><td>Control Test</td><td>0.081</td><td>11:58</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:58</td></tr> <tr><td>Control Test</td><td>0.000</td><td>11:58</td></tr> <tr><td>Average</td><td>0.0817</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7072</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	11:54	Control Test	0.082	11:55	Air Blank	0.000	11:56	Control Test	0.082	11:56	Air Blank	0.000	11:57	Control Test	0.081	11:58	Air Blank	0.000	11:58	Control Test	0.000	11:58	Average	0.0817		Std Dev	0.0006		Rel Std Dev(%)	0.7072		<table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:00</td></tr> <tr><td>Control Test</td><td>0.204</td><td>12:01</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:01</td></tr> <tr><td>Control Test</td><td>0.204</td><td>12:02</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:03</td></tr> <tr><td>Control Test</td><td>0.203</td><td>12:04</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:04</td></tr> <tr><td>Control Test</td><td>0.000</td><td>12:04</td></tr> <tr><td>Average</td><td>0.2037</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2835</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	12:00	Control Test	0.204	12:01	Air Blank	0.000	12:01	Control Test	0.204	12:02	Air Blank	0.000	12:03	Control Test	0.203	12:04	Air Blank	0.000	12:04	Control Test	0.000	12:04	Average	0.2037		Std Dev	0.0006		Rel Std Dev(%)	0.2835		<table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:07</td></tr> <tr><td>Control Test</td><td>0.079</td><td>12:08</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:08</td></tr> <tr><td>Control Test</td><td>0.080</td><td>12:08</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:09</td></tr> <tr><td>Control Test</td><td>0.079</td><td>12:09</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:10</td></tr> <tr><td>Control Test</td><td>0.000</td><td>12:10</td></tr> <tr><td>Average</td><td>0.0793</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7277</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	12:07	Control Test	0.079	12:08	Air Blank	0.000	12:08	Control Test	0.080	12:08	Air Blank	0.000	12:09	Control Test	0.079	12:09	Air Blank	0.000	12:10	Control Test	0.000	12:10	Average	0.0793		Std Dev	0.0006		Rel Std Dev(%)	0.7277	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:49																																																																																																																																																	
Control Test	0.050	11:50																																																																																																																																																	
Air Blank	0.000	11:50																																																																																																																																																	
Control Test	0.050	11:51																																																																																																																																																	
Air Blank	0.000	11:52																																																																																																																																																	
Control Test	0.050	11:52																																																																																																																																																	
Air Blank	0.000	11:53																																																																																																																																																	
Control Test	0.000	11:53																																																																																																																																																	
Average	0.0500																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:54																																																																																																																																																	
Control Test	0.082	11:55																																																																																																																																																	
Air Blank	0.000	11:56																																																																																																																																																	
Control Test	0.082	11:56																																																																																																																																																	
Air Blank	0.000	11:57																																																																																																																																																	
Control Test	0.081	11:58																																																																																																																																																	
Air Blank	0.000	11:58																																																																																																																																																	
Control Test	0.000	11:58																																																																																																																																																	
Average	0.0817																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7072																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:00																																																																																																																																																	
Control Test	0.204	12:01																																																																																																																																																	
Air Blank	0.000	12:01																																																																																																																																																	
Control Test	0.204	12:02																																																																																																																																																	
Air Blank	0.000	12:03																																																																																																																																																	
Control Test	0.203	12:04																																																																																																																																																	
Air Blank	0.000	12:04																																																																																																																																																	
Control Test	0.000	12:04																																																																																																																																																	
Average	0.2037																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2835																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:07																																																																																																																																																	
Control Test	0.079	12:08																																																																																																																																																	
Air Blank	0.000	12:08																																																																																																																																																	
Control Test	0.080	12:08																																																																																																																																																	
Air Blank	0.000	12:09																																																																																																																																																	
Control Test	0.079	12:09																																																																																																																																																	
Air Blank	0.000	12:10																																																																																																																																																	
Control Test	0.000	12:10																																																																																																																																																	
Average	0.0793																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7277																																																																																																																																																		
<p>Operator's Signature: <i>DELL</i></p>	<p>Operator's Signature: <i>DELL</i></p>	<p>Operator's Signature: <i>DELL</i></p>	<p>Operator's Signature: <i>DELL</i></p>																																																																																																																																																

1/23/19  
*DELL*

*DELL*



# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
4700 Terminal Drive, Suite 1  
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-005551, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-005551</u>	UNCERTAINTY* ±	
Owning Agency:	<u>MIAMI DADE PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>01/15/2019</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>16:14</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.  
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.  
\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

**TRACEABILITY INFORMATION**  
This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 July 2018  
Issuing Authority: Alcohol Testing Program

01/15/2019 Date  
David Reyes-Rivera  
**DAVID E REYES-RIVERA,**  
Department Inspector

Service • Integrity • Respect • Quality

*6/1/19*

*(Handwritten initials)*