



INSTRUMENT PROCESSING SHEET

Agency Gainesville PDS/N 80-006384Florida Department of
Law EnforcementDate In 7/3/2019 DI Completion Date 7/9/19☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake Performed By <u>QPM</u> ☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____	Quality Checks Performed By <u>BK</u> ☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>175</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP102</u> 32 mm <u>.167</u> (.139 - .169) 36 mm <u>.179</u> (.156 - .190) 53 mm <u>.250</u> (.228 - .278) 103 mm <u>.507</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28427</u> ☒ Stability Checks <table border="1" style="width:100%; border-collapse: collapse; margin-top: 5px;"> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> <tr> <td>0.050</td> <td><u>SD1018</u></td> <td><u>201707D</u> <u>7/25/19</u></td> </tr> <tr> <td>0.080</td> <td><u>SD3962</u></td> <td><u>201707E</u> <u>7/25/19</u></td> </tr> <tr> <td>0.200</td> <td><u>G2078</u></td> <td><u>201707C</u> <u>7/24/19</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>AG916501</u> <u>6/14/21</u></td> </tr> </table>	Simulator	Serial #	Lot #/Exp	0.050	<u>SD1018</u>	<u>201707D</u> <u>7/25/19</u>	0.080	<u>SD3962</u>	<u>201707E</u> <u>7/25/19</u>	0.200	<u>G2078</u>	<u>201707C</u> <u>7/24/19</u>	0.080 DGS	N/A	<u>AG916501</u> <u>6/14/21</u>	Flow Calibration Performed By _____ Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) Maintenance Performed By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ Temperature Checks Performed By <u>BK</u> ☒ Lab Temp °C <u>21.04</u> External Digital Therm. ID#: <u>300505</u> ☒ 34°C +/- .2 Serial #: <u>SD1018</u> ☒ 34°C +/- .2 Serial #: <u>SD3962</u> ☒ 34°C +/- .2 Serial #: <u>G2078</u>																																	
Simulator	Serial #	Lot #/Exp																																																
0.050	<u>SD1018</u>	<u>201707D</u> <u>7/25/19</u>																																																
0.080	<u>SD3962</u>	<u>201707E</u> <u>7/25/19</u>																																																
0.200	<u>G2078</u>	<u>201707C</u> <u>7/24/19</u>																																																
0.080 DGS	N/A	<u>AG916501</u> <u>6/14/21</u>																																																
Final Release Date FDLE JUL 09 2019 Alcohol Testing Program	Calibration Adjustment Performed By <u>BK</u> Barometric Pressure Gauge <u>1011</u> ID # <u>30793</u> <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> <tr> <td>0.000</td> <td><u>SD1016</u></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td><u>SD1022</u></td> <td><u>19080</u></td> <td><u>3/4/21</u></td> </tr> <tr> <td>0.100</td> <td><u>SD3964</u></td> <td><u>18200</u></td> <td><u>7/3/20</u></td> </tr> <tr> <td>0.200</td> <td><u>SD1025</u></td> <td><u>19040</u></td> <td><u>1/29/21</u></td> </tr> <tr> <td>0.300</td> <td><u>SD1024</u></td> <td><u>18110</u></td> <td><u>4/2/20</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>17817180A2</u></td> <td><u>8/5/19</u></td> </tr> </table> ☒ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> <tr> <td>0.050</td> <td><u>SD1018</u></td> <td><u>201707D</u></td> <td><u>7/25/19</u></td> </tr> <tr> <td>0.080</td> <td><u>SD3962</u></td> <td><u>201707E</u></td> <td><u>7/25/19</u></td> </tr> <tr> <td>0.200</td> <td><u>G2078</u></td> <td><u>201707C</u></td> <td><u>7/24/19</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>AG916501</u></td> <td><u>6/14/21</u></td> </tr> </table> Notes/Suggested Service: _____ _____ _____ _____ _____ _____		Simulator	Serial Number	Lot Number	Expiration	0.000	<u>SD1016</u>	N/A	N/A	0.040	<u>SD1022</u>	<u>19080</u>	<u>3/4/21</u>	0.100	<u>SD3964</u>	<u>18200</u>	<u>7/3/20</u>	0.200	<u>SD1025</u>	<u>19040</u>	<u>1/29/21</u>	0.300	<u>SD1024</u>	<u>18110</u>	<u>4/2/20</u>	0.080 DGS	N/A	<u>17817180A2</u>	<u>8/5/19</u>	Simulator	Serial Number	Lot Number	Expiration	0.050	<u>SD1018</u>	<u>201707D</u>	<u>7/25/19</u>	0.080	<u>SD3962</u>	<u>201707E</u>	<u>7/25/19</u>	0.200	<u>G2078</u>	<u>201707C</u>	<u>7/24/19</u>	0.080 DGS	N/A	<u>AG916501</u>	<u>6/14/21</u>
Simulator	Serial Number	Lot Number	Expiration																																															
0.000	<u>SD1016</u>	N/A	N/A																																															
0.040	<u>SD1022</u>	<u>19080</u>	<u>3/4/21</u>																																															
0.100	<u>SD3964</u>	<u>18200</u>	<u>7/3/20</u>																																															
0.200	<u>SD1025</u>	<u>19040</u>	<u>1/29/21</u>																																															
0.300	<u>SD1024</u>	<u>18110</u>	<u>4/2/20</u>																																															
0.080 DGS	N/A	<u>17817180A2</u>	<u>8/5/19</u>																																															
Simulator	Serial Number	Lot Number	Expiration																																															
0.050	<u>SD1018</u>	<u>201707D</u>	<u>7/25/19</u>																																															
0.080	<u>SD3962</u>	<u>201707E</u>	<u>7/25/19</u>																																															
0.200	<u>G2078</u>	<u>201707C</u>	<u>7/24/19</u>																																															
0.080 DGS	N/A	<u>AG916501</u>	<u>6/14/21</u>																																															
Department Inspection Performed By <u>BK</u> Barometric Pressure ID# <u>28662</u> Gauge <u>1012</u> Instrument <u>1010</u> Mouth Alcohol Solution Lot # <u>2018-B</u> Acetone Stock Solution Lot # <u>2018-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> <tr> <td>0.000</td> <td><u>G4444</u></td> </tr> <tr> <td>Interferent</td> <td><u>G6621</u></td> </tr> <tr> <td>0.050</td> <td><u>SD1018</u></td> </tr> <tr> <td>0.080</td> <td><u>SD3962</u></td> </tr> <tr> <td>0.200</td> <td><u>G2078</u></td> </tr> </table> Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Calibration Adjustment ☒ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____ ☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use <u>QPM 7/9/19</u> Tech Review / Date <u>J. Baker 7/9/19</u> Admin Review / Date			Simulator	Serial Number	0.000	<u>G4444</u>	Interferent	<u>G6621</u>	0.050	<u>SD1018</u>	0.080	<u>SD3962</u>	0.200	<u>G2078</u>																																				
Simulator	Serial Number																																																	
0.000	<u>G4444</u>																																																	
Interferent	<u>G6621</u>																																																	
0.050	<u>SD1018</u>																																																	
0.080	<u>SD3962</u>																																																	
0.200	<u>G2078</u>																																																	

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: GAINESVILLE PD

Time of Inspection: 12:31

Date of Inspection: 07/09/2019

Serial Number: 80-006384

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG916501 Exp: 06/14/2021
0.000	0.049	0.082	0.202	0.080
0.000	0.050	0.082	0.203	0.080
0.000	0.050	0.082	0.204	0.080
0.000	0.050	0.082	0.204	0.080
0.000	0.050	0.083	0.203	0.080
0.000	0.050	0.082	0.204	0.079
0.000	0.050	0.082	0.204	0.080
0.000	0.050	0.082	0.203	0.079
0.000	0.050	0.082	0.204	0.079
0.000	0.050	0.082	0.204	0.079

Standard Deviations	0.0003	0.0003	0.0007	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

pgm

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Brett Kirkland

BRETT H KIRKLAND

Signature and Printed Name

07/09/2019
Date

7/9/19
22

80-006384
BK
7/9/19
Stability Check

GAINESVILLE PD
Intoxilyzer - Alcohol Analyzer
Model 8000
07/09/2019
Software: 8100.27
SN 80-006384

Test	g/210L	Time
Air Blank	0.000	08:56
Control Test	0.076	08:57
Air Blank	0.000	08:58
Control Test	0.077	08:58
Air Blank	0.000	08:59
Control Test	0.077	09:00
Air Blank	0.000	09:00
Control Test Stats		
Average	0.0767	
Std Dev	0.0006	
Rel Std Dev(%)	0.7531	

BK

Operator's Signature

GAINESVILLE PD
Intoxilyzer - Alcohol Analyzer
Model 8000
07/09/2019
Software: 8100.27
SN 80-006384

Test	g/210L	Time
Air Blank	0.000	09:03
Control Test	0.046	09:04
Air Blank	0.000	09:04
Control Test	0.046	09:05
Air Blank	0.000	09:06
Control Test	0.047	09:06
Air Blank	0.000	09:07
Control Test Stats		
Average	0.0463	
Std Dev	0.0006	
Rel Std Dev(%)	1.2461	

BK

Operator's Signature

GAINESVILLE PD
Intoxilyzer - Alcohol Analyzer
Model 8000
07/09/2019
Software: 8100.27
SN 80-006384

Test	g/210L	Time
Air Blank	0.000	09:08
Control Test	0.191	09:09
Air Blank	0.000	09:09
Control Test	0.193	09:10
Air Blank	0.000	09:11
Control Test	0.194	09:11
Air Blank	0.000	09:12
Control Test Stats		
Average	0.1927	
Std Dev	0.0015	
Rel Std Dev(%)	0.7928	

BK

Operator's Signature

GAINESVILLE PD
Intoxilyzer - Alcohol Analyzer
Model 8000
07/09/2019
Software: 8100.27
SN 80-006384

Test	g/210L	Time
Air Blank	0.000	09:13
Control Test	0.080	09:14
Air Blank	0.000	09:14
Control Test	0.080	09:15
Air Blank	0.000	09:15
Control Test	0.080	09:16
Air Blank	0.000	09:16
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

BK

Operator's Signature

Adam 7/9/19
22



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006384, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006384</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>GAINESVILLE PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>07/09/2019</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>12:31</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

07/09/2019

Date

Brett Kirkland

BRETT H KIRKLAND,
Department Inspector

FDLE/ATP Form 69 July 2018

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1

7/9/19
22

GAINESVILLE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006384
07/09/2019 09:40:46

Auto Calibration
Max Power Res Value = 78
Auto Range Res Value = 55

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
3um lo = 12857, 9um lo = 13229

Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 0.0600 (-0.0060)
Sample #2 = 0.0430 (0.0190)
Sample #3 = 0.0710 (0.0140)
Sample #4 = 0.0820 (0.0260)
Avg % Abs = 0.0653 (0.0197)
STD DEV = 0.0201 (0.0060)
REL STD DEV = 30.778 (30.649)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 0.1540 (-0.0040)
Sample #2 = 0.1620 (-0.0090)
Sample #3 = 0.1580 (-0.0070)
Sample #4 = 0.1370 (0.0060)
Avg % Abs = 0.1523 (-0.0033)
STD DEV = 0.0134 (0.0081)
REL STD DEV = 8.815 (244.336)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
3um lo = 12848, 9um lo = 13227

Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 0.7470 (0.0000)
Sample #2 = 0.7930 (0.0000)
Sample #3 = 0.7880 (0.0050)
Sample #4 = 0.7470 (0.0310)
Avg % Abs = 0.7760 (0.0120)
STD DEV = 0.0252 (0.0166)
REL STD DEV = 3.252 (138.694)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 1.5530 (-0.0040)
Sample #2 = 1.5770 (-0.0130)
Sample #3 = 1.5730 (-0.0120)
Sample #4 = 1.5480 (0.0010)
Avg % Abs = 1.5660 (-0.0080)
STD DEV = 0.0157 (0.0078)
REL STD DEV = 1.004 (97.628)

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
3um lo = 12842, 9um lo = 13225

Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 1.8180 (-0.0120)
Sample #2 = 1.8580 (-0.0130)
Sample #3 = 1.8480 (-0.0010)
Sample #4 = 1.8040 (0.0220)
Avg % Abs = 1.8367 (0.0027)
STD DEV = 0.0287 (0.0178)
REL STD DEV = 1.564 (666.966)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 3.6230 (0.0110)
Sample #2 = 3.6440 (-0.0070)
Sample #3 = 3.6500 (-0.0010)
Sample #4 = 3.6150 (0.0060)
Avg % Abs = 3.6363 (-0.0007)
STD DEV = 0.0187 (0.0065)
REL STD DEV = 0.515 (975.961)

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
3um lo = 12840, 9um lo = 13222

Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 3.5150 (-0.0110)
Sample #2 = 3.4860 (0.0260)
Sample #3 = 3.5060 (0.0190)
Sample #4 = 3.5180 (0.0190)
Avg % Abs = 3.5033 (0.0213)
STD DEV = 0.0162 (0.0040)
REL STD DEV = 0.461 (18.944)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 6.8380 (-0.0090)
Sample #2 = 6.8410 (0.0030)
Sample #3 = 6.8660 (-0.0020)
Sample #4 = 6.8760 (-0.0070)
Avg % Abs = 6.8610 (-0.0020)
STD DEV = 0.0180 (0.0050)
REL STD DEV = 0.263 (250.000)

Sol Value = 0.300 g/210L ***
Fit value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
3um lo = 12836, 9um lo = 13220
Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 5.1340 (-0.0080)
Sample #2 = 5.1030 (0.0300)
Sample #3 = 5.1480 (0.0190)
Sample #4 = 5.1080 (0.0430)
Avg % Abs = 5.1197 (0.0307)
STD DEV = 0.0247 (0.0120)
REL STD DEV = 0.482 (39.176)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 9.8820 (-0.0290)
Sample #2 = 9.8720 (0.0060)
Sample #3 = 9.9060 (0.0030)
Sample #4 = 9.9100 (0.0140)
Avg % Abs = 9.8960 (0.0077)
STD DEV = 0.0209 (0.0057)
REL STD DEV = 0.211 (74.168)

AUTO CAL DATA
Channel 1
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.065
Std Dev = 0.02 Rel Std Dev = 30.78
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.776
Std Dev = 0.03 Rel Std Dev = 3.25
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.837
Std Dev = 0.03 Rel Std Dev = 1.56
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.503
Std Dev = 0.02 Rel Std Dev = 0.46
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.120
Std Dev = 0.02 Rel Std Dev = 0.48
Zero Order Coef = -170.15
First Order Coef = 2627.03
Second Order Coef = 38.59
Standard Deviation = 15.986573

Channel 2
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.152
Std Dev = 0.01 Rel Std Dev = 8.82
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.566
Std Dev = 0.02 Rel Std Dev = 1.00
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.636
Std Dev = 0.02 Rel Std Dev = 0.51
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.861
Std Dev = 0.02 Rel Std Dev = 0.26
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.896
Std Dev = 0.02 Rel Std Dev = 0.21
Zero Order Coef = -194.91
First Order Coef = 1309.01
Second Order Coef = 15.59
Standard Deviation = 8.015811

Solution Stats Quadratic Fit Chan 1
Act Fit Residual
g/210L g/210L g/210L
0.000 0.000 -0.0000
0.040 0.040 0.0003
0.100 0.100 -0.0005
0.200 0.200 0.0004
0.300 0.300 -0.0001

Solution Stats Quadratic Fit Chan 2
Act Fit Residual
g/210L g/210L g/210L
0.000 0.000 -0.0001
0.040 0.040 0.0002
0.100 0.100 -0.0002
0.200 0.200 0.0001
0.300 0.300 -0.0000

Sol Value = 0.080 g/210L ***
Fit value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1
Channel 1
Sample #1 = 3334.00
Sample #2 = 3346.00
Sample #3 = 3372.00
Sample #4 = 3271.00
Average Result = 3329.6667
STD DEV = 52.4436
REL STD DEV = 1.575
Channel 2
Sample #1 = 3403.00
Sample #2 = 3395.00
Sample #3 = 3383.00
Sample #4 = 3382.00
Average Result = 3386.6667
STD DEV = 7.2342
REL STD DEV = 0.214
Dry Gas H2O Adjust Results
Barometric Pressure = 1011
3 um H2O Adjust (mg/l*10,000) = 480
9 um H2O Adjust (mg/l*10,000) = 423
AUTO CAL PASS

Optical Bench Cal
80-006384
BK
7/9/19 7/9/19
J2
RBM

Best stability
80-006384
BK
7/9/19

GAINESVILLE PD
Intoxilyzer - Alcohol Analyzer
Model 8000
07/09/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:30
Control Test	0.049	10:31
Air Blank	0.000	10:31
Control Test	0.050	10:32
Air Blank	0.000	10:32
Control Test	0.050	10:33
Air Blank	0.000	10:34
Control Test Stats		
Average	0.0497	
Std Dev	0.0006	
Rel Std Dev(%)	1.1625	

BK
Operator's Signature

GAINESVILLE PD
Intoxilyzer - Alcohol Analyzer
Model 8000
07/09/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:35
Control Test	0.082	10:36
Air Blank	0.000	10:36
Control Test	0.082	10:37
Air Blank	0.000	10:37
Control Test	0.082	10:38
Air Blank	0.000	10:39
Control Test Stats		
Average	0.0820	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

BK
Operator's Signature

GAINESVILLE PD
Intoxilyzer - Alcohol Analyzer
Model 8000
07/09/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:40
Control Test	0.202	10:41
Air Blank	0.000	10:41
Control Test	0.203	10:42
Air Blank	0.000	10:43
Control Test	0.204	10:43
Air Blank	0.000	10:44
Control Test Stats		
Average	0.2030	
Std Dev	0.0010	
Rel Std Dev(%)	0.4926	

BK
Operator's Signature

GAINESVILLE PD
Intoxilyzer - Alcohol Analyzer
Model 8000
07/09/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:45
Control Test	0.079	10:45
Air Blank	0.000	10:46
Control Test	0.080	10:46
Air Blank	0.000	10:47
Control Test	0.079	10:47
Air Blank	0.000	10:48
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

QUS

BK
Operator's Signature

WGB

7/9/19
22