



INSTRUMENT PROCESSING SHEET

Agency Escambia County SOS/N 80-006435

Florida Department of Law Enforcement

Date In 07/01/2019 DI Completion Date 7/12/19 Ship P/U H/D CMI EE

Intake Performed By JA

Annual
 Registration
 Return from CMI / EE

Visual Inspection:
 Case Handle
 Keyboard Dry Gas Shelf
 Feet Breath Tube
 Ports Screws Tight

Other Equipment/ Accessories:
 Power cord Printer Cable
 Static Bag 12V DC Cable

Notes: _____

Quality Checks Performed By gym

Breath Tube Screen
 Replace External O-Rings
 Instrument Set Up Verified
 R-Value 202
 Flow Verification (L/s)
 Flow Column # ATP102
 32 mm .140 (.139 - .169)
 36 mm .156 (.156 - .190)
 53 mm .234 (.228 - .278)
 103 mm .502 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Flow Calibration Performed By _____

Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)

Flow Column # _____
 32 mm _____ (.139 - .169)
 36 mm _____ (.156 - .190)
 53 mm _____ (.228 - .278)
 103 mm _____ (.447 - .547)

Simulator	Serial #	Lot #/Exp
0.050	SD1012	201707D 7/25/19
0.080	DR1279	201707E 7/25/19
0.200	SD1013	201707C 7/24/19
0.080 DGS	N/A	AG916501 6/14/21

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Temperature Checks Performed By gym

Lab Temp °C 21.3
 External Digital Therm. ID#: 300503
 34°C +/-2 Serial #: 501018
 34°C +/-2 Serial #: 503962
 34°C +/-2 Serial #: 62078

Final Release Date

FDLE

JUL 16 2019

Alcohol Testing Program

Calibration Adjustment Performed By _____

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

Department Inspection Performed By gym

Barometric Pressure ID# 28427
 Gauge 1012 Instrument 1009
 Mouth Alcohol Solution Lot # 2018-B
 Acetone Stock Solution Lot # 2019-A

Simulator	Serial Number
0.000	64444
Interferent	66621
0.050	501018
0.080	503962
0.200	62078

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

Attachments

Form 41 Post-Stability Checks
 Stability Checks Flow Calibration
 Calibration Certificate Form 40
 Calibration Adjustment Other _____

Notes/Suggested Service: _____

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

SP 7/16/19 [Signature] 7/16/19
 Tech Review / Date Admin Review / Date

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: ESCAMBIA COUNTY SO
Time of Inspection: 10:26

Date of Inspection: 07/12/2019

Serial Number: 80-006435
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG916501 Exp: 06/14/2021
0.000	0.048	0.080	0.198	0.081
0.000	0.048	0.081	0.201	0.081
0.000	0.049	0.080	0.201	0.081
0.000	0.048	0.080	0.200	0.081
0.000	0.049	0.081	0.200	0.081
0.000	0.049	0.080	0.200	0.081
0.000	0.049	0.080	0.201	0.081
0.000	0.048	0.080	0.200	0.081
0.000	0.049	0.080	0.200	0.081
0.000	0.049	0.080	0.200	0.081

Standard Deviations	0.0005	0.0004	0.0008	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

SP

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Patrick J Murphy

PATRICK J MURPHY

Signature and Printed Name

07/12/2019
Date

*7/16/19
JD*

ESCAMBIA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006435
07/11/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:21
Control Test	0.049	12:22
Air Blank	0.000	12:22
Control Test	0.049	12:23
Air Blank	0.000	12:23
Control Test	0.049	12:24
Air Blank	0.000	12:25
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

ESCAMBIA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006435
07/11/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:26
Control Test	0.079	12:27
Air Blank	0.000	12:27
Control Test	0.080	12:28
Air Blank	0.000	12:29
Control Test	0.080	12:29
Air Blank	0.000	12:30
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

ESCAMBIA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006435
07/11/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:32
Control Test	0.198	12:33
Air Blank	0.000	12:33
Control Test	0.199	12:34
Air Blank	0.000	12:34
Control Test	0.199	12:35
Air Blank	0.000	12:36
Control Test Stats		
Average	0.1987	
Std Dev	0.0006	
Rel Std Dev(%)	0.2906	

P. Murphy
Operator's Signature

P. Murphy
Operator's Signature

P. Murphy
Operator's Signature

ESCAMBIA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006435
07/11/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:38
Control Test	0.081	12:38
Air Blank	0.000	12:38
Control Test	0.081	12:39
Air Blank	0.000	12:39
Control Test	0.081	12:39
Air Blank	0.000	12:40
Control Test Stats		
Average	0.0810	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

D65

SP

P. Murphy
Operator's Signature

7/16/19
JD



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006435, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006435</u>	UNCERTAINTY* ±	
Owning Agency:	<u>ESCAMBIA COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>07/12/2019</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>10:26</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

07/12/2019

Date

Patrick J Murphy

PATRICK J MURPHY,
Department Inspector

SP

FDLE/ATP Form 69 July 2018

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

7/16/19
SP



INSTRUMENT PROCESSING SHEET

Agency Escambia County SO

S/N 80-006435

Florida Department of Law Enforcement

Date In 12/21/18

DI Completion Date 1/4/2019

Ship P/U H/D CMI EE

Intake Performed By <u>SQC</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	Quality Checks Performed By <u>QAM</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>200</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP102</u> 32 mm <u>.152</u> (.139 - .169) 36 mm <u>.167</u> (.156 - .190) 53 mm <u>.234</u> (.228 - .278) 103 mm <u>.511</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>30793</u> <input checked="" type="checkbox"/> Stability Checks	Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)															
Final Release Date <p style="text-align: center;">FDLE JAN 07 2019 Alcohol Testing Program</p>	<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD1021</td> <td>201707D 7/25/19</td> </tr> <tr> <td>0.080</td> <td>DR1275</td> <td>201707E 7/25/19</td> </tr> <tr> <td>0.200</td> <td>SD1013</td> <td>201707C 7/24/19</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG805701 2/26/20</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	SD1021	201707D 7/25/19	0.080	DR1275	201707E 7/25/19	0.200	SD1013	201707C 7/24/19	0.080 DGS	N/A	AG805701 2/26/20	Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Temperature Checks Performed By <u>QAM</u> <input checked="" type="checkbox"/> Lab Temp °C <u>21.6</u> External Digital Therm. ID#: <u>300505</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>SD1021</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>DR1275</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>SD1013</u>
Simulator	Serial #	Lot #/Exp															
0.050	SD1021	201707D 7/25/19															
0.080	DR1275	201707E 7/25/19															
0.200	SD1013	201707C 7/24/19															
0.080 DGS	N/A	AG805701 2/26/20															

Calibration Adjustment Performed By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection Performed By <u>QAM</u> Barometric Pressure ID# <u>30793</u> Gauge <u>1016</u> Instrument <u>1014</u> Mouth Alcohol Solution Lot # <u>2018-B</u> Acetone Stock Solution Lot # <u>2018-A</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>611621</td> </tr> <tr> <td>Interferent</td> <td>DR3855</td> </tr> <tr> <td>0.050</td> <td>SD1021</td> </tr> <tr> <td>0.080</td> <td>DR1275</td> </tr> <tr> <td>0.200</td> <td>SD1013</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	611621	Interferent	DR3855	0.050	SD1021	0.080	DR1275	0.200	SD1013
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.300																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number	Lot Number	Expiration																																																										
0.050																																																													
0.080																																																													
0.200																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.000	611621																																																												
Interferent	DR3855																																																												
0.050	SD1021																																																												
0.080	DR1275																																																												
0.200	SD1013																																																												

Notes/Suggested Service: _____

Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Form 40 <input type="checkbox"/> Other _____	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use
<u>J.D. [Signature]</u> 1/7/19 Tech Review / Date	<u>Brith [Signature]</u> 1/7/19 Admin Review / Date

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: ESCAMBIA COUNTY SO
Time of Inspection: 10:21

Date of Inspection: 01/04/2019

Serial Number: 80-006435
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG805701 Exp: 02/26/2020
0.000	0.048	0.080	0.200	0.080
0.000	0.049	0.080	0.200	0.080
0.000	0.048	0.080	0.201	0.080
0.000	0.049	0.080	0.200	0.080
0.000	0.048	0.080	0.200	0.080
0.000	0.049	0.080	0.201	0.080
0.000	0.049	0.080	0.200	0.080
0.000	0.049	0.080	0.201	0.080
0.000	0.049	0.080	0.200	0.080
0.000	0.049	0.080	0.200	0.080

Standard Deviations	0.0004	0.0000	0.0004	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Patrick J Murphy PATRICK J MURPHY
Signature and Printed Name

01/04/2019
Date

1/7/19
BK

ESCAMBIA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006435
01/04/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:48
Control Test	0.049	07:49
Air Blank	0.000	07:49
Control Test	0.048	07:50
Air Blank	0.000	07:51
Control Test	0.048	07:51
Air Blank	0.000	07:52
Control Test Stats		
Average	0.0483	
Std Dev	0.0006	
Rel Std Dev(%)	1.1945	

P Murphy
Operator's Signature

ESCAMBIA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006435
01/04/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:58
Control Test	0.079	07:58
Air Blank	0.000	07:59
Control Test	0.079	08:00
Air Blank	0.000	08:00
Control Test	0.080	08:01
Air Blank	0.000	08:01
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

P Murphy
Operator's Signature

ESCAMBIA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006435
01/04/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:03
Control Test	0.200	08:03
Air Blank	0.000	08:04
Control Test	0.199	08:05
Air Blank	0.000	08:05
Control Test	0.200	08:06
Air Blank	0.000	08:06
Control Test Stats		
Average	0.1997	
Std Dev	0.0006	
Rel Std Dev(%)	0.2892	

P Murphy
Operator's Signature

ESCAMBIA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006435
01/04/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:07
Control Test	0.081	08:08
Air Blank	0.000	08:08
Control Test	0.081	08:09
Air Blank	0.000	08:09
Control Test	0.080	08:09
Air Blank	0.000	08:10
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

DGS

P Murphy
Operator's Signature

1/7/19
JK



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006435, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006435</u>	UNCERTAINTY* ±
Owning Agency:	<u>ESCAMBIA COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>01/04/2019</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>10:21</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

01/04/2019

Date

Patrick J Murphy

PATRICK J MURPHY,
Department Inspector

FDLE/ATP Form 69 July 2018

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

11/7/19
JP
SK

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ESCAMBIA COUNTY SO
Time of Inspection: 08:51

Date of Inspection: 01/03/2019

Serial Number: 80-006435
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: _____

Remarks:
SKIPPED AI TO OPERATE INSTRUMENT

NOT A COMPLIANCE CHECK

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Patrick J Murphy

PATRICK J MURPHY

Signature and Printed Name

01/03/2019
Date

1/7/19
BK

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: Escambia SO Serial Number: 80-006435
 Time of Inspection: 0725 Date of Inspection: 12/19/2018 Software:

CHECK OR TEST	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		
Alcohol Free Subject Test: 0.000		
Mouth Alcohol Test: Slope Not Met		
Interferent Detect Test: Interferent Detect		
Diagnostic Check (Post-Inspection): OK		

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: _____


Remarks:

Attempted to change password on instrument before inspections. The instrument does not accept the new password and the password is now unknown. Thus, Agency Inspectors are locked out of the instrument to complete this month's Agency Inspection. The instrument was sent to ATP/FDLE on 12/19/18 for a software update.

Wanted to verify compliance.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


James A. Varies

 Signature and Printed Name

12/19/2018
Date

12/19
JAV
BK