FDLE
Florida Department of Law Enforcement

INSTRUMENT PROCESSING SHEET

Agency Florida Highway Patrol Miami S/N 80-006626

Date In 01/07/2019 DI Completion Date 01/07/2019 ☑Ship ☐P/U ☐H/D ☐CMI ☐EE

Intake Performed By DERR

- ☑ Annual
- ☐ Registration
- ☐ Return from CMI / EE

Visual Inspection:

- ☑ Case ☑ Handle
- ☑ Keyboard ☑ Dry Gas Shelf
- ☑ Feet ☑ Breath Tube
- ☑ Ports ☑ Screws Tight

Other Equipment/ Accessories:

- ☐ Power cord ☐ Printer Cable
- ☑ Static Bag ☐ 12V DC Cable

Notes: _____

Final Release Date

FDLE

JAN 15 2019

Alcohol Testing Program

Quality Checks Performed By DERR

- ☑ Breath Tube Screen
- ☑ Replace External O-Rings
- ☑ Instrument Set Up Verified
- ☑ R-Value 211
- ☑ Flow Verification (L/s)

Flow Column # ATP 106

32 mm	.156	(.139 - .169)
36 mm	.183	(.156 - .190)
53 mm	.246	(.228 - .278)
103 mm	.503	(.447 - .547)

- ☑ Barometric Pressure Check

Gauge ID # 28663

- ☑ Stability Checks

Simulator	Serial #	Lot #/Exp
0.050	SD3967	201707D 07/25/2019
0.080	SD3968	201707E 07/25/2019
0.200	SD3969	201707C 07/24/2019
0.080 DGS	N/A	AG805701 02/26/2020

Flow Calibration Performed By _____

Flow Column # _____

- ☐ 5L/min – 17mm
- ☐ 15L/min – 53mm
- ☐ 30L/min – 103mm
- ☐ R-Value _____
- ☐ Post Calibration Verification (L/s)

Flow Column # _____

32 mm		(.139 - .169)
36 mm		(.156 - .190)
53 mm		(.228 - .278)
103 mm		(.447 - .547)

Maintenance Performed By DERR

- ☐ Battery Replacement
- ☐ Dry Gas Regulator Replacement
- ☐ Breath Tube Replacement
- ☑ Other Form Load/Change Pass

Temperature Checks Performed By DERR

- ☑ Lab Temp °C 22.07C

External Digital Therm. ID#: 300918

- ☑ 34°C +-.2 Serial #: SD3967
- ☑ 34°C +-.2 Serial #: SD3968
- ☑ 34°C +-.2 Serial #: SD3969

Calibration Adjustment Performed By _____

Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

☐ Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

Department Inspection Performed By DERR

Barometric Pressure ID# 68639

Gauge 1020 Instrument 1019

Mouth Alcohol Solution Lot # 2017-B

Acetone Stock Solution Lot # 2018-A

Simulator	Serial Number
0.000	SD3965
Interferent	SD3966
0.050	SD3967
0.080	SD3968
0.200	SD3969

Attachments

- ☑ Form 41
- ☑ Stability Checks
- ☑ Calibration Certificate
- ☐ Calibration Adjustment
- ☐ Post-Stability Checks
- ☐ Flow Calibration
- ☐ Form 40
- ☐ Other _____

Notes/Suggested Service: E-mailed ☑ APPROVED

Approved date 1/8/2019

DERR

Digitally signed by DERR
Date: 2020.05.20 13:43:40 -04'00'

- ☑ **Instrument Complies with Chapter 11D-8, FAC**
- ☐ **Instrument Does Not Comply with Chapter 11D-8, FAC**
- ☑ **Return to/Place into Evidentiary Use**
- ☐ **Remain Out of Evidentiary Use**
- ☑ **Conduct an Agency Inspection Before Evidentiary Use**

DERR 1/15/19 Tech Review / Date

1/15/19 Admin Review / Date

Florida Department of Law Enforcement
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP MIAMI
Serial Number: 80-006626
Time of Inspection: 15:04
Date of Inspection: 01/07/2019
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG805701 Exp: 02/26/2020
0.000	0.047	0.079	0.198	0.081
0.000	0.048	0.079	0.199	0.081
0.000	0.048	0.079	0.199	0.081
0.000	0.049	0.079	0.199	0.081
0.000	0.048	0.079	0.198	0.080
0.000	0.048	0.079	0.199	0.080
0.000	0.048	0.079	0.198	0.080
0.000	0.048	0.079	0.198	0.081
0.000	0.049	0.079	0.197	0.080
0.000	0.049	0.079	0.198	0.080

Standard Deviations	0.0006	0.0000	0.0006	0.0005

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David E Reyes Rivera — DAVID E REYES-RIVERA
Signature and Printed Name

01/07/2019
Date

1/15/19

FDLE/ATP Form 41 –Revised August 2005

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-006626	Florida Highway Patrol Miami	01/07/2019	DCL

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
SN: SD3967 Temp: 34.07c	SN: SD3968 Temp: 34.07c	SN: SD3969 Temp:34.07c	Lot AG805701
0.047 to 0.053 ☑	0.077 to 0.083 ☑	0.194 to 0.206 ☑	0.077 to 0.083 ☑

FHP MIAMI
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006626
01/07/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:15
Control Test	0.048	12:16
Air Blank	0.000	12:16
Control Test	0.048	12:17
Air Blank	0.000	12:18
Control Test	0.048	12:18
Air Blank	0.000	12:19
Control Test Stats		
Average	0.0480	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DCL
Operator's Signature

FHP MIAMI
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006626
01/07/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:21
Control Test	0.079	12:22
Air Blank	0.000	12:22
Control Test	0.079	12:23
Air Blank	0.000	12:23
Control Test	0.079	12:24
Air Blank	0.000	12:25
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DCL
Operator's Signature

FHP MIAMI
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006626
01/07/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:27
Control Test	0.197	12:27
Air Blank	0.000	12:28
Control Test	0.198	12:29
Air Blank	0.000	12:29
Control Test	0.199	12:30
Air Blank	0.000	12:30
Control Test Stats		
Average	0.1980	
Std Dev	0.0010	
Rel Std Dev(%)	0.5051	

DCL
Operator's Signature

FHP MIAMI
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006626
01/07/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:32
Control Test	0.081	12:32
Air Blank	0.000	12:33
Control Test	0.081	12:33
Air Blank	0.000	12:33
Control Test	0.081	12:34
Air Blank	0.000	12:34
Control Test Stats		
Average	0.0810	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DCL
Operator's Signature

1/7/19

PGM

Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006626 , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	80-006626
Owning Agency:	FHP MIAMI
Calibration Date:	01/07/2019
Calibration Time:	15:04

	UNCERTAINTY* ±
0.050 g/ 210 L	0.004
0.080 g/ 210 L	0.004
0.200 g/ 210 L	0.007
0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

TRACEABILITY INFORMATION
This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

01/07/2019
Date

DAVID E REYES-RIVERA,
Department Inspector

11/5/19

FDLE
Florida Department of Law Enforcement

INSTRUMENT PROCESSING SHEET

Agency Florida Highway Patrol Miami S/N 80-006626

Date In 01/07/2019 DI Completion Date 01/07/2019 ☑Ship ☐P/U ☐H/D ☐CMI ☐EE

Intake Performed By DLR

☑ Annual
☐ Registration
☐ Return from CMI / EE

Visual Inspection:
☑ Case ☑ Handle
☑ Keyboard ☑ Dry Gas Shelf
☑ Feet ☑ Breath Tube
☑ Ports ☑ Screws Tight

Other Equipment/ Accessories:
☐ Power cord ☐ Printer Cable
☑ Static Bag ☐ 12V DC Cable

Notes: _____

Final Release Date

FDLE

JAN 15 2019

Alcohol Testing Program

Quality Checks Performed By DLR

☑ Breath Tube Screen
☑ Replace External O-Rings
☑ Instrument Set Up Verified
☑ R-Value 211
☑ Flow Verification (L/s)
Flow Column # ATP 106
32 mm .156 (.139 - .169)
36 mm .183 (.156 - .190)
53 mm .246 (.228 - .278)
103 mm .503 (.447 - .547)
☑ Barometric Pressure Check
Gauge ID # 28663
☑ Stability Checks

Simulator	Serial #	Lot #/Exp
0.050	SD3967	201707D 07/25/2019
0.080	SD3968	201707E 07/25/2019
0.200	SD3969	201707C 07/24/2019
0.080 DGS	N/A	AG805701 02/26/2020

Flow Calibration Performed By _____

Flow Column # _____
☐ 5L/min – 17mm
☐ 15L/min – 53mm
☐ 30L/min – 103mm
☐ R-Value _____
☐ Post Calibration Verification (L/s)
Flow Column # _____
32 mm _____ (.139 - .169)
36 mm _____ (.156 - .190)
53 mm _____ (.228 - .278)
103 mm _____ (.447 - .547)

Maintenance Performed By DLR

☐ Battery Replacement
☐ Dry Gas Regulator Replacement
☐ Breath Tube Replacement
☑ Other Form Load/Change Pass

Temperature Checks Performed By DLR

☑ Lab Temp °C 22.07C
External Digital Therm. ID#: 300918
☑ 34°C +-.2 Serial #: SD3967
☑ 34°C +-.2 Serial #: SD3968
☑ 34°C +-.2 Serial #: SD3969

Calibration Adjustment Performed By _____

Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

☐ Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

Department Inspection Performed By DLR

Barometric Pressure ID# 68639
Gauge 1020 Instrument 1019
Mouth Alcohol Solution Lot # 2017-B
Acetone Stock Solution Lot # 2018-A

Simulator	Serial Number
0.000	SD3965
Interferent	SD3966
0.050	SD3967
0.080	SD3968
0.200	SD3969

Attachments

☑ Form 41	☐ Post-Stability Checks
☑ Stability Checks	☐ Flow Calibration
☑ Calibration Certificate	☐ Form 40
☐ Calibration Adjustment	☐ Other _____

Notes/Suggested Service: E-mailed ☑ APPROVED

☑ **Instrument Complies with Chapter 11D-8, FAC**
☐ **Instrument Does Not Comply with Chapter 11D-8, FAC**
☑ **Return to/Place into Evidentiary Use**
☐ **Remain Out of Evidentiary Use**
☑ **Conduct an Agency Inspection Before Evidentiary Use**

DLR 1/15/19 Tech Review / Date

Admin Review / Date 1/15/19

Florida Department of Law Enforcement
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP MIAMI
Serial Number: 80-006626
Time of Inspection: 15:04
Date of Inspection: 01/07/2019
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG805701 Exp: 02/26/2020
0.000	0.047	0.079	0.198	0.081
0.000	0.048	0.079	0.199	0.081
0.000	0.048	0.079	0.199	0.081
0.000	0.049	0.079	0.199	0.081
0.000	0.048	0.079	0.198	0.080
0.000	0.048	0.079	0.199	0.080
0.000	0.048	0.079	0.198	0.080
0.000	0.048	0.079	0.198	0.081
0.000	0.049	0.079	0.197	0.080
0.000	0.049	0.079	0.198	0.080

Standard Deviations	0.0006	0.0000	0.0006	0.0005

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

PJM

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David E Reyes Rivera — DAVID E REYES-RIVERA

Signature and Printed Name

01/07/2019
Date

1/15/19
JO

FDLE/ATP Form 41 –Revised August 2005

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-006626	Florida Highway Patrol Miami	01/07/2019	DCLL

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
SN: SD3967 Temp: 34.07c	SN: SD3968 Temp: 34.07c	SN: SD3969 Temp:34.07c	Lot AG805701
0.047 to 0.053 ☑	0.077 to 0.083 ☑	0.194 to 0.206 ☑	0.077 to 0.083 ☑

FHP MIAMI
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006626
01/07/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:15
Control Test	0.048	12:16
Air Blank	0.000	12:16
Control Test	0.048	12:17
Air Blank	0.000	12:18
Control Test	0.048	12:18
Air Blank	0.000	12:19
Control Test Stats		
Average	0.0480	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DCLL
Operator's Signature

11/5/19

FHP MIAMI
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006626
01/07/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:21
Control Test	0.079	12:22
Air Blank	0.000	12:22
Control Test	0.079	12:23
Air Blank	0.000	12:23
Control Test	0.079	12:24
Air Blank	0.000	12:25
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DCLL
Operator's Signature

FHP MIAMI
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006626
01/07/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:27
Control Test	0.197	12:27
Air Blank	0.000	12:28
Control Test	0.198	12:29
Air Blank	0.000	12:29
Control Test	0.199	12:30
Air Blank	0.000	12:30
Control Test Stats		
Average	0.1980	
Std Dev	0.0010	
Rel Std Dev(%)	0.5051	

DCLL
Operator's Signature

FHP MIAMI
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006626
01/07/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:32
Control Test	0.081	12:32
Air Blank	0.000	12:33
Control Test	0.081	12:33
Air Blank	0.000	12:33
Control Test	0.081	12:34
Air Blank	0.000	12:34
Control Test Stats		
Average	0.0810	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DCLL
Operator's Signature

PGM

Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006626 , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	80-006626
Owning Agency:	FHP MIAMI
Calibration Date:	01/07/2019
Calibration Time:	15:04

	UNCERTAINTY* ±
0.050 g/ 210 L	0.004
0.080 g/ 210 L	0.004
0.200 g/ 210 L	0.007
0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

TRACEABILITY INFORMATION
This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

01/07/2019
Date

DAVID E REYES-RIVERA,
Department Inspector

11/5/19