

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: Fort Walton Beach Police Dep  
Time of Inspection:

Date of Inspection:

Serial Number: 80-006931  
Software:

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK			Date and/or Time Adjusted		
Minimum Sample Volume Check: OK			Barometric Pressure Sensor Check: OK		
Alcohol Free Subject Test: 0.000			Mouth Alcohol Test: Slope Not Met		
Interferent Detect Test: Interferent Detect			Diagnostic Check (Post-Inspection): OK		

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Standard Deviations				
---------------------	--	--	--	--

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: \_\_\_\_\_ Number of Simulators Used: \_\_\_\_\_

Remarks:  
At CMI

As of 12/31/19  
YBK

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.  
I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

\_\_\_\_\_  
Signature and Printed Name

\_\_\_\_\_  
Date