



INSTRUMENT PROCESSING SHEET

Agency Ft. Lauderdale PDS/N 80-006939Florida Department of
Law EnforcementDate In 01/18/19DI Completion Date 1/22/19 Ship P/U H/D CMI EE

Intake Performed By <u>SQC</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	Quality Checks Performed By <u>JD</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>162</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP-105</u> 32 mm <u>.152</u> (.139 - .169) 36 mm <u>.171</u> (.156 - .190) 53 mm <u>.250</u> (.228 - .278) 103 mm <u>.523</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28421</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse; margin-top: 5px;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>SD1012</u></td> <td><u>201707D</u> <u>7/25/19</u></td> </tr> <tr> <td>0.080</td> <td><u>DR1279</u></td> <td><u>201707E</u> <u>7/25/19</u></td> </tr> <tr> <td>0.200</td> <td><u>DR3856</u></td> <td><u>201707C</u> <u>7/24/19</u></td> </tr> <tr> <td>0.080 DGS</td> <td><u>N/A</u></td> <td><u>AG805701</u> <u>2/26/20</u></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	<u>SD1012</u>	<u>201707D</u> <u>7/25/19</u>	0.080	<u>DR1279</u>	<u>201707E</u> <u>7/25/19</u>	0.200	<u>DR3856</u>	<u>201707C</u> <u>7/24/19</u>	0.080 DGS	<u>N/A</u>	<u>AG805701</u> <u>2/26/20</u>	Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Temperature Checks Performed By <u>JD</u> <input checked="" type="checkbox"/> Lab Temp °C <u>21.6</u> External Digital Therm. ID#: <u>300504</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD1012</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>DR1279</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>DR3856</u>																																												
Simulator	Serial #	Lot #/Exp																																																											
0.050	<u>SD1012</u>	<u>201707D</u> <u>7/25/19</u>																																																											
0.080	<u>DR1279</u>	<u>201707E</u> <u>7/25/19</u>																																																											
0.200	<u>DR3856</u>	<u>201707C</u> <u>7/24/19</u>																																																											
0.080 DGS	<u>N/A</u>	<u>AG805701</u> <u>2/26/20</u>																																																											
Final Release Date <p style="text-align: center; font-size: 1.2em;">FDLE</p> <p style="text-align: center; font-size: 1.2em;">JAN 22 2019</p> <p style="text-align: center;">Alcohol Testing Program</p>																																																													
Calibration Adjustment Performed By <u>JD</u> Barometric Pressure Gauge <u>281028</u> ID # <u>28662</u> <table border="1" style="width:100%; border-collapse: collapse; margin-bottom: 5px;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>G2834</u></td> <td><u>N/A</u></td> <td><u>N/A</u></td> </tr> <tr> <td>0.040</td> <td><u>SD1022</u></td> <td><u>17410</u></td> <td><u>12/6/19</u></td> </tr> <tr> <td>0.100</td> <td><u>SD3964</u></td> <td><u>18070</u></td> <td><u>2/26/20</u></td> </tr> <tr> <td>0.200</td> <td><u>SD1025</u></td> <td><u>17340</u></td> <td><u>10/9/19</u></td> </tr> <tr> <td>0.300</td> <td><u>SD1024</u></td> <td><u>18110</u></td> <td><u>4/2/20</u></td> </tr> <tr> <td>0.080 DGS</td> <td><u>N/A</u></td> <td><u>17817080A2</u></td> <td><u>8/5/19</u></td> </tr> </tbody> </table> <input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse; margin-bottom: 5px;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>SD1012</u></td> <td><u>201707D</u></td> <td><u>7/25/19</u></td> </tr> <tr> <td>0.080</td> <td><u>DR1279</u></td> <td><u>201707E</u></td> <td><u>7/25/19</u></td> </tr> <tr> <td>0.200</td> <td><u>DR3856</u></td> <td><u>201707C</u></td> <td><u>7/24/19</u></td> </tr> <tr> <td>0.080 DGS</td> <td><u>N/A</u></td> <td><u>AG805701</u></td> <td><u>2/26/20</u></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000	<u>G2834</u>	<u>N/A</u>	<u>N/A</u>	0.040	<u>SD1022</u>	<u>17410</u>	<u>12/6/19</u>	0.100	<u>SD3964</u>	<u>18070</u>	<u>2/26/20</u>	0.200	<u>SD1025</u>	<u>17340</u>	<u>10/9/19</u>	0.300	<u>SD1024</u>	<u>18110</u>	<u>4/2/20</u>	0.080 DGS	<u>N/A</u>	<u>17817080A2</u>	<u>8/5/19</u>	Simulator	Serial Number	Lot Number	Expiration	0.050	<u>SD1012</u>	<u>201707D</u>	<u>7/25/19</u>	0.080	<u>DR1279</u>	<u>201707E</u>	<u>7/25/19</u>	0.200	<u>DR3856</u>	<u>201707C</u>	<u>7/24/19</u>	0.080 DGS	<u>N/A</u>	<u>AG805701</u>	<u>2/26/20</u>	Department Inspection Performed By <u>JD</u> Barometric Pressure ID# <u>26932</u> Gauge <u>1028</u> Instrument <u>1029</u> Mouth Alcohol Solution Lot # <u>2018-B</u> Acetone Stock Solution Lot # <u>2018-A</u> <table border="1" style="width:100%; border-collapse: collapse; margin-bottom: 5px;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>G2408</u></td> </tr> <tr> <td>Interferent</td> <td><u>G2882</u></td> </tr> <tr> <td>0.050</td> <td><u>SD1012</u></td> </tr> <tr> <td>0.080</td> <td><u>DR1279</u></td> </tr> <tr> <td>0.200</td> <td><u>DR3856</u></td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	<u>G2408</u>	Interferent	<u>G2882</u>	0.050	<u>SD1012</u>	0.080	<u>DR1279</u>	0.200	<u>DR3856</u>
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000	<u>G2834</u>	<u>N/A</u>	<u>N/A</u>																																																										
0.040	<u>SD1022</u>	<u>17410</u>	<u>12/6/19</u>																																																										
0.100	<u>SD3964</u>	<u>18070</u>	<u>2/26/20</u>																																																										
0.200	<u>SD1025</u>	<u>17340</u>	<u>10/9/19</u>																																																										
0.300	<u>SD1024</u>	<u>18110</u>	<u>4/2/20</u>																																																										
0.080 DGS	<u>N/A</u>	<u>17817080A2</u>	<u>8/5/19</u>																																																										
Simulator	Serial Number	Lot Number	Expiration																																																										
0.050	<u>SD1012</u>	<u>201707D</u>	<u>7/25/19</u>																																																										
0.080	<u>DR1279</u>	<u>201707E</u>	<u>7/25/19</u>																																																										
0.200	<u>DR3856</u>	<u>201707C</u>	<u>7/24/19</u>																																																										
0.080 DGS	<u>N/A</u>	<u>AG805701</u>	<u>2/26/20</u>																																																										
Simulator	Serial Number																																																												
0.000	<u>G2408</u>																																																												
Interferent	<u>G2882</u>																																																												
0.050	<u>SD1012</u>																																																												
0.080	<u>DR1279</u>																																																												
0.200	<u>DR3856</u>																																																												
Notes/Suggested Service: <u>Please change level 2 password to something unique. JD</u> _____ _____ _____	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use <u>SP 1/22/19</u> <u>Britt Kirkland 1/23/19</u> Tech Review / Date Admin Review / Date																																																												

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FORT LAUDERDALE PD
Time of Inspection: 09:42

Date of Inspection: 01/22/2019

Serial Number: 80-006939
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG805701 Exp: 02/26/2020
0.000	0.049	0.081	0.199	0.079
0.000	0.049	0.081	0.200	0.078
0.000	0.050	0.080	0.200	0.079
0.000	0.049	0.080	0.200	0.079
0.000	0.049	0.079	0.199	0.078
0.000	0.050	0.080	0.200	0.078
0.000	0.049	0.080	0.200	0.078
0.000	0.049	0.079	0.200	0.079
0.000	0.050	0.080	0.199	0.079
0.000	0.049	0.080	0.200	0.078

Standard Deviations	0.0004	0.0006	0.0004	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

SP
BK
1/22/19

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

TJ Graham

THOMAS J GRAHAM
Signature and Printed Name

01/22/2019
Date

80-006939

1/18/19 JD

INTOXILYZER 8000
Instrument Initialization
11:11 01/18/2019

FORT LAUDERDALE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006939
01/18/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:07
Control Test	0.049	12:08
Air Blank	0.000	12:08
Control Test	0.048	12:09
Air Blank	0.000	12:10
Control Test	0.048	12:10
Air Blank	0.000	12:11
Control Test Stats		
Average	0.0483	
Std Dev	0.0006	
Rel Std Dev(%)	1.1945	

JD

Operator's Signature

FORT LAUDERDALE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006939
01/18/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:13
Control Test	0.078	12:13
Air Blank	0.000	12:14
Control Test	0.078	12:15
Air Blank	0.000	12:15
Control Test	0.078	12:16
Air Blank	0.000	12:16
Control Test Stats		
Average	0.0780	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

JD

Operator's Signature

FORT LAUDERDALE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006939
01/18/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:18
Control Test	0.192	12:18
Air Blank	0.000	12:19
Control Test	0.193	12:20
Air Blank	0.000	12:20
Control Test	0.193	12:21
Air Blank	0.000	12:21
Control Test Stats		
Average	0.1927	
Std Dev	0.0006	
Rel Std Dev(%)	0.2997	

JD

Operator's Signature

FORT LAUDERDALE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006939
01/18/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:23
Control Test	0.077	12:23
Air Blank	0.000	12:23
Control Test	0.077	12:24
Air Blank	0.000	12:24
Control Test	0.077	12:25
Air Blank	0.000	12:25
Control Test Stats		
Average	0.0770	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

JD

Operator's Signature

SP BK 1/22/19



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006939, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006939</u>	UNCERTAINTY* ±	
Owning Agency:	<u>FORT LAUDERDALE PD</u>	0.050 g/210 L	0.004
Calibration Date:	<u>01/22/2019</u>	0.080 g/210 L	0.004
Calibration Time:	<u>09:42</u>	0.200 g/210 L	0.007
		0.080 g/210 L Dry Gas Control	0.005

All results are reported in g/210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

01/22/2019

Date
THOMAS J GRAHAM,
Department Inspector

FDLE/ATP Form 69 July 2018
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

SP BK 1/22/19

1/22/19 80-006939
 Calibration Adjustment

JD

<<<<< CHANNEL 2 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 1.5190 (-0.0040)
 Sample #2 = 1.5270 (0.0090)
 Sample #3 = 1.5050 (0.0230)
 Sample #4 = 1.5170 (0.0340)
 Avg % Abs = 1.5163 (0.0220)
 STD DEV = 0.0110 (0.0125)
 REL STD DEV = 0.726 (56.954)

Sol Value = 0.100 g/210L ***
 Fit Value = 0.4762 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12553, 9um Io = 13141
 <<<<< CHANNEL 1 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 1.8400 (-0.0160)
 Sample #2 = 1.8360 (0.0200)
 Sample #3 = 1.7860 (0.0700)
 Sample #4 = 1.8300 (0.0730)
 Avg % Abs = 1.8173 (0.0543)
 STD DEV = 0.0273 (0.0298)
 REL STD DEV = 1.502 (54.794)

<<<<< CHANNEL 2 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 3.4970 (-0.0200)
 Sample #2 = 3.4930 (-0.0060)
 Sample #3 = 3.4650 (0.0200)
 Sample #4 = 3.4990 (0.0160)
 Avg % Abs = 3.4857 (0.0100)
 STD DEV = 0.0181 (0.0140)
 REL STD DEV = 0.521 (140.000)
 Sol Value = 0.200 g/210L ***
 Fit Value = 0.9524 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12538, 9um Io = 13135

<<<<< CHANNEL 1 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 3.4890 (-0.0040)
 Sample #2 = 3.4610 (0.0320)
 Sample #3 = 3.4670 (0.0500)
 Sample #4 = 3.4700 (0.0680)
 Avg % Abs = 3.4660 (0.0500)
 STD DEV = 0.0046 (0.0180)
 REL STD DEV = 0.132 (36.000)

FORT LAUDERDALE PD
 Intoxilyzer - Alcolho Analyzer
 Model 8000 SN 80-006939
 01/22/2019 07:18:52
 Auto Calibration
 Max Power Res Value = 103
 Auto Range Res Value = 72

Sol Value = 0.000 g/210L ***
 Fit Value = 0.0000 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12595, 9um Io = 13160
 <<<<< CHANNEL 1 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 0.0850 (-0.0140)
 Sample #2 = 0.0320 (0.0680)
 Sample #3 = 0.0540 (0.1140)
 Sample #4 = 0.0200 (0.1590)
 Avg % Abs = 0.0353 (0.1137)
 STD DEV = 0.0172 (0.0455)
 REL STD DEV = 48.802 (40.030)

<<<<< CHANNEL 2 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 0.1730 (-0.0040)
 Sample #2 = 0.1700 (0.0160)
 Sample #3 = 0.1690 (0.0380)
 Sample #4 = 0.1430 (0.0620)
 Avg % Abs = 0.1607 (0.0387)
 STD DEV = 0.0153 (0.0230)
 REL STD DEV = 9.528 (59.501)
 Sol Value = 0.040 g/210L ***
 Fit Value = 0.1905 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12570, 9um Io = 13150

<<<<< CHANNEL 1 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 0.7930 (-0.0060)
 Sample #2 = 0.7700 (0.0330)
 Sample #3 = 0.7570 (0.0620)
 Sample #4 = 0.7810 (0.0860)
 Avg % Abs = 0.7693 (0.0603)
 STD DEV = 0.0120 (0.0265)
 REL STD DEV = 1.562 (43.988)

<<<<< CHANNEL 1 >>>>>
 Sol Val = 0.0000 mg/l or 0.000 g/210L
 % Abs = 0.035
 Std Dev = 0.02 Rel Std Dev = 48.80
 Sol Val = 0.1905 mg/l or 0.040 g/210L
 % Abs = 0.769
 Std Dev = 0.01 Rel Std Dev = 1.56
 Sol Val = 0.4762 mg/l or 0.100 g/210L
 % Abs = 1.817
 Std Dev = 0.03 Rel Std Dev = 1.50
 Sol Val = 0.9524 mg/l or 0.200 g/210L
 % Abs = 3.466
 Std Dev = 0.00 Rel Std Dev = 0.13
 Sol Val = 1.4286 mg/l or 0.300 g/210L
 % Abs = 5.140
 Std Dev = 0.02 Rel Std Dev = 0.41
 Zero Order Coef = -133.43
 First Order Coef = 2673.27
 Second Order Coef = 26.64
 Standard Deviation = 52.555611

<<<<< CHANNEL 2 >>>>>
 Sol Val = 0.0000 mg/l or 0.000 g/210L
 % Abs = 0.161
 Std Dev = 0.02 Rel Std Dev = 9.53
 Sol Val = 0.1905 mg/l or 0.040 g/210L
 % Abs = 1.516
 Std Dev = 0.01 Rel Std Dev = 0.73
 Sol Val = 0.4762 mg/l or 0.100 g/210L
 % Abs = 3.466
 Std Dev = 0.02 Rel Std Dev = 0.52
 Sol Val = 0.9524 mg/l or 0.200 g/210L
 % Abs = 6.557
 Std Dev = 0.02 Rel Std Dev = 0.24
 Sol Val = 1.4286 mg/l or 0.300 g/210L
 % Abs = 9.531
 Std Dev = 0.01 Rel Std Dev = 0.13
 Zero Order Coef = -235.21
 First Order Coef = 1393.10
 Second Order Coef = 13.80
 Standard Deviation = 21.685789

<<<<< CHANNEL 1 >>>>>
 Sol Value = 0.080 g/210L ***
 Fit Value = 0.3810 mg/l %%%
 Samples Taken = 4, Discarded = 1
 <<<<< CHANNEL 2 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 3503.00
 Sample #2 = 3460.00
 Sample #3 = 3444.00
 Sample #4 = 3362.00
 Avg % Abs = 3422.0000
 STD DEV = 52.5738
 REL STD DEV = 1.536
 <<<<< CHANNEL 1 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 3343.00
 Sample #2 = 3353.00
 Sample #3 = 3367.00
 Sample #4 = 3359.00
 Avg % Abs = 3359.6667
 STD DEV = 7.0238
 REL STD DEV = 0.209

<<<<< CHANNEL 2 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 9.5460 (0.0040)
 Sample #2 = 9.5400 (0.0380)
 Sample #3 = 9.5170 (0.0580)
 Sample #4 = 9.5360 (0.0600)
 Avg % Abs = 9.5310 (0.0520)
 STD DEV = 0.0123 (0.0122)
 REL STD DEV = 0.129 (23.395)

Solution Stats Quadratic Fit Chan 1

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	-0.001	0.0008
0.040	0.041	-0.0007
0.100	0.101	-0.0011
0.200	0.198	0.0015
0.300	0.301	-0.0005

Solution Stats Quadratic Fit Chan 2

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	-0.000	0.0002
0.040	0.040	-0.0001
0.100	0.101	-0.0006
0.200	0.199	0.0006
0.300	0.300	-0.0002

 Dry Gas H2O Adjust Results *****
 Barometric Pressure = 1028
 3 um H2O Adjust (mg/l*10.000) = 387
 9 um H2O Adjust (mg/l*10.000) = 450

 ***** AUTO CAL PASS *****

SP BK 1/22/19

80-006939

Post Calibration Adjustment Stabilities

1/22/19
JD

FORT LAUDERDALE PD
Intoxilyzer - Alcohol Analyzer
Model 8000
01/22/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:56
Control Test	0.049	07:57
Air Blank	0.000	07:58
Control Test	0.050	07:58
Air Blank	0.000	07:59
Control Test	0.050	08:00
Air Blank	0.000	08:00
Control Test Stats		
Average	0.0497	
Std Dev	0.0006	
Rel. Std Dev(%)	1.1625	

JD
Operator's Signature

FORT LAUDERDALE PD
Intoxilyzer - Alcohol Analyzer
Model 8000
01/22/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:01
Control Test	0.080	08:02
Air Blank	0.000	08:03
Control Test	0.080	08:03
Air Blank	0.000	08:04
Control Test	0.080	08:05
Air Blank	0.000	08:05
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel. Std Dev(%)	0.0000	

JD
Operator's Signature

FORT LAUDERDALE PD
Intoxilyzer - Alcohol Analyzer
Model 8000
01/22/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:06
Control Test	0.199	08:07
Air Blank	0.000	08:08
Control Test	0.199	08:08
Air Blank	0.000	08:09
Control Test	0.199	08:09
Air Blank	0.000	08:10
Control Test Stats		
Average	0.1990	
Std Dev	0.0000	
Rel. Std Dev(%)	0.0000	

JD
Operator's Signature

FORT LAUDERDALE PD
Intoxilyzer - Alcohol Analyzer
Model 8000
01/22/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:11
Control Test	0.079	08:11
Air Blank	0.000	08:12
Control Test	0.079	08:12
Air Blank	0.000	08:13
Control Test	0.079	08:13
Air Blank	0.000	08:14
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel. Std Dev(%)	0.0000	

DGS

JD
Operator's Signature

SP
BK
1/22/19