INSTRUMENT PROCESSING SHEET

FDLE Florida Department of Law Enforcement

Agency Walton County S/N 80-007173

Date In 8/30/2019 DI Completion Date 9/5/19 ☒Ship ☐P/U ☐H/D ☐CMI ☐EE

Intake Performed By DK

- ☑ Annual
- ☐ Registration
- ☑ Return from CMI / EE

Visual Inspection:

- ☑ Case ☑ Handle
- ☑ Keyboard ☑ Dry Gas Shelf
- ☑ Feet ☑ Breath Tube
- ☑ Ports ☑ Screws Tight

Other Equipment/ Accessories:

- ☑ Power cord ☐ Printer Cable
- ☑ Static Bag ☐ 12V DC Cable

Notes: _____

Final Release Date

FDLE

SEP 06 2019

Alcohol Testing Program

Quality Checks Performed By DKM

- ☒ Breath Tube Screen
- ☐ Replace External O-Rings
- ☒ Instrument Set Up Verified
- ☒ R-Value 183
- ☒ Flow Verification (L/s)

Flow Column # ATP102

32 mm	.156	(.139 - .169)
36 mm	.167	(.156 - .190)
53 mm	.242	(.228 - .278)
103 mm	.503	(.447 - .547)

- ☒ Barometric Pressure Check

Gauge ID # 28427

- ☒ Stability Checks

Simulator	Serial #	Lot #/Exp
0.050	SD1018	201905A 5/14/21
0.080	SD3962	201905B 5/14/21
0.200	G2078	201904D 4/30/21
0.080 DGS	N/A	A6916501 6/14/21

Flow Calibration Performed By _____

Flow Column # _____

- ☐ 5L/min – 17mm
- ☐ 15L/min – 53mm
- ☐ 30L/min – 103mm
- ☐ R-Value _____
- ☐ Post Calibration Verification (L/s)

Flow Column # _____

32 mm	_____	(.139 - .169)
36 mm	_____	(.156 - .190)
53 mm	_____	(.228 - .278)
103 mm	_____	(.447 - .547)

Maintenance Performed By _____

- ☐ Battery Replacement
- ☐ Dry Gas Regulator Replacement
- ☐ Breath Tube Replacement
- ☐ Other _____

Temperature Checks Performed By DKM

- ☒ Lab Temp °C 21.1

External Digital Therm. ID#: 300503

- ☒ 34°C +-.2 Serial #: SD1018
- ☒ 34°C +-.2 Serial #: SD3962
- ☒ 34°C +-.2 Serial #: G2078

Calibration Adjustment Performed By _____

Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

☐ Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

Department Inspection Performed By DKM

Barometric Pressure ID# 28427

Gauge 1009 Instrument 1009

Mouth Alcohol Solution Lot # 2018-B

Acetone Stock Solution Lot # 2019-A

Simulator	Serial Number
0.000	G4444
Interferent	G6621
0.050	SD1018
0.080	SD3962
0.200	G2078

Attachments

- ☒ Form 41
- ☒ Stability Checks
- ☒ Calibration Certificate
- ☐ Calibration Adjustment
- ☐ Post-Stability Checks
- ☐ Flow Calibration
- ☐ Form 40
- ☐ Other _____

Notes/Suggested Service: _____

- ☒ **Instrument Complies with Chapter 11D-8, FAC**
- ☐ **Instrument Does Not Comply with Chapter 11D-8, FAC**
- ☒ **Return to/Place into Evidentiary Use**
- ☐ **Remain Out of Evidentiary Use**
- ☒ **Conduct an Agency Inspection Before Evidentiary Use**

SP 9/10/19 Tech Review / Date

Britt Kirkland 9/6/19 Admin Review / Date

Florida Department of Law Enforcement
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: WALTON COUNTY SO
Time of Inspection: 12:06
Date of Inspection: 09/05/2019
Serial Number: 80-007173
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG916501 Exp: 06/14/2021
0.000	0.048	0.078	0.197	0.080
0.000	0.048	0.079	0.199	0.080
0.000	0.048	0.079	0.199	0.080
0.000	0.048	0.079	0.199	0.080
0.000	0.049	0.079	0.199	0.080
0.000	0.049	0.079	0.199	0.079
0.000	0.048	0.079	0.199	0.080
0.000	0.048	0.079	0.199	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.079	0.199	0.079
Standard Deviations	0.0005	0.0003	0.0006	0.0005

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

SP BK 9/6/19

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Patrick J Murphy
Signature and Printed Name

PATRICK J MURPHY

09/05/2019
Date

FDLE/ATP Form 41 –Revised August 2005

WALTON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007173
09/05/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:20
Control Test	0.047	08:21
Air Blank	0.000	08:21
Control Test	0.047	08:22
Air Blank	0.000	08:22
Control Test	0.047	08:23
Air Blank	0.000	08:24
Control Test Stats		
Average	0.0470	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

P Murphy
Operator's Signature

WALTON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007173
09/05/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:25
Control Test	0.078	08:26
Air Blank	0.000	08:26
Control Test	0.077	08:27
Air Blank	0.000	08:28
Control Test	0.079	08:28
Air Blank	0.000	08:29
Control Test Stats		
Average	0.0780	
Std Dev	0.0010	
Rel Std Dev(%)	1.2821	

P Murphy
Operator's Signature

WALTON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007173
09/05/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:32
Control Test	0.198	08:33
Air Blank	0.000	08:33
Control Test	0.198	08:34
Air Blank	0.000	08:35
Control Test	0.198	08:35
Air Blank	0.000	08:36
Control Test Stats		
Average	0.1980	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

P Murphy
Operator's Signature

WALTON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007173
09/05/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:37
Control Test	0.080	08:38
Air Blank	0.000	08:38
Control Test	0.080	08:39
Air Blank	0.000	08:39
Control Test	0.080	08:40
Air Blank	0.000	08:40
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

P Murphy
Operator's Signature

SP BK 9/6/19

Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007173 , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	80-007173
Owning Agency:	WALTON COUNTY SO
Calibration Date:	09/05/2019
Calibration Time:	12:06

	UNCERTAINTY* ±
0.050 g/ 210 L	0.004
0.080 g/ 210 L	0.004
0.200 g/ 210 L	0.007
0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

TRACEABILITY INFORMATION
This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

09/05/2019
Date

Patrick J Murphy
PATRICK J MURPHY,
Department Inspector

SP
BK
9/6/19

FDLE
Florida Department of Law Enforcement

INSTRUMENT PROCESSING SHEET

Agency Walton County SO S/N 80-007173

Date In 04/10/2019 DI Completion Date 04/24/2019 ☑Ship ☐P/U ☐H/D ☐CMI ☐EE

Intake Performed By DP

- ☑ Annual
- ☐ Registration
- ☐ Return from CMI / EE

Visual Inspection:

- ☑ Case ☑ Handle
- ☑ Keyboard ☑ Dry Gas Shelf
- ☑ Feet ☑ Breath Tube
- ☑ Ports ☑ Screws Tight

Other Equipment/ Accessories:

- ☐ Power cord ☐ Printer Cable
- ☐ Static Bag ☐ 12V DC Cable

Notes: _____

Final Release Date

FDLE

APR 25 2019

Alcohol Testing Program

Quality Checks Performed By JD

- ☑ Breath Tube Screen
- ☑ Replace External O-Rings
- ☑ Instrument Set Up Verified
- ☑ R-Value 179
- ☑ Flow Verification (L/s)

Flow Column # ATP-103

32 mm	.156	(.139 - .169)
36 mm	.175	(.156 - .190)
53 mm	.242	(.228 - .278)
103 mm	.503	(.447 - .547)

- ☑ Barometric Pressure Check

Gauge ID # 28421

- ☑ Stability Checks

Simulator	Serial #	Lot #/Exp
0.050	SD1012	201707D
		07/25/2019
0.080	DR1279	201707E
		07/25/2019
0.200	SD1013	201707C
		07/24/2019
0.080 DGS	N/A	AG831804
		11/14/2020

Flow Calibration Performed By _____

Flow Column # _____

- ☐ 5L/min – 17mm
- ☐ 15L/min – 53mm
- ☐ 30L/min – 103mm
- ☐ R-Value _____
- ☐ Post Calibration Verification (L/s)

Flow Column # _____

32 mm		(.139 - .169)
36 mm		(.156 - .190)
53 mm		(.228 - .278)
103 mm		(.447 - .547)

Maintenance Performed By _____

- ☐ Battery Replacement
- ☐ Dry Gas Regulator Replacement
- ☐ Breath Tube Replacement
- ☐ Other _____

Temperature Checks Performed By JD

- ☑ Lab Temp °C 22.2

External Digital Therm. ID#: 300503

- ☑ 34°C +-.2 Serial #: SD1012
- ☑ 34°C +-.2 Serial #: DR1279
- ☑ 34°C +-.2 Serial #: SD1013

Calibration Adjustment Performed By _____

Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

☐ Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

Department Inspection Performed By JD

Barometric Pressure ID# 28427

Gauge 1015 Instrument 1015

Mouth Alcohol Solution Lot # 2018-B

Acetone Stock Solution Lot # 2018-A

Simulator	Serial Number
0.000	G2408
Interferent	G2882
0.050	SD1012
0.080	DR1279
0.200	SD1013

Attachments

- ☑ Form 41
- ☑ Stability Checks
- ☑ Calibration Certificate
- ☐ Calibration Adjustment
- ☐ Post-Stability Checks
- ☐ Flow Calibration
- ☐ Form 40
- ☐ Other _____

Notes/Suggested Service: _____

- ☑ **Instrument Complies with Chapter 11D-8, FAC**
- ☐ **Instrument Does Not Comply with Chapter 11D-8, FAC**
- ☑ **Return to/Place into Evidentiary Use**
- ☐ **Remain Out of Evidentiary Use**
- ☑ **Conduct an Agency Inspection Before Evidentiary Use**

DDM 4/25/19 Tech Review / Date

Brett Kirkland 4/25/19 Admin Review / Date

Florida Department of Law Enforcement
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: WALTON COUNTY SO
Time of Inspection: 11:30

Date of Inspection: 04/24/2019

Serial Number: 80-007173
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG831804 Exp: 11/14/2020
0.000	0.049	0.080	0.199	0.078
0.000	0.049	0.080	0.199	0.078
0.000	0.049	0.080	0.199	0.078
0.000	0.049	0.080	0.199	0.078
0.000	0.049	0.079	0.199	0.078
0.000	0.049	0.079	0.200	0.078
0.000	0.050	0.080	0.199	0.077
0.000	0.050	0.080	0.199	0.078
0.000	0.050	0.080	0.199	0.078
0.000	0.050	0.080	0.199	0.078
Standard Deviations	0.0005	0.0004	0.0003	0.0003

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

BJM

Remarks:

BK
4/25/19

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

TJ Graham

THOMAS J GRAHAM
Signature and Printed Name

04/24/2019
Date

FDLE/ATP Form 41 –Revised August 2005

80-007173

4/24/19

JD

WALTON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007173
04/24/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:21
Control Test	0.048	08:22
Air Blank	0.000	08:22
Control Test	0.049	08:23
Air Blank	0.000	08:24
Control Test	0.049	08:24
Air Blank	0.000	08:25
Control Test Stats		
Average	0.0487	
Std Dev	0.0006	
Rel Std Dev(%)	1.1863	

BBM
BK
4/25/19 JD

Operator's Signature

WALTON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007173
04/24/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:26
Control Test	0.078	08:27
Air Blank	0.000	08:27
Control Test	0.079	08:28
Air Blank	0.000	08:28
Control Test	0.078	08:29
Air Blank	0.000	08:30
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

JD

Operator's Signature

WALTON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007173
04/24/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:31
Control Test	0.196	08:32
Air Blank	0.000	08:32
Control Test	0.197	08:33
Air Blank	0.000	08:34
Control Test	0.197	08:34
Air Blank	0.000	08:35
Control Test Stats		
Average	0.1967	
Std Dev	0.0006	
Rel Std Dev(%)	0.2936	

JD

Operator's Signature

WALTON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007173
04/24/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:36
Control Test	0.078	08:36
Air Blank	0.000	08:37
Control Test	0.078	08:37
Air Blank	0.000	08:38
Control Test	0.078	08:38
Air Blank	0.000	08:39
Control Test Stats		
Average	0.0780	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

JD

Operator's Signature

Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007173 , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	80-007173
Owning Agency:	WALTON COUNTY SO
Calibration Date:	04/24/2019
Calibration Time:	11:30

	UNCERTAINTY* ±
0.050 g/ 210 L	0.004
0.080 g/ 210 L	0.004
0.200 g/ 210 L	0.007
0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

TRACEABILITY INFORMATION
This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

04/24/2019
Date

THOMAS J GRAHAM,
Department Inspector

BK
4/25/19

Issues

① Unable to recall tests

② BTO Diagnostics failed on 4/4/19 so I Attempted inspection on 4/5/19, see attached.

Spoke to Mr. Murphy on 4/5/19

RECEIVED
APR 10 2019
FDLE
Alcohol Testing Program

WALTON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007173
04/05/2019
Software: 8100.27

RECEIVED
APR 10 2019
FDLE
Alcohol Testing Program

| DIAGNOSTICS |

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

RECEIVED
APR 10 2019
FDLE
Alcohol Testing Program

WALTON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007173
04/05/2019
Software: 8100.27

DRY CONTROL TEST

Test	g/210L	Time
Air Blank	0.000	09:07
Control Test	0.078	09:07
Air Blank	0.000	09:08

Operator's Signature

RECEIVED
APR 10 2019
FDLE
Alcohol Testing Program

WALTON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007173
04/05/2019
Software: 8100.27

| WET CONTROL TEST |

.08 g/210L

Test	g/210L	Time
Air Blank	0.000	08:58
Control Test	0.071	08:58
Air Blank	0.000	08:59

Operator's Signature

RECEIVED
APR 10 2019
FDLE
Alcohol Testing Program

WALTON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007173
04/05/2019
Software: 8100.27

| WET CONTROL TEST |

.20 g/210L

Test	g/210L	Time
Air Blank	0.000	08:56
Control Test	0.190	08:56
Air Blank	0.000	08:57

Operator's Signature

Safety Solutions
3249 Edinburgh Castle Drive
Milton, FL 32583

RECEIVED
APR 10 2019
FDLE
Alcohol Testing Program

Temp	Time	Date
33.97 C ,	06:05 pm	Feb/16/2019
33.98 C ,	06:05 pm	Feb/16/2019
33.99 C ,	06:06 pm	Feb/16/2019
33.98 C ,	06:06 pm	Feb/16/2019
33.97 C ,	06:07 pm	Feb/16/2019
33.97 C ,	06:07 pm	Feb/16/2019
33.97 C ,	06:08 pm	Feb/16/2019
33.97 C ,	06:08 pm	Feb/16/2019
33.98 C ,	06:09 pm	Feb/16/2019
33.98 C ,	06:09 pm	Feb/16/2019
33.98 C ,	06:10 pm	Feb/16/2019
33.99 C ,	06:10 pm	Feb/16/2019
33.97 C ,	06:11 pm	Feb/16/2019
33.98 C ,	06:11 pm	Feb/16/2019
33.98 C ,	06:12 pm	Feb/16/2019
33.98 C ,	06:12 pm	Feb/16/2019
33.98 C ,	06:13 pm	Feb/16/2019
33.99 C ,	06:13 pm	Feb/16/2019
33.98 C ,	06:14 pm	Feb/16/2019
33.99 C ,	06:14 pm	Feb/16/2019
33.99 C ,	06:15 pm	Feb/16/2019
33.99 C ,	06:15 pm	Feb/16/2019
33.99 C ,	06:16 pm	Feb/16/2019
33.99 C ,	06:16 pm	Feb/16/2019
33.98 C ,	06:17 pm	Feb/16/2019
34.02 C ,	06:17 pm	Feb/16/2019
34.01 C ,	06:18 pm	Feb/16/2019
34.01 C ,	06:18 pm	Feb/16/2019
33.99 C ,	06:19 pm	Feb/16/2019
33.99 C ,	06:19 pm	Feb/16/2019
33.99 C ,	06:20 pm	Feb/16/2019
33.97 C ,	06:20 pm	Feb/16/2019
33.97 C ,	06:21 pm	Feb/16/2019
33.97 C ,	06:21 pm	Feb/16/2019
33.97 C ,	06:22 pm	Feb/16/2019
33.97 C ,	06:22 pm	Feb/16/2019
33.98 C ,	06:23 pm	Feb/16/2019
33.96 C ,	06:23 pm	Feb/16/2019
33.96 C ,	06:24 pm	Feb/16/2019
33.97 C ,	06:24 pm	Feb/16/2019

Certificate of Calibration

Manufacturer: Guth

Model# 34c

Serial# G1950

The above listed simulator was calibrated using the Oakton Temp 360 Datalogging RTD Thermometer Model, Serial 917535, which is NIST (National Institute of Standards and Technology) traceable. The standards used during calibration meet and/or exceed the manufacturer's requirements for proper calibration methods.

Technician Signature

February 16, 2019
Date of Calibration

Standard Deviation

0.0127

Safety Solutions
3249 Edinburgh Castle Drive
Milton, FL 32583

RECEIVED
APR 10 2019
FDLE
Alcohol Testing Program

Temp	Time	Date
34.00 C ,	04:15 pm	Feb/16/2019
34.00 C ,	04:15 pm	Feb/16/2019
34.01 C ,	04:16 pm	Feb/16/2019
34.01 C ,	04:16 pm	Feb/16/2019
34.01 C ,	04:17 pm	Feb/16/2019
34.01 C ,	04:17 pm	Feb/16/2019
34.00 C ,	04:18 pm	Feb/16/2019
34.00 C ,	04:18 pm	Feb/16/2019
33.99 C ,	04:19 pm	Feb/16/2019
34.00 C ,	04:19 pm	Feb/16/2019
34.00 C ,	04:20 pm	Feb/16/2019
34.00 C ,	04:20 pm	Feb/16/2019
34.01 C ,	04:21 pm	Feb/16/2019
34.01 C ,	04:21 pm	Feb/16/2019
34.00 C ,	04:22 pm	Feb/16/2019
34.00 C ,	04:22 pm	Feb/16/2019
34.00 C ,	04:23 pm	Feb/16/2019
33.99 C ,	04:23 pm	Feb/16/2019
34.00 C ,	04:24 pm	Feb/16/2019
34.00 C ,	04:24 pm	Feb/16/2019
34.00 C ,	04:25 pm	Feb/16/2019
34.00 C ,	04:25 pm	Feb/16/2019
34.01 C ,	04:26 pm	Feb/16/2019
34.00 C ,	04:26 pm	Feb/16/2019
34.00 C ,	04:27 pm	Feb/16/2019
34.00 C ,	04:27 pm	Feb/16/2019
34.00 C ,	04:28 pm	Feb/16/2019
34.01 C ,	04:28 pm	Feb/16/2019
34.00 C ,	04:29 pm	Feb/16/2019
34.00 C ,	04:29 pm	Feb/16/2019
33.99 C ,	04:30 pm	Feb/16/2019
33.99 C ,	04:30 pm	Feb/16/2019
33.99 C ,	04:31 pm	Feb/16/2019
33.99 C ,	04:31 pm	Feb/16/2019
33.99 C ,	04:32 pm	Feb/16/2019
34.01 C ,	04:32 pm	Feb/16/2019
34.00 C ,	04:33 pm	Feb/16/2019
34.01 C ,	04:33 pm	Feb/16/2019
34.00 C ,	04:34 pm	Feb/16/2019
34.00 C ,	04:34 pm	Feb/16/2019

Certificate of Calibration

Manufacturer: Guth

Model# 34c

Serial# G6483

The above listed simulator was calibrated using the Oakton Temp 360 Datalogging RTD Thermometer Model, Serial 917535, which is NIST (National Institute of Standards and Technology) traceable. The standards used during calibration meet and/or exceed the manufacturer's requirements for proper calibration methods.

Technician Signature

February 16, 2019
Date of Calibration

Standard Deviation

0.0065

Safety Solutions
3249 Edinburgh Castle Drive
Milton, FL 32583

RECEIVED
APR 10 2019
FDLE
Alcohol Testing Program

Temp	Time	Date
33.97 C ,	02:51 pm	Feb/16/2019
33.97 C ,	02:52 pm	Feb/16/2019
33.97 C ,	02:52 pm	Feb/16/2019
33.96 C ,	02:53 pm	Feb/16/2019
33.97 C ,	02:53 pm	Feb/16/2019
33.99 C ,	02:54 pm	Feb/16/2019
33.98 C ,	02:54 pm	Feb/16/2019
33.97 C ,	02:55 pm	Feb/16/2019
33.97 C ,	02:55 pm	Feb/16/2019
33.97 C ,	02:56 pm	Feb/16/2019
33.97 C ,	02:56 pm	Feb/16/2019
33.98 C ,	02:57 pm	Feb/16/2019
33.98 C ,	02:57 pm	Feb/16/2019
33.98 C ,	02:58 pm	Feb/16/2019
33.98 C ,	02:58 pm	Feb/16/2019
33.98 C ,	02:59 pm	Feb/16/2019
33.97 C ,	02:59 pm	Feb/16/2019
33.98 C ,	03:00 pm	Feb/16/2019
33.97 C ,	03:00 pm	Feb/16/2019
33.98 C ,	03:01 pm	Feb/16/2019
33.98 C ,	03:01 pm	Feb/16/2019
33.98 C ,	03:02 pm	Feb/16/2019
33.98 C ,	03:02 pm	Feb/16/2019
33.98 C ,	03:03 pm	Feb/16/2019
33.98 C ,	03:03 pm	Feb/16/2019
33.98 C ,	03:04 pm	Feb/16/2019
33.98 C ,	03:04 pm	Feb/16/2019
33.97 C ,	03:05 pm	Feb/16/2019
33.96 C ,	03:05 pm	Feb/16/2019
33.97 C ,	03:06 pm	Feb/16/2019
33.98 C ,	03:06 pm	Feb/16/2019
33.97 C ,	03:07 pm	Feb/16/2019
33.97 C ,	03:07 pm	Feb/16/2019
33.96 C ,	03:08 pm	Feb/16/2019
33.96 C ,	03:08 pm	Feb/16/2019
33.96 C ,	03:09 pm	Feb/16/2019
33.96 C ,	03:09 pm	Feb/16/2019
33.97 C ,	03:10 pm	Feb/16/2019
33.96 C ,	03:10 pm	Feb/16/2019
33.96 C ,	03:11 pm	Feb/16/2019

Certificate of Calibration

Manufacturer: Guth

Model# 2100

Serial# DR1455

The above listed simulator was calibrated using the Oakton Temp 360 Datalogging RTD Thermometer Model, Serial 917535, which is NIST (National Institute of Standards and Technology) traceable. The standards used during calibration meet and/or exceed the manufacturer's requirements for proper calibration methods.

Technician Signature

February 16, 2019
Date of Calibration

Standard Deviation

0.0080

Safety Solutions
3249 Edinburgh Castle Drive
Milton, FL 32583

RECEIVED
APR 10 2019
FDLE
Alcohol Testing Program

Temp	Time	Date
33.98 C ,	03:29 pm	Feb/16/2019
33.97 C ,	03:29 pm	Feb/16/2019
33.96 C ,	03:30 pm	Feb/16/2019
33.98 C ,	03:30 pm	Feb/16/2019
33.97 C ,	03:31 pm	Feb/16/2019
33.97 C ,	03:31 pm	Feb/16/2019
33.97 C ,	03:32 pm	Feb/16/2019
33.97 C ,	03:32 pm	Feb/16/2019
33.96 C ,	03:33 pm	Feb/16/2019
33.97 C ,	03:33 pm	Feb/16/2019
33.97 C ,	03:34 pm	Feb/16/2019
33.97 C ,	03:34 pm	Feb/16/2019
33.97 C ,	03:35 pm	Feb/16/2019
33.96 C ,	03:35 pm	Feb/16/2019
33.97 C ,	03:36 pm	Feb/16/2019
33.97 C ,	03:36 pm	Feb/16/2019
33.97 C ,	03:37 pm	Feb/16/2019
33.98 C ,	03:37 pm	Feb/16/2019
33.96 C ,	03:38 pm	Feb/16/2019
33.96 C ,	03:38 pm	Feb/16/2019
33.97 C ,	03:39 pm	Feb/16/2019
33.97 C ,	03:39 pm	Feb/16/2019
33.97 C ,	03:40 pm	Feb/16/2019
33.96 C ,	03:40 pm	Feb/16/2019
33.96 C ,	03:41 pm	Feb/16/2019
33.97 C ,	03:41 pm	Feb/16/2019
33.98 C ,	03:42 pm	Feb/16/2019
33.97 C ,	03:42 pm	Feb/16/2019
33.96 C ,	03:43 pm	Feb/16/2019
33.96 C ,	03:43 pm	Feb/16/2019
33.97 C ,	03:44 pm	Feb/16/2019
33.96 C ,	03:44 pm	Feb/16/2019
33.97 C ,	03:45 pm	Feb/16/2019
33.98 C ,	03:45 pm	Feb/16/2019
33.97 C ,	03:46 pm	Feb/16/2019
33.97 C ,	03:46 pm	Feb/16/2019
33.96 C ,	03:47 pm	Feb/16/2019
33.97 C ,	03:47 pm	Feb/16/2019
33.98 C ,	03:48 pm	Feb/16/2019
33.98 C ,	03:48 pm	Feb/16/2019

Certificate of Calibration

Manufacturer: Guth

Model# 2100

Serial# DR4848

The above listed simulator was calibrated using the Oakton Temp 360 Datalogging RTD Thermometer Model, Serial 917535, which is NIST (National Institute of Standards and Technology) traceable. The standards used during calibration meet and/or exceed the manufacturer's requirements for proper calibration methods.

Technician Signature

February 16, 2019
Date of Calibration

Standard Deviation

0.0066

Safety Solutions
3249 Edinburgh Castle Drive
Milton, FL 32583

RECEIVED
APR 10 2019
FDLE
Alcohol Testing Program

Temp	Time	Date
34.01 C ,	02:09 pm	Feb/16/2019
34.01 C ,	02:10 pm	Feb/16/2019
34.01 C ,	02:10 pm	Feb/16/2019
34.00 C ,	02:11 pm	Feb/16/2019
34.01 C ,	02:11 pm	Feb/16/2019
34.01 C ,	02:12 pm	Feb/16/2019
34.01 C ,	02:12 pm	Feb/16/2019
34.02 C ,	02:13 pm	Feb/16/2019
34.01 C ,	02:13 pm	Feb/16/2019
34.02 C ,	02:14 pm	Feb/16/2019
34.00 C ,	02:14 pm	Feb/16/2019
34.01 C ,	02:15 pm	Feb/16/2019
34.01 C ,	02:15 pm	Feb/16/2019
34.01 C ,	02:16 pm	Feb/16/2019
34.01 C ,	02:16 pm	Feb/16/2019
34.01 C ,	02:17 pm	Feb/16/2019
34.01 C ,	02:17 pm	Feb/16/2019
34.01 C ,	02:18 pm	Feb/16/2019
34.02 C ,	02:18 pm	Feb/16/2019
34.02 C ,	02:19 pm	Feb/16/2019
34.01 C ,	02:19 pm	Feb/16/2019
34.01 C ,	02:20 pm	Feb/16/2019
34.01 C ,	02:20 pm	Feb/16/2019
34.01 C ,	02:21 pm	Feb/16/2019
34.01 C ,	02:21 pm	Feb/16/2019
34.01 C ,	02:22 pm	Feb/16/2019
34.02 C ,	02:22 pm	Feb/16/2019
34.01 C ,	02:23 pm	Feb/16/2019
34.01 C ,	02:23 pm	Feb/16/2019
34.01 C ,	02:24 pm	Feb/16/2019
34.01 C ,	02:24 pm	Feb/16/2019
34.00 C ,	02:25 pm	Feb/16/2019
34.00 C ,	02:25 pm	Feb/16/2019
33.99 C ,	02:26 pm	Feb/16/2019
34.01 C ,	02:26 pm	Feb/16/2019
34.01 C ,	02:27 pm	Feb/16/2019
34.01 C ,	02:27 pm	Feb/16/2019
34.01 C ,	02:28 pm	Feb/16/2019
34.02 C ,	02:28 pm	Feb/16/2019
34.01 C ,	02:29 pm	Feb/16/2019

Certificate of Calibration

Manufacturer: Guth

Model# 2100

Serial# DR4943

The above listed simulator was calibrated using the Oakton Temp 360 Datalogging RTD Thermometer Model, Serial 917535, which is NIST (National Institute of Standards and Technology) traceable. The standards used during calibration meet and/or exceed the manufacturer's requirements for proper calibration methods.

Technician Signature

February 16, 2019
Date of Calibration

Standard Deviation

0.0059

RECEIVED
APR 10 2019
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: WALTON COUNTY SO
Time of Inspection: 10:50 Date of Inspection: 03/23/2019
Serial Number: 80-007173
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-T… …OK	Yes	
Alcohol …	Yes	
	Yes	
I… …etect	Yes	
Di…	Yes	

Alcohol Test (g/21…	…10L Test …709A …12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:980558 Exp: 07/09/2021
0.000		0.195	0.079
0.000		0.197	0.079
0.000	0.078	0.198	0.079

Number of … 5

Remarks:

COPY
Placed in service when recieved from FDLE

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JONATHAN A GAYDOS
Signature and Printed Name

03/23/2019
Date

FDLE/ATP Form 40 – March 2004

RECEIVED
APR 10 2019
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: WALTON COUNTY SO
Serial Number: 80-007173
Time of Inspection: 08:47
Date of Inspection: 04/05/2019
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000	0.045	0.072 / 0.072		
0.000	0.046	0.073 / 0.074		
0.000	0.047	0.074 / 0.074		

Number of Simulators Used: 5

Remarks:
08: Control Outside Tolerance, ABORTING . ABORTING, REPLACING 0.08Non-compliance: CONTROL OUTSIDE TOLERANCE.

Out of service, Sending to FDLE

The above instrument complies ( ) does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Jonathan Gaydos — JONATHAN A GAYDOS
Signature and Printed Name

04/05/2019
Date

RECEIVED
APR 10 2019
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: WALTON COUNTY SO
Time of Inspection: 09:11 **Date of Inspection:** 04/05/2019
Serial Number: 80-001312
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:754250 Exp: 04/03/2020
0.000	0.047	0.076	0.197	0.080
0.000	0.048	0.077	0.199	0.080
0.000	0.049	0.077	0.199	0.080

Number of Simulators Used: 5

Remarks:

This is Some Solutions as 80-007173

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JONATHAN A GAYDOS
Signature and Printed Name

04/05/2019
Date

FDLE/ATP Form 40 – March 2004

RECEIVED
APR 10 2019
FDLE
Alcohol Testing Program

FLORIDA DEPARTMENT OF LAW ENFORCEMENT
ALCOHOL TESTING PROGRAM
BREATH ALCOHOL TEST AFFIDAVIT

Instrument Type: Intoxilyzer 8000
Instrument Registered To: WALTON COUNTY SO
Instrument Serial Number: 80-007173 Software: 8100.27
Date of Test: 04/04/2019

Date of Last Agency Inspection: 03/23/2019
Observation Period Began: 02:31
Subject's Name: SONA L SHUKRI DOB: 04/10/1984 Sex: F

The subject was observed for at least twenty-minutes prior to the administration of the breath test to ensure that the subject did not take anything orally and did not regurgitate.

Results:

Test	g/210L	Time
Diagnostics Check	OK	02:56
Air Blank	0.000	02:56
Control Test	0.078	02:57
Air Blank	0.000	02:57
Subject Sample #1	0.197	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample #2	0.189	03:03
Air Blank	0.000	03:04
Control Test	0.077	03:04
Air Blank	0.000	03:05
Diagnostics Check	Fail*	03:05
Air Blank	0.000	03:05

* Voltage/Current Test

Cylinder Lot: 980558
Exp: 07/09/2021

State of Florida, County of Walton.

Personally appeared before me the undersigned authority, who (✓) is personally known to me or (__) produced _____ as identification, and who after being placed under oath, states:

I LUIS E SALAS, hold a valid Breath Test Operator permit issued by the Florida Department of Law Enforcement, I administered the above breath test to the subject named above in accordance with Chapter 11D-8, Florida Administrative Code, and this form is a true and accurate report of that breath test.

Breath Test Operator: _____ #339 Date: 04/04/2019
Signature

Sworn to (or affirmed) before me this 4th day of April, 2019

_____ 451 C. Parsons
Signature of Notary Public-State of Florida Printed Name of Notary Public-State of Florida

Note: Pursuant to section 117.10, Florida Statutes, law enforcement officers, correctional officers, traffic accident investigation officers and traffic infraction enforcement officers are notaries public when engaged in the performance of official duties. In accordance with section 316.1934(5), F.S., this completed form is admissible without further authentication and is presumptive proof of the results herein. To be used in accordance with Section 316.1934(5), F.S., and in administrative proceedings pursuant to 322.2615, F.S.

FDLE REQUEST FOR REGISTRATION

Florida Department of
Law Enforcement

MAKE AND MODEL OF INSTRUMENT: Intoxilyzer 8000

SERIAL NUMBER: 80-007173

OWNING AGENCY: Walton County S.O.

DATE OF DEPARTMENT INSPECTION: 2/12/2019

AGENCY INSPECTOR: William Curtis

ADDRESS: 752 Triple G Road

CITY, STATE, ZIP: DeFuniak Springs, FL 32433-5606

TELEPHONE NUMBER: 850-419-5684

FAX NUMBER:

EMAIL ADDRESS (if available): wcurtis@waltonso.org

For Program Office Use Only:

- ☒ **Registration Issued** BK
- ☒ **Instrument Added to Evidentiary Instrument Database** SP
- ☒ **Instrument Added to Monthly Statistics Database** SP
- ☒ **Contact Information Added to Instrument Database** SP

2/14/19

FDLE
Florida Department of Law Enforcement

INSTRUMENT PROCESSING SHEET

Agency Walton County SO S/N 80-007173

Date In 01/28/2019 DI Completion Date 2/12/19 ☒Ship ☐P/U ☐H/D ☐CMI ☐EE

Intake Performed By JB

- ☑ Annual
- ☑ Registration
- ☐ Return from CMI / EE

Visual Inspection:

- ☑ Case ☑ Handle
- ☑ Keyboard ☑ Dry Gas Shelf
- ☑ Feet ☑ Breath Tube
- ☑ Ports ☑ Screws Tight

Other Equipment/ Accessories:

- ☑ Power cord ☐ Printer Cable
- ☑ Static Bag ☑ 12V DC Cable

Notes: _____

Final Release Date

FDLE

FEB 14 2019

Alcohol Testing Program

Quality Checks Performed By PJM

- ☑ Breath Tube Screen
- ☒ Replace External O-Rings
- ☒ Instrument Set Up Verified
- ☒ R-Value 178
- ☒ Flow Verification (L/s)

Flow Column # ATP102

32 mm	.195	(.139 - .169)
36 mm	.228	(.156 - .190)
53 mm	.292	(.228 - .278)
103 mm	.632	(.447 - .547)

- ☒ Barometric Pressure Check

Gauge ID # 30793

- ☒ Stability Checks

Simulator	Serial #	Lot #/Exp
0.050	SD1021	201707D
		07/25/2019
0.080	DR1275	201707E
		07/25/2019
0.200	SD1013	201707C
		07/24/2019
0.080 DGS	N/A	RG831804
		11/14/20

Flow Calibration Performed By PJM

Flow Column # ATP103

- ☒ 5L/min – 17mm
- ☒ 15L/min – 53mm
- ☒ 30L/min – 103mm
- ☒ R-Value 178
- ☒ Post Calibration Verification (L/s)

Flow Column # ATP102

32 mm	.148	(.139 - .169)
36 mm	.167	(.156 - .190)
53 mm	.234	(.228 - .278)
103 mm	.482	(.447 - .547)

Maintenance Performed By _____

- ☐ Battery Replacement
- ☐ Dry Gas Regulator Replacement
- ☐ Breath Tube Replacement
- ☐ Other _____

Temperature Checks Performed By PJM

- ☒ Lab Temp °C 21.8

External Digital Therm. ID#: 300505

- ☒ 34°C +-.2 Serial #: SD1021
- ☒ 34°C +-.2 Serial #: DR1275
- ☒ 34°C +-.2 Serial #: SD1013

Calibration Adjustment Performed By _____

Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

☐ Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

Department Inspection Performed By PJM

Barometric Pressure ID# 30793

Gauge 1019 Instrument 1018

Mouth Alcohol Solution Lot # 2018-B

Acetone Stock Solution Lot # 2018-A

Simulator	Serial Number
0.000	G11621
Interferent	DR3855
0.050	SD1021
0.080	DR1275
0.200	SD1013

Attachments

☒ Form 41	☐ Post-Stability Checks
☒ Stability Checks	☒ Flow Calibration
☒ Calibration Certificate	☐ Form 40
☐ Calibration Adjustment	☐ Other _____

Notes/Suggested Service: _____

- ☒ **Instrument Complies with Chapter 11D-8, FAC**
- ☐ **Instrument Does Not Comply with Chapter 11D-8, FAC**
- ☒ **Return to/Place into Evidentiary Use**
- ☐ **Remain Out of Evidentiary Use**
- ☒ **Conduct an Agency Inspection Before Evidentiary Use**

SP 2/14/19 — Tech Review / Date

[signature] 2/14/19 — Admin Review / Date

Florida Department of Law Enforcement
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: WALTON COUNTY SO
Time of Inspection: 11:52
Date of Inspection: 02/12/2019
Serial Number: 80-007173
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG831804 Exp: 11/14/2020
0.000	0.048	0.079	0.198	0.078
0.000	0.048	0.079	0.199	0.078
0.000	0.049	0.080	0.199	0.078
0.000	0.049	0.080	0.199	0.077
0.000	0.048	0.080	0.199	0.077
0.000	0.049	0.081	0.199	0.078
0.000	0.049	0.080	0.199	0.078
0.000	0.049	0.080	0.199	0.077
0.000	0.049	0.080	0.199	0.078
0.000	0.049	0.080	0.199	0.078

Standard Deviations	0.0004	0.0005	0.0003	0.0004

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

SP

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Patrick J Murphy
PATRICK J MURPHY
Signature and Printed Name

02/12/2019
Date

2/14/19

FDLE/ATP Form 41 –Revised August 2005

WALTON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007173
02/12/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:02
Control Test	0.048	09:03
Air Blank	0.000	09:03
Control Test	0.048	09:04
Air Blank	0.000	09:05
Control Test	0.048	09:05
Air Blank	0.000	09:06

Control Test Stats

Average	0.0480
Std Dev	0.0000
Rel Std Dev(%)	0.0000

P Murphy
Operator's Signature

WALTON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007173
02/12/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:07
Control Test	0.079	09:08
Air Blank	0.000	09:08
Control Test	0.079	09:09
Air Blank	0.000	09:10
Control Test	0.080	09:10
Air Blank	0.000	09:11

Control Test Stats

Average	0.0793
Std Dev	0.0006
Rel Std Dev(%)	0.7277

P Murphy
Operator's Signature

WALTON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007173
02/12/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:12
Control Test	0.198	09:13
Air Blank	0.000	09:13
Control Test	0.197	09:14
Air Blank	0.000	09:15
Control Test	0.197	09:15
Air Blank	0.000	09:16

Control Test Stats

Average	0.1973
Std Dev	0.0006
Rel Std Dev(%)	0.2926

P Murphy
Operator's Signature

WALTON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007173
02/12/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:18
Control Test	0.077	09:18
Air Blank	0.000	09:19
Control Test	0.077	09:19
Air Blank	0.000	09:20
Control Test	0.077	09:20
Air Blank	0.000	09:20

Control Test Stats

Average	0.0770
Std Dev	0.0000
Rel Std Dev(%)	0.0000

DGS

P Murphy
Operator's Signature

CMI INC.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007173
02/11/2019
Software: 8100.27

Flow Rate Calibration********
1: Rate (Liters/min) = 5
SQRT(Diff)) = 6.926
2: Rate (Liters/min) = 15
SQRT(Diff)) = 11.000
3: Rate (Liters/min) = 30
SQRT(Diff)) = 21.281
Dependent Data Scale Factor = 100000 L/min
Independent Data Scale Factor = 256
Rounded Slope = 659
Rounded Intercept = -537540
Correlation = 0.99160

SP

2/4/19

Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007173 , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	80-007173
Owning Agency:	WALTON COUNTY SO
Calibration Date:	02/12/2019
Calibration Time:	11:52

	UNCERTAINTY* ±
0.050 g/ 210 L	0.004
0.080 g/ 210 L	0.004
0.200 g/ 210 L	0.007
0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

TRACEABILITY INFORMATION
This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

02/12/2019
Date

Patrick J Murphy
PATRICK J MURPHY,
Department Inspector

2/4/19

Graham, Thomas

From: Hagan, Pam <pjhagan@alcoholtest.com>
Sent: Tuesday, January 29, 2019 9:32 AM
To: Graham, Thomas
Subject: Re: 80-007173

Good Morning Sir...

You never bug me!! The instrument belongs to Walton County SO. If you need more info., just let me know.

Pam

Sent from my iPhone

On Jan 29, 2019, at 8:20 AM, Graham, Thomas <ThomasGraham@fdle.state.fl.us<mailto:ThomasGraham@fdle.state.fl.us>> wrote:

Hi Pam,

Sorry to bug you again. Have you been able to determine which agency purchased Intox 8000 80-007173?

TJ Graham, D-ABFT-FA
Quality Assurance Manager
Alcohol Testing Program
Florida Dept. of Law Enforcement
Office: 850-617-1285
Cell: 850-445-7938

PLEASE NOTE : Florida has a very broad public records law. Most written communications to or from state officials regarding state business are considered to be public records and will be made available to the public and the media upon request. Your e-mail messages may, therefore, be subject to public disclosure.

Sent from my Verizon, Samsung Galaxy smartphone

This email has been scanned by the Symantec Email Security.cloud service.
For more information please visit http://www.symanteccloud.com

2/14/19 SP