



Florida Department of Law Enforcement

DEPARTMENT INSPECTOR FIELD NOTES

Agency: FHP Instrument Serial Number: 80-000785

Department Inspection, Agency Inspection/Agency Contact, Records Audit. Includes checkboxes for Registration, Agency Inspection Notification, and Records Audit items.

Comments: A.I. reported a failed mouth alcohol test, an RFI caused by his cell phone, and a control outside of tolerance, all operator errors. He judged the inspection as non-compliant though it met the rule requirements, and completed a successful second inspection.

Instrument/Area, Equipment, Supplies. Includes checkboxes for Clean/Dry, Proper Number of Simulators, and various supply items.

Comments: (Empty space for additional notes)

Action: Instrument Complies with Chapter 11D-8, FAC; Instrument Does Not Comply with Chapter 11D-8, FAC; Remain in Evidentiary Use; Return to/Place into Evidentiary Use; Remove from Evidentiary Use; Remain Out of Evidentiary Use; Conduct an Agency Inspection Before Evidentiary Use. Includes a Comments section.

Patrick Murphy (Handwritten signature)

Signature of Alcohol Testing Program Staff Member

5/14/19 (Handwritten date)

Date

Original - FDLE

Copy - Agency