



Florida Department of Law Enforcement

DEPARTMENT INSPECTOR FIELD NOTES

Agency: Chattahoochee P.D. Instrument Serial Number: 80-000887

Department Inspection: Agency Inspection/Agency Contact: Records Audit:
Registration, Agency Inspection Notification, Barometric Pressure, Date of Notification/Contact, Agency Inspector, Instrument Set Up Verified

Comments: A.I. reports a failed inspection due to a faulty simulator. Inspection was repeated once the problem was fixed and all measurements were within the acceptable range.

Instrument/Area: Equipment: Supplies:
Clean/Dry, Proper Number of Simulators, Distilled/Deionized Water, Secure, Checked Simulators for Air Leak, Mouth Alcohol Solution, Limited Access, Resistant Seal and Proper Temperature, Acetone Stock Solution, Alcohol Reference Solution, Dry Gas Standard, Mouth Pieces

Comments:

Action: Instrument Complies with Chapter 11D-8, FAC, Instrument Does Not Comply with Chapter 11D-8, FAC, Remain in Evidentiary Use, Return to/Place into Evidentiary Use, Remove from Evidentiary Use, Remain Out of Evidentiary Use, Conduct an Agency Inspection Before Evidentiary Use

Patrick Murphy

Signature of Alcohol Testing Program Staff Member

7/3/19

Date

Original - FDLE

Copy - Agency

CITY of CHATTAHOOCHEE

OFFICE of the CHIEF OF POLICE

32 JEFFERSON STREET • CHATTAHOOCHEE, FL 32324
PHONE: (850) 663-4383 • FAX # (850) 663-8510

June 25th 2019

Inspector Patrick Murphy

Alcohol Testing Program

PO Box 1489

Tallahassee, Florida 32302

Inspector Murphy,

Please find enclosed this months' Form 40 for the monthly Agency Inspection of Intoxilyzer 8000 series instrument #80-000887, which is located and has been returned to service at the Chattahoochee Police Department. All results were within acceptable range.

The first inspection was outside tolerance (.20) due to simulator seal. In running the second test I found the simulator container was broken.



Chief FC O'Brian, Agency Inspector

Chattahoochee Police Department

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CHATTAHOOCHEE PD

Serial Number: 80-000887

Time of Inspection: 15:09

Date of Inspection: 06/25/2019

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#: 201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot #: Exp:
0.000	0.050	0.080	0.174 / 0.170	
0.000	0.050	0.080	0.174 / 0.175	
0.000	0.050	0.080	0.175 / 0.178	

Number of Simulators Used: 5

Remarks:

20: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

*.20 ↓ poor simulator seal
.20 ↓ broken g less thread on simulator container*

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



FREDERICK C OBRIAN

Signature and Printed Name

06/25/2019

Date