



Florida Department of Law Enforcement

DEPARTMENT INSPECTOR FIELD NOTES

Agency: Pensacola P.D. Instrument Serial Number: 80-007159

Department Inspection, Agency Inspection/Agency Contact, Records Audit. Includes checkboxes for Registration, Agency Inspection Notification, and Records Audit items.

Comments: A.I. reported failed inspection due to unusually cold ambient conditions. He raised the room temperature and repeated the inspection with success.

Instrument/Area, Equipment, Supplies. Includes checkboxes for Clean/Dry, Proper Number of Simulators, Distilled/Deionized Water, etc.

Comments: (Empty space for additional notes)

Action: Instrument Complies with Chapter 11D-8, FAC. Includes checkboxes for Evidentiary Use and Agency Inspection Before Evidentiary Use.

Signature of Alcohol Testing Program Staff Member (Handwritten: Patrick Murphy)

Date (Handwritten: 8/15/2019)

Original - FDLE Copy - Agency



Florida Department of Law Enforcement

DEPARTMENT INSPECTOR FIELD NOTES

Agency: Pensacola P.D.

Instrument Serial Number: 80-007159

Department Inspection: <input type="checkbox"/> Registration <input type="checkbox"/> Request for Registration Attached <input type="checkbox"/> Annual <input type="checkbox"/> Inspection After Repair <input type="checkbox"/> Other: _____ <input type="checkbox"/> Barometric Pressure: _____ Gauge _____ Instrument <input type="checkbox"/> Instrument Set Up Verified	Agency Inspection/Agency Contact: <input checked="" type="checkbox"/> Agency Inspection Notification <input type="checkbox"/> Agency Contact <input type="checkbox"/> Other _____ Date of Notification/Contact: May 3, 2019 Agency Inspector: James Daniels	Records Audit: The following records were audited: <input type="checkbox"/> Agency Inspection Reports <input type="checkbox"/> Breath Test Instrument Repair/Maintenance Records <input type="checkbox"/> Instrument Registration <input type="checkbox"/> Dry Gas Standard Certificate(s) of Analysis <input type="checkbox"/> Other: _____ <input type="checkbox"/> Other: _____
---	---	--

Comments:
A.I. reported failed inspection due to low 0.05 solution readings. Solution was replaced and the subsequent inspection was successful.

Instrument/Area: <input type="checkbox"/> Clean/Dry <input type="checkbox"/> Secure <input type="checkbox"/> Limited Access <input type="checkbox"/> Other: _____	Equipment: <input type="checkbox"/> Proper Number of Simulators <input type="checkbox"/> Checked Simulators for Air Leak Resistant Seal and Proper Temperature <input type="checkbox"/> Class A Glassware	Supplies: <input type="checkbox"/> Distilled/Deionized Water <input type="checkbox"/> Mouth Alcohol Solution <input type="checkbox"/> Acetone Stock Solution <input type="checkbox"/> Alcohol Reference Solution <input type="checkbox"/> Dry Gas Standard <input type="checkbox"/> Mouth Pieces
--	---	---

Comments:

Action: <input type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input type="checkbox"/> Remain in Evidentiary Use <input type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remove from Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use	Comments:
--	------------------

Patrick Murphy
Signature of Alcohol Testing Program Staff Member

5/13/19
Date

Original - FDLE Copy - Agency