Kirkland, Brett

From:

Collito, Brad <bcollito@hernandosheriff.org>

Sent:

Monday, October 28, 2019 11:05 AM

To:

Ben Larison; Conrad Juergensmeyer; David Vitola; Alcohol Testing Program; Ivette

Miranda; Klucznik, Shaun; Lisa Yeager (Iyeager@sao5.org); Paula Tyrell; Quinones,

Daniel; Robinson, Cyrus; Sonya Wolthuis; Stevens, Teresa; Wegner, Steven

Cc:

Brown, Kelly W.; Origon, Lance

Subject:

Hernando County Sheriff's Office October 2019 Agency

Attachments:

October 2019 Agency Inspections.pdf

Good Afternoon,

Attached are the October 2019 Agency Inspections for the Hernando County Sheriff's Office.

Instrument 5249 was inspected with no issues to report. Instrument 5249 is currently being utilized as our primary instrument.

Instrument 814 was not inspected. Instrument 814 is currently awaiting service at CMI, Inc as of 10/16/2019. Instrument 814 lost power and was not able to be powered back on during September 2019.

Instrument 812 was inspected with no issues to report. Instrument 812 is currently being utilized as our back up instrument.

All of our instruments have had their annual inspections completed by FDLE.

All of our simulators have been calibrated by TechSource and we are now on a one year calibration cycle for our simulators.

If I can be of additional assistance, please do not hesitate to contact me.

Respectfully, Sgt. B. Collito

Sergeant Brad Collito, CCM Classification Unit Hernando County Detention Center 16425 Spring Hill Dr Brooksville, FL 34604 T(352)797-3426 F(352)544-2352 bcollito@hernandosheriff.org

Pinsky, David

From: Collito, Brad <bcollito@hernandosheriff.org>
Sent: Thursday, September 26, 2019 9:11 AM

To: Origon, Lance

Cc: Ben Larison; Conrad Juergensmeyer; Alcohol Testing Program; Ivette Miranda; Klucznik,

Shaun; Paula Tyrell; Quinones, Daniel; Robinson, Cyrus; Sonya Wolthuis; Stevens,

Teresa; Wegner, Steven; Brown, Kelly W.

Subject: Re: Intozxilyzer 80-000814 Update

Good morning,

We have two operational instruments at this time.

Sgt. B. Collito

Hernando County Detention Center

On Sep 26, 2019, at 9:08 AM, Origon, Lance < LOrigon@hernandosheriff.org < mailto:LOrigon@hernandosheriff.org >> wrote:

Sgt Collito,

Thank you for the update.

Just to confirm we still have one operational instrument, correct?

Corporal Lance Origon

S.T.E.P. / S.W.A.T. Assistant Team Leader / F.T.O.

Hernando County Sheriff's Office www.hernandosheriff.org/">www.hernandosheriff.org/ 18900 Cortez Blvd, Brooksville, Florida 34601 Office phone 352-754-6830 Extension 53959 Cell phone 352-442-8197

From: Collito, Brad

Sent: Wednesday, September 25, 2019 12:42 PM

To: Ben Larison; Conrad Juergensmeyer; FDLE; Ivette Miranda; Klucznik, Shaun; Paula Tyrell; Quinones, Daniel; Robinson,

Cyrus; Sonya Wolthuis; Stevens, Teresa; Wegner, Steven

Cc: Brown, Kelly W.; Origon, Lance Subject: Intozxilyzer 80-000814 Update

Good afternoon,

I received a telephone call today from FDLE in reference to our inoperable instrument. FDLE has advised me to complete a handwritten Form 40 Agency Inspection Report due to the inability to complete an inspection on 80-000814 this month. FDLE was unable to power up the instrument. Per our request, the instrument will be sent to CMI, Inc for diagnosis and repair.

Thank you

Sergeant Brad Collito, CCM Classification Unit

Hernando County Detention Center 16425 Spring Hill Dr Brooksville, FL 34604 T(352)797-3426 F(352)544-2352

Bcollito@hernandosheriff.org<mailto:Bcollito@hernandosheriff.org>

Kirkland, Brett

From:

Collito, Brad <bcollito@hernandosheriff.org>

Sent:

Sunday, September 15, 2019 11:06 PM

To:

Ben Larison; Conrad Juergensmeyer; Alcohol Testing Program; Ivette Miranda; Klucznik,

Shaun; Paula Tyrell; Quinones, Daniel; Robinson, Cyrus; Sonya Wolthuis; Stevens, Teresa;

Wegner, Steven

Cc:

Brown, Kelly W.; Origon, Lance

Subject:

Hernando County Sheriff's Office September 2019 Agency Inspections

Attachments:

September 2019 Agency Inspections.pdf

Good Afternoon,

Attached are the September 2019 Agency Inspections for the Hernando County Sheriff's Office.

Instrument 5249 was inspected with no issues to report. Instrument 5249 is currently being utilized as our primary instrument.

Instrument 814 was not inspected. Instrument 814 lost power and was not able to be powered back on. No agency inspection was able to be completed on this instrument due to the inability to power it on. Instrument 814 will be shipped back to FDLE on 9/16/2019 for repair at CMI.

Instrument 812 was returned by Mary Bowermaster of the Citrus County Detention Center on 9/15/2019. Instrument 812 was immediately inspected with no issues to report. Instrument 812 was placed back into service and is currently being utilized as our back up instrument.

All of our instruments have had their annual inspections completed by FDLE.

All of our simulators have been calibrated by TechSource and we are now on a one year calibration cycle for our simulators.

If I can be of additional assistance, please do not hesitate to contact me.

Respectfully, Sgt. B. Collito

Sergeant Brad Collito, CCM Classification Unit Hernando County Detention Center 16425 Spring Hill Dr Brooksville, FL 34604 T(352)797-3426 F(352)544-2352 bcollito@hernandosheriff.org

Deputy Brad Collito Intake/ Release Unit Hernando County Detention Center 16425 Spring Hill Dr Brooksville, FL 34604 T(352)544-2334 F(352)544-2349

Respectfully,

Pinsky, David	
From: Sent: To: Cc: Subject: Attachments:	Collito, Brad <bcollito@hernandosheriff.org> Sunday, August 25, 2019 1:08 PM Ben Larison; Conrad Juergensmeyer; Alcohol Testing Program; Ivette Miranda; Klucznik, Shaun; Paula Tyrell; Quinones, Daniel; Robinson, Cyrus; Sonya Wolthuis; Stevens, Teresa; Wegner, Steven Hillman, Bill; Origon, Lance Hernando County Sheriff's Office August 2019 Agency Inspections August 2019 Agency Inspections.pdf; Citrus Custody Receipt.pdf</bcollito@hernandosheriff.org>
Good Afternoon,	
Attached are the August 2019 Ag	ency Inspections for the Hernando County Sheriff's Office.
Instrument 5249 was inspected winstrument.	vith no issues to report. Instrument 5249 is currently being utilized as our back-up
Instrumen <mark>t 814 was i</mark> nspected. Ir	nstrument 814 is currently being utilized as our primary instrument.
Instrument 812 was transferred to Copies of the most recent Agency	th no issues to report. Upon receiving telephonic direction from Dr. Kirkland, o Mary Bowermaster from the Citrus County Detention Center after being inspected. Inspection and Annual Inspection were transferred with the instrument. Ms. ceipt upon accepting the instrument (attached). Upon return, I will be completing an ing it back into service.
All of our instruments have had th	eir annual inspections completed by FDLE.
All of our simulators have been ca simulators.	librated by TechSource and we are now on a one year calibration cycle for our
f I can be of additional assistance,	please do not hesitate to contact me.

Sgt. B. Collito

Sergeant Brad Collito, CCM

Classification Unit

Hernando County Detention Center

16425 Spring Hill Dr

Brooksville, FL 34604

T(352)797-3426 F(352)544-2352

bcollito@hernandosheriff.org<mailto:bcollito@hernandosheriff.org>

<mailto:bcollito@hernandosheriff.org>