

80-000827
SLL 1/18/19

827

- PASSES MONTHLY INSPECTION
- INSTRUMENT SHOWS O PSI ON SCREEN FOR DRY GAS STANDARD, THE ~~GAUGE~~ GAUGE SHOWS THE PROPER PSI.

LT. JEFFERY RAKER
Jeffery Raker

(352) 843-4215

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MARION COUNTY SO
Time of Inspection: 09:50

Date of Inspection: 01/14/2019

Serial Number: 80-000827
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:875088 Exp: 12/12/2020
/ 0.000	0.050	0.080	0.199	0.079
/ 0.000	0.050	0.081	0.201	0.079
/ 0.000	0.050	0.080	0.202	0.080


Number of Simulators Used: 5

Remarks:

00: Ambient Fail.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



JEFFERY A RAKER

Signature and Printed Name

01/14/2019
Date