

Pinsky, David

From: Dan Lyons <dlyo@OSCEOLA.ORG>
Sent: Saturday, August 3, 2019 12:52 AM
To: Graham, Thomas; 'Atkinson, Karsona'; 'McKnight, Jamie (JMcknight@sao9.org)'; 'SherryForcier@flhsmv.gov'; Alcohol Testing Program
Cc: Michael Hernandez; Brian Adams
Subject: monthly intox inspection report OSCO
Attachments: 02 August 2019 Monthly Intox Inspec.pdf

ALCON,

The monthly Agency Inspection Report conducted 08/02/2019 on the following instruments;
80-000966, 80-000968, 80-001715, 80-003935, 80-003937, 80-004587 and 80-005935.

Ref to instrument 80-005937, Interferent Detect Test; I did not connect the simulator. Simulator was connected and the instrument was re-tested, not re-started. The re-test results are within standards.

Ref to instrument 80-004587, 0.05g/210L Test; I did not connect the simulator. Simulator was connected and the instrument was re-tested, not re-started. The re-test results are within standards.

All intox inspections are complete and in compliance with 11D-8.

V/R

Dan Lyons
Osceola County Sheriff's Office
Spec Ops/Traffic/DUI Enforcement Unit
Agency 407-348-1100
Agency Cell 407-269-1980
Cell 321-960-2066

Attention: The information contained in this E-mail message is privileged and confidential information intended only for the use of the individual(s) named above. If the reader of this message is not the intended recipient, you are hereby notified that any dissemination, distribution or copy of this communication is strictly prohibited. If you have received this communication in error, please contact the sender by reply E-mail and destroy all copies of the original message.

From: Dan Lyons
To: [Graham, Thomas](#); ["McKnight, Jamie \(JMcknight@sao9.org\)"](#); ["Atkinson, Karsona"](#); ["SherryForcier@flhsmv.gov"](#); [George Maroney](#)
Subject: Monthly intox inspections/ 24 Jan 2019
Date: Friday, January 25, 2019 2:09:46 AM
Attachments: [24 Jan 2019 Monthly Intox Inspec.pdf](#)

ALCON,

Instruments 80-000968 , 80-003937 and 80-004587 have returned from Annual inspections with FDLE.

Agency Inspection Report conducted 01/24/2019 on the following instruments;
80-000966, 80-000968, 80-001715, 80-003935, 80-003937 , 80-004587 and 80-005935.

Ref to instrument 80-003935, 0.05g/210L Test, Received an outside tolerance. Re-set lid to simulator, the instrument was re-tested. The re-test results are with in standards. The inspection was not Re-Started.

Ref to instrument 80-000968, AF/MA. MA did not indicate SNM. Rinsed with additional mouth alcohol solution and re-tested. The re-test results are with in standards. The inspection was not Re-Started.

Instruments 80-001715 and 80-003935 have been sent off to FDLE for Annual inspections.

All intox inspections are complete and in compliance with 11D-8.

V/R

Dan Lyons
Osceola County Sheriff's Office
Spec Ops/Traffic/DUI Enforcement Unit
Agency 407-348-1100
Agency Cell 407-269-1980
Cell 321-960-2066

Attention: The information contained in this E-mail message is privileged and confidential information intended only for the use of the individual(s) named above. If the reader of this message is not the intended recipient, you are hereby notified that any dissemination, distribution or copy of this communication is strictly prohibited. If you have received this communication in error, please contact the sender by reply E-mail and destroy all copies of the original message.

From: Dan Lyons
To: "McKnight, Jamie (JMcknight@sao9.org)"; "SherryForcier@flhsmv.gov"; "Atkinson, Karsona"; Graham, Thomas
Cc: George Maroney; Jacob Beekman
Subject: monthly intox inspections Jan 2019
Date: Monday, January 07, 2019 2:15:46 AM
Attachments: 06 Jan 2019 Monthly intox inspec.pdf

ALCON,

Agency Inspection Report conducted 01/06/2019 on the following instruments;
80-000966, 80-000968, 80-001715, 80-003935, 80-003937, 80-004587 and 80-005935.

Ref to instrument 80-000966, The Alcohol Free Subject Test: showed a 0.009, due to residual mouth alcohol from the previous Mouth Alcohol test. The test was completed and noted as not within standards. The instrument was re-tested and the results are with in standards.

Ref to instrument 80-000968, AF/MA initial reading of SNM for AF. Re-tested, Had another Deputy provide AF. The re-test results are with in standards

Ref to instrument 80-003937, 0.05g/210L Test, re-set lid to simulator, re-tested. 0.20g/210L Test, re-set lid to simulator. The re-test results are with in standards. The inspection was not Re-Started.

Instruments 80-000968 and 80-003937 have been sent off to FDLE for Annual inspections.

All intox inspections are complete and in compliance with 11D-8.

V/R

Dan Lyons
Osceola County Sheriff's Office
Spec Ops/Traffic/DUI Enforcement Unit
Agency 407-348-1100
Agency Cell 407-269-1980
Cell 321-960-2066

Attention: The information contained in this E-mail message is privileged and confidential information intended only for the use of the individual(s) named above. If the reader of this message is not the intended recipient, you are hereby notified that any dissemination, distribution or copy of this communication is strictly prohibited. If you have received this communication in error, please contact the sender by reply E-mail and destroy all copies of the original message.