



Highlands County Sheriff's Office

Sheriff Paul Blackman

400 S. Eucalyptus Street
Sebring, Florida 33870
863-402-7200
www.highlandssheriff.org



Dually Accredited

July 04, 2019

Mrs. Shayla Platt
Florida Department of Law Enforcement
Alcohol Testing Program
PO Box 1489
Tallahassee, FL 32302

REF: July 04, 2019 Inspection

Dear Mrs. Platt:

I am mailing a copy of the Agency Inspection Reports for the month of July 04, 2019. Instrument 80-001054 and 80-001043 have passed inspection.

However for 1054 during the .020 test at the end I accidentally hit the enter instead of the escape to indicate I had an error which of course I did not. So I entered 00 as the reason for the error and aborted the entire test. I performed a second inspection with only 1 minor problem. During the mouth alcohol part I had too much solution in my mouth and gave an improper sample message. Cleared/rinsed my mouth and repeated that part with no issue.

Sgt. Glasco
Agency Inspector

Cc: File

1054 SP

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO
Time of Inspection: 22:41

Date of Inspection: 07/04/2019

Serial Number: 80-001054
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#: 201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 13518080A6 Exp: 08/05/2020
0.000	0.047	0.077	0.194	0.079
0.000	0.047	0.077	0.196	0.079
0.000	0.047	0.077	0.195	0.079

Number of Simulators Used: 5

Remarks:

A F / M A: Improper Sample.

1st attempt I had too much alcohol solution in my mouth.
2nd attempt no issues, passed

A TRUE COPY
PAUL BLACKMAN, SHERIFF
HIGHLANDS COUNTY, FLORIDA
BY _____ D.S.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

GREGORY S GLASCO

Signature and Printed Name

07/04/2019
Date

1054 SP

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO
Time of Inspection: 22:02

Date of Inspection: 07/04/2019

Serial Number: 80-001054
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000	0.047	0.193		
0.000	0.047	0.194		
0.000	0.047	0.194		


Number of Simulators Used: 5

Remarks:
OOHIT WRONG KEY.

During .20 test I hit the exception key by mistake and aborted the test/inspection and started over. My fault not Instruments I believe I hooked up the wrong solution simulator.

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



 Signature and Printed Name GREGORY S GLASCO

07/04/2019
Date