

DEPARTMENT INSPECTOR FIELD NOTES

Agency: Eglin Air Force Base

Instrument Serial Number: 80-001155

Department Inspection:	Agency Inspection/Agency Contact:	Records Audit:
<input type="checkbox"/> Registration <input type="checkbox"/> Request for Registration Attached <input type="checkbox"/> Annual <input type="checkbox"/> Inspection After Repair <input type="checkbox"/> Other: _____ <input type="checkbox"/> Barometric Pressure: Gauge Instrument <input type="checkbox"/> Instrument Set Up Verified	<input checked="" type="checkbox"/> Agency Inspection Notification <input type="checkbox"/> Agency Contact <input type="checkbox"/> Other _____ Date of Notification/Contact: <u>3/4/2019</u> Agency Inspector: <u>James Blackwell</u>	The following records were audited: <input type="checkbox"/> Agency Inspection Reports <input type="checkbox"/> Breath Test Instrument Repair/Maintenance Records <input type="checkbox"/> Instrument Registration <input type="checkbox"/> Dry Gas Standard Certificate(s) of Analysis <input type="checkbox"/> Other: _____ <input type="checkbox"/> Other: _____
Comments: A.I. reported two failed inspections when the instrument failed to detect interferent properly. Instrument had just been seen by us and agency acetone stock solution was old. We sent them new bottles of acetone stock solution.		

Instrument/Area:	Equipment:	Supplies:
<input type="checkbox"/> Clean/Dry <input type="checkbox"/> Secure <input type="checkbox"/> Limited Access <input type="checkbox"/> Other: _____	<input type="checkbox"/> Proper Number of Simulators <input type="checkbox"/> Checked Simulators for Air Leak Resistant Seal and Proper Temperature <input type="checkbox"/> Class A Glassware	<input type="checkbox"/> Distilled/Deionized Water <input type="checkbox"/> Mouth Alcohol Solution <input type="checkbox"/> Acetone Stock Solution <input type="checkbox"/> Alcohol Reference Solution <input type="checkbox"/> Dry Gas Standard <input type="checkbox"/> Mouth Pieces
Comments: <hr/> <hr/> <hr/>		

Action:	
<input type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC	Comments: <hr/> <hr/> <hr/> <hr/>
<input type="checkbox"/> Remain in Evidentiary Use <input type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remove from Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use	
<input type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use	

Patrick Murphy
Signature of Alcohol Testing Program Staff Member

3/5/19
Date

Original – FDLE

Copy – Agency