

Murphy, Patrick

From: Enterkin, Wayne <wenterkin@sroso.net>
Sent: Wednesday, January 02, 2019 10:05 AM
To: Murphy, Patrick
Cc: Nelson, Duron
Subject: Non-Compliant Agency Inspections - 80-000772, 80-001302, 80-001303
Attachments: SHR_C25819010208560.pdf

Dr. Murphy,

In accordance with FAC 11D-8, please find agency inspection reports for instruments; 80-000772, 80-001302, 80-001303.

Agency Inspector Sgt. Nelson performed said inspections on 12/31/2018 and received non-compliance results from all three instruments.

After the non-compliance results, the area housing the instruments was ventilated to reduce or eliminate any alcohol contamination of the ambient air, and the tests were repeated with all instruments passing.

Please contact me if you have any questions.

V/r,

Lieutenant Wayne Enterkin
Department of Administration
Professional Standards Unit
Santa Rosa Sheriff's Office
(850) 981-2242



CONFIDENTIALITY NOTICE: This message and any attachments are for the sole use of the intended recipient (s) and may contain confidential and privileged information that is exempt from public disclosure. Any unauthorized review, use, disclosure, or distribution is prohibited. If you have received this message in error please contact the sender (by phone or reply electronic mail) and then destroy all copies of the original message.