

Pinsky, David

From: Dan Lyons <dlyo@OSCEOLA.ORG>
Sent: Saturday, August 3, 2019 12:52 AM
To: Graham, Thomas; 'Atkinson, Karsona'; 'McKnight, Jamie (JMcknight@sao9.org)'; 'SherryForcier@flhsmv.gov'; Alcohol Testing Program
Cc: Michael Hernandez; Brian Adams
Subject: monthly intox inspection report OSCO
Attachments: 02 August 2019 Monthly Intox Inspec.pdf

ALCON,

The monthly Agency Inspection Report conducted 08/02/2019 on the following instruments;
80-000966, 80-000968, 80-001715, 80-003935, 80-003937, 80-004587 and 80-005935.

Ref to instrument 80-005937, Interferent Detect Test; I did not connect the simulator. Simulator was connected and the instrument was re-tested, not re-started. The re-test results are with in standards.

Ref to instrument 80-004587, 0.05g/210L Test; I did not connect the simulator. Simulator was connected and the instrument was re-tested, not re-started. The re-test results are with in standards.

All intox inspections are complete and in compliance with 11D-8.

V/R

Dan Lyons
Osceola County Sheriff's Office
Spec Ops/Traffic/DUI Enforcement Unit
Agency 407-348-1100
Agency Cell 407-269-1980
Cell 321-960-2066

Attention: The information contained in this E-mail message is privileged and confidential information intended only for the use of the individual(s) named above. If the reader of this message is not the intended recipient, you are hereby notified that any dissemination, distribution or copy of this communication is strictly prohibited. If you have received this communication in error, please contact the sender by reply E-mail and destroy all copies of the original message.

Graham, Thomas

From: Dan Lyons <dlyo@OSCEOLA.ORG>
Sent: Tuesday, March 19, 2019 3:02 AM
To: Graham, Thomas; 'McKnight, Jamie (JMcknight@sao9.org)'; 'Atkinson, Karsona'; 'SherryForcier@flhsmv.gov'; George Maroney
Subject: Monthly Intox Inspections March 2019
Attachments: 18 Mar 2019 Monthly Intox Inspec.pdf

Follow Up Flag: FollowUp
Flag Status: Flagged

3/19/19
JA

ALCON,

Agency Inspection Report conducted 03/18/2019 on the following instruments;
80-000968, 80-001715, 80-003935, 80-003937, 80-004587 and 80-005935.

Ref to instrument 80-001715, 0.05g/210L Test, Received an outside tolerance. Re-set lid to simulator, the instrument was re-tested. The re-test results are with in standards. The inspection was not Re-Started.

Instruments 80-000966 is still out at CMI for Annual Inspection and repairs.

All intox inspections are complete and in compliance with 11D-8.

V/R

Dan Lyons
Osceola County Sheriff's Office
Spec Ops/Traffic/DUI Enforcement Unit
Agency 407-348-1100
Agency Cell 407-269-1980
Cell 321-960-2066

Attention: The information contained in this E-mail message is privileged and confidential information intended only for the use of the individual(s) named above. If the reader of this message is not the intended recipient, you are hereby notified that any dissemination, distribution or copy of this communication is strictly prohibited. If you have received this communication in error, please contact the sender by reply E-mail and destroy all copies of the original message.