

## Graham, Thomas

---

**From:** Lawrence Kolbicka <LKolbicka@pinellas-park.com>  
**Sent:** Thursday, April 04, 2019 3:36 PM  
**To:** Graham, Thomas  
**Subject:** PPPD Intox Issues

**Follow Up Flag:** FollowUp  
**Flag Status:** Flagged

Mr Graham,

I completed both Pinellas Park Police Department agency inspections today and had two issues.

1 - instrument 1042. I did not get a interferent detect on my interferent sim. I checked and the top was on loosely. I tightened it and had no further insuses.

2 - Instrument 1732. I got an ambient fail on the first dry gas standard. I found that I did not connect the hose to the instrument from the dry gas. I connected the hose and had no further issues.

Thank You,

Larry

--

OFC Lawrence R Kolbicka II  
Special Operations Unit  
Traffic Homicide Investigator  
Traffic Enforcement Officer  
Agency Inspector

Pinellas Park Police Department