

Pinsky, David

From: Dan Lyons <dlyo@OSCEOLA.ORG>
Sent: Thursday, September 12, 2019 11:01 AM
To: 'McKnight, Jamie (JMcknight@sao9.org)'; 'SherryForcier@flhsmv.gov'; Alcohol Testing Program; 'Atkinson, Karsona'
Cc: Michael Hernandez; Brian Adams
Subject: Monthly intox inspection report, Sept 2019
Attachments: 10 Sept 2019 Monthly Intox Inspec.pdf

ALCON,

The monthly Agency Inspection Report conducted 09/10/2019 on the following instruments; 80-000966, 80-000968, 80-001715, 80-003935, 80-003937, 80-004587 and 80-005935.

Ref to instrument 80-003937, AF/MA Test; volume not met and the instrument was re-tested, not re-started. The re-test results are with in standards.

Ref to instrument 80-004587, 0.08g/210L Dry Gas Std Test; I did not connect the dry gas tube. Tube was connected and the instrument was re-tested, not re-started. The re-test results are with in standards.

I also conducted an additional Inspection report on 06/06/2019 on instrument 80-005935 as part of a training evolution for the ASA's assigned with our State Attorney's Office.

All intox inspections are complete and in compliance with 11D-8.

V/R

Dan Lyons
Osceola County Sheriff's Office
Spec Ops/Traffic/DUI Enforcement Unit
Agency 407-348-1100
Agency Cell 407-269-1980
Cell 321-960-2066

Attention: The information contained in this E-mail message is privileged and confidential information intended only for the use of the individual(s) named above. If the reader of this message is not the intended recipient, you are hereby notified that any dissemination, distribution or copy of this communication is strictly prohibited. If you have received this communication in error, please contact the sender by reply E-mail and destroy all copies of the original message.

Pinsky, David

From: Dan Lyons <dlyo@OSCEOLA.ORG>
Sent: Saturday, August 3, 2019 12:52 AM
To: Graham, Thomas; 'Atkinson, Karsona'; 'McKnight, Jamie (JMcknight@sao9.org)'; 'SherryForcier@flhsmv.gov'; Alcohol Testing Program
Cc: Michael Hernandez; Brian Adams
Subject: monthly intox inspection report OSCO
Attachments: 02 August 2019 Monthly Intox Inspec.pdf

ALCON,

The monthly Agency Inspection Report conducted 08/02/2019 on the following instruments;
80-000966, 80-000968, 80-001715, 80-003935, 80-003937, 80-004587 and 80-005935.

Ref to instrument 80-005937, Interferent Detect Test; I did not connect the simulator. Simulator was connected and the instrument was re-tested, not re-started. The re-test results are with in standards.

Ref to instrument 80-004587, 0.05g/210L Test; I did not connect the simulator. Simulator was connected and the instrument was re-tested, not re-started. The re-test results are with in standards.

All intox inspections are complete and in compliance with 11D-8.

V/R

Dan Lyons
Osceola County Sheriff's Office
Spec Ops/Traffic/DUI Enforcement Unit
Agency 407-348-1100
Agency Cell 407-269-1980
Cell 321-960-2066

Attention: The information contained in this E-mail message is privileged and confidential information intended only for the use of the individual(s) named above. If the reader of this message is not the intended recipient, you are hereby notified that any dissemination, distribution or copy of this communication is strictly prohibited. If you have received this communication in error, please contact the sender by reply E-mail and destroy all copies of the original message.

Graham, Thomas

From: Dan Lyons <dlyo@OSCEOLA.ORG>
Sent: Wednesday, April 03, 2019 9:54 PM
To: Graham, Thomas; 'Atkinson, Karsona'; 'McKnight, Jamie (JMcknight@sao9.org)'; 'SherryForcier@flhsmv.gov'; George Maroney
Subject: Monthly Intox Inspections, April 2019

Follow Up Flag: FollowUp
Flag Status: Flagged

ALCON,

Agency Inspection was conducted 04/03/2019 on the following instruments;
80-000968, 80-001715, 80-003935, 80-003937, 80-004587 and 80-005935.

Ref to instrument **80-004587**, 0.08g/210L Test, Received an outside tolerance on test 1, I did not connect the simulator. I connected the simulator to finish the test. The instrument was re-tested and within standards. I then double tapped the the enter key, which closed out the test. I then marked it out of compliance and re-started the test. The re-start test results are with in standards.

Instruments 80-000966 is still out at CMI for Annual Inspection and repairs.

All intox inspections are complete and in compliance with 11D-8.

V/R

Dan Lyons
Osceola County Sheriff's Office
Spec Ops/Traffic/DUI Enforcement Unit
Agency 407-348-1100
Agency Cell 407-269-1980
Cell 321-960-2066

Attention: The information contained in this E-mail message is privileged and confidential information intended only for the use of the individual(s) named above. If the reader of this message is not the intended recipient, you are hereby notified that any dissemination, distribution or copy of this communication is strictly prohibited. If you have received this communication in error, please contact the sender by reply E-mail and destroy all copies of the original message.

Graham, Thomas

From: Dan Lyons <dlyo@OSCEOLA.ORG>
Sent: Friday, February 08, 2019 6:15 PM
To: Graham, Thomas; 'McKnight, Jamie (JMcknight@sao9.org)'; 'Atkinson, Karsona'; 'SherryForcier@flhsmv.gov'; George Maroney
Subject: RE: Monthly intox inspections/ 07 Feb 2019

Follow Up Flag: FollowUp
Flag Status: Flagged

ALCON,

Instruments 80-001715 and 80-003935 have returned from Annual inspections with FDLE.

Agency Inspection Report conducted 02/07/2019 on the following instruments;
80-000966, 80-000968, 80-001715, 80-003935, 80-003937, 80-004587 and 80-005935.

Ref to instrument 80-003935, 0.08g/210L Test, Received an outside tolerance. Simulator not connected. Simulator connected, the instrument was re-tested. The re-test results are with in standards. The inspection was not Re-Started.

Ref to instrument 80-004587, prior to completion of the inspection, lost power. After power back on, re-started, then re-tested. Re-test, Received AF/MA; Improper sample provided. AF breath provider blew to soon. Provided instruction on when to provide proper sample. Test was re-started. The re-test results are with in standards. The inspection was not Re-Started.

Ref to instrument 80-005935, 0.08g/210L Test, Received an outside tolerance on first test. Re-set lid to simulator, the instrument was re-tested. The re-test results are with in standards. The inspection was not Re-Started.

Instruments 80-005935 and 80-000966 have been sent off to FDLE for Annual inspections.

All intox inspections are complete and in compliance with 11D-8.

V/R

Dan Lyons
Osceola County Sheriff's Office
Spec Ops/Traffic/DUI Enforcement Unit
Agency 407-348-1100
Agency Cell 407-269-1980
Cell 321-960-2066

Attention: The information contained in this E-mail message is privileged and confidential information intended only for the use of the individual(s) named above. If the reader of this message is not the intended recipient, you are hereby notified that any dissemination, distribution or copy of this communication is strictly prohibited. If you have received this communication in error, please contact the sender by reply E-mail and destroy all copies of the original message.

Graham, Thomas

From: Graham, Thomas
Sent: Friday, February 08, 2019 7:24 AM
To: 'Dan Lyons'
Subject: RE: Monthly intox inspections
Attachments: 80-004587 AI Feb 2019.pdf

2/8/19
JD

Dan,

Thank you for informing me about your power issue. I am glad you repeated the test. The only Agency Inspection that uploaded for 80-004587 is attached. I am guessing this is your repeated inspection. Please document the power incident and send me a copy for our file. If you have any further questions, please feel free to call.

TJ

From: Dan Lyons [mailto:dlyo@OSCEOLA.ORG]
Sent: Thursday, February 07, 2019 11:57 PM
To: Graham, Thomas
Subject: Monthly intox inspections

TJ,

I received the 2 intoxs back and conducted my monthly inspections. The last 2 intoxs will be out in the morning to you.

While doing the intox's I had a power outage. Instrument 80-004587, I know had completed the dry gas test and I believe was finished with the diagnostic check.

Power went out and I restarted the instrument and went to recall the test and it showed no test to recall. I then uploaded and it did up load.

If so can you send me a copy of the pg 40 if one was uploaded. I erred on the side of caution in case I lost it or it was not complete and I retested the instrument. I will attached that with an explanation of loss of power and attach that to the outgoing inspection results.

Thanks,
V/R

Dan Lyons
Osceola County Sheriff's Office
Spec Ops/Traffic/DUI Enforcement Unit
Agency 407-348-1100
Agency Cell 407-269-1980
Cell 321-960-2066

The reader of this message is not the intended recipient, you are hereby notified that any dissemination, distribution or copy of this communication is strictly prohibited. If you have received this communication in error, please contact the sender by reply E-mail and destroy all copies of the original message.