

Kirkland, Brett

From: Collito, Brad <bcollito@hernandosheriff.org>
Sent: Monday, October 28, 2019 11:05 AM
To: Ben Larison; Conrad Juergensmeyer; David Vitola; Alcohol Testing Program; Ivette Miranda; Klucznik, Shaun; Lisa Yeager (lyeager@sao5.org); Paula Tyrell; Quinones, Daniel; Robinson, Cyrus; Sonya Wolthuis; Stevens, Teresa; Wegner, Steven
Cc: Brown, Kelly W.; Origon, Lance
Subject: Hernando County Sheriff's Office October 2019 Agency
Attachments: October 2019 Agency Inspections.pdf

Good Afternoon,

Attached are the October 2019 Agency Inspections for the Hernando County Sheriff's Office.

Instrument 5249 was inspected with no issues to report. Instrument 5249 is currently being utilized as our primary instrument.

Instrument 814 was not inspected. Instrument 814 is currently awaiting service at CMI, Inc as of 10/16/2019. Instrument 814 lost power and was not able to be powered back on during September 2019.

Instrument 812 was inspected with no issues to report. Instrument 812 is currently being utilized as our back up instrument.

All of our instruments have had their annual inspections completed by FDLE.

All of our simulators have been calibrated by TechSource and we are now on a one year calibration cycle for our simulators.

If I can be of additional assistance, please do not hesitate to contact me.

Respectfully,
Sgt. B. Collito

Sergeant Brad Collito, CCM
Classification Unit
Hernando County Detention Center
16425 Spring Hill Dr
Brooksville, FL 34604
T(352)797-3426 F(352)544-2352
bcollito@hernandosheriff.org

Please note: Florida has a very broad Public Records Law. Most written communications to or from the Hernando County Sheriff's Office are public records available to anyone upon request. Your e-mail, including your e-mail address, may therefore, be subject to public disclosure.

Kirkland, Brett

From: Collito, Brad <bcollito@hernandosheriff.org>
Sent: Sunday, September 15, 2019 11:06 PM
To: Ben Larison; Conrad Juergensmeyer; Alcohol Testing Program; Ivette Miranda; Klucznik, Shaun; Paula Tyrell; Quinones, Daniel; Robinson, Cyrus; Sonya Wolthuis; Stevens, Teresa; Wegner, Steven
Cc: Brown, Kelly W.; Origon, Lance
Subject: Hernando County Sheriff's Office September 2019 Agency Inspections
Attachments: September 2019 Agency Inspections.pdf

Good Afternoon,

Attached are the September 2019 Agency Inspections for the Hernando County Sheriff's Office.

Instrument 5249 was inspected with no issues to report. Instrument 5249 is currently being utilized as our primary instrument.

Instrument 814 was not inspected. Instrument 814 lost power and was not able to be powered back on. No agency inspection was able to be completed on this instrument due to the inability to power it on. Instrument 814 will be shipped back to FDLE on 9/16/2019 for repair at CMI.

Instrument 812 was returned by Mary Bowermaster of the Citrus County Detention Center on 9/15/2019. Instrument 812 was immediately inspected with no issues to report. Instrument 812 was placed back into service and is currently being utilized as our back up instrument.

All of our instruments have had their annual inspections completed by FDLE.

All of our simulators have been calibrated by TechSource and we are now on a one year calibration cycle for our simulators.

If I can be of additional assistance, please do not hesitate to contact me.

Respectfully,
Sgt. B. Collito

Sergeant Brad Collito, CCM
Classification Unit
Hernando County Detention Center
16425 Spring Hill Dr
Brooksville, FL 34604
T(352)797-3426 F(352)544-2352
bcollito@hernandosheriff.org

Deputy Brad Collito
Intake/ Release Unit

Hernando County Detention Center
16425 Spring Hill Dr
Brooksville, FL 34604
T(352)544-2334 F(352)544-2349

Please note: Florida has a very broad Public Records Law. Most written communications to or from the Hernando County Sheriff's Office are public records available to anyone upon request. Your e-mail, including your e-mail address, may therefore, be subject to public disclosure.

Pinsky, David

From: Collito, Brad <bcollito@hernandosheriff.org>
Sent: Sunday, August 25, 2019 1:08 PM
To: Ben Larison; Conrad Juergensmeyer; Alcohol Testing Program; Ivette Miranda; Klucznik, Shaun; Paula Tyrell; Quinones, Daniel; Robinson, Cyrus; Sonya Wolthuis; Stevens, Teresa; Wegner, Steven
Cc: Hillman, Bill; Origen, Lance
Subject: Hernando County Sheriff's Office August 2019 Agency Inspections
Attachments: August 2019 Agency Inspections.pdf; Citrus Custody Receipt.pdf

Good Afternoon,

Attached are the August 2019 Agency Inspections for the Hernando County Sheriff's Office.

Instrument 5249 was inspected with no issues to report. Instrument 5249 is currently being utilized as our back-up instrument.

Instrument 814 was inspected. Instrument 814 is currently being utilized as our primary instrument.

Instrument 812 was inspected with no issues to report. Upon receiving telephonic direction from Dr. Kirkland, Instrument 812 was transferred to Mary Bowermaster from the Citrus County Detention Center after being inspected. Copies of the most recent Agency Inspection and Annual Inspection were transferred with the instrument. Ms. Bowermaster signed a custody receipt upon accepting the instrument (attached). Upon return, I will be completing an additional inspection prior to placing it back into service.

All of our instruments have had their annual inspections completed by FDLE.

All of our simulators have been calibrated by TechSource and we are now on a one year calibration cycle for our simulators.

If I can be of additional assistance, please do not hesitate to contact me.

Respectfully,

Sgt. B. Collito

Sergeant Brad Collito, CCM

Classification Unit

Hernando County Detention Center

16425 Spring Hill Dr

Brooksville, FL 34604

T(352)797-3426 F(352)544-2352

bcollito@hernandosheriff.org<mailto:bcollito@hernandosheriff.org>

<mailto:bcollito@hernandosheriff.org>

Please note: Florida has a very broad Public Records Law. Most written communications to or from the Hernando County Sheriff's Office are public records available to anyone upon request. Your e-mail, including your e-mail address, may therefore, be subject to public disclosure.