

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT – INTOXILYZER 8000

Agency: HILLSBOROUGH CO SO
Time of Inspection: 17:08

Date of Inspection: 11/16/2020

Serial Number: 80-000808
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24419080A2 * Exp: 10/05/2021
0.000	0.049	0.080	0.199	0.082
0.000	0.048	0.080	0.200	0.081
0.000	0.049	0.080	0.200	0.081

Number of Simulators Used: 5

Remarks:


AI J. DAIL #254236 PRESENT FOR INSPECTION.

Julius 254236

** CORRECT LOT NUMBER IS 22419080A2 2/16/09*

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



ROGER G SKIPPER

Signature and Printed Name

11/16/2020
Date

*Reviewed
11-16-20
11/16/20
4/16*

#808 SP

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HILLSBOROUGH CO SO
Time of Inspection: 16:30

Date of Inspection: 02/05/2020

Serial Number: 80-000808
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#: 201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#: 201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 24419000A2 Exp: 10/05/2021
0.000	0.048	0.079	0.202	0.078
0.000	0.048	0.080	0.203	0.078
0.000	0.049	0.080	0.203	0.078

Number of Simulators Used: 5

DGS LOT # 22419080A2 2-7-20

Remarks:

AI J. DAIL #254236 PRESENT FOR INSPECTION.
Open this 2/5/20

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

ROGER G SKIPPER

Signature and Printed Name

02/05/2020
Date

*Reviewed
2-6-20
MSM 2:11 PM
4/20*