

Florida Department of Law Enforcement

Alcohol Testing Program

AGENCY INSPECTION REPORT -- INTOXILYZER 8000

Agency: DIXIE COUNTY SO
 Time of Inspection: 10:26

Date of Inspection: 08/25/2020

Serial Number: 80-000810
 Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) <i>AD1910D</i> Lot#: 10/22/2021 Exp: 10/22/2021	0.08g/210L Test (g/210L) <i>201908D</i> Lot#: 08/07/2021 Exp: 08/07/2021	0.20g/210L Test (g/210L) <i>201902AC</i> Lot#: 02/20/2021 Exp: 02/20/2021 <i>3021</i>	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 22419080A3 Exp: 10/05/2021
0.000	0.047	0.076	0.197	0.078
0.000	0.047	0.077	0.197	0.078
0.000	0.048	0.077	0.197	0.077

Number of Simulators Used: 4

All corrected on 8-28-2020 at 9:57 AM
Matthew Mandola

Remarks:

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

MATTHEW MANDOLA

Signature and Printed Name

Matthew Mandola

08/25/2020

Date