## Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: SUMTER COUNTY SO Time of Inspection: 21:21

Date of Inspection: 08/26/2020

Serial Number: 80-000815

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		110
Diagnostic Check (Pre-Inspection): OK		No
OK	Yes	
Alcohol Free Subject Test: 0.000	165	
Mouth Alcohol Test: Slope Not Met	Yes	
	Yes	
Interferent Detect Test: Interferent Detect	165	
Diagnostic Check (Post-Inspection): OK	Yes	
g	77.0	
	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:22419080A2 Exp: 10/05/2021
0.000	0.047	0.078	0.187 / 0.197	0.079
0.000	0.047	0.077	0.191 / 0.197	0.078
0.000	0.047	0.078	0.192 / 0.197	0.078

Number of Simulators Used: 5

Remarks.

A F / M A: . 20: Control Outside Tolerance.

I forgot to do the mouth wash. I completed the mouth washon the Second afterpt. On the 0.205/2101 Test their was radio interference. I completed the second test without am interference

74m 615 11/14/2020 Amended

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name

08/26/2020 Date



## AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Agency: <u>Sumter County Sheriff's Office</u> Instrument Serial Number: <u>80-000815</u>

AGENCY INSPECTION DATA REVIEW							
Agency Inspector: Terence Hollis		Date of Inspection: 08/26/2020	Time of Inspection: 21:21:56				
□ Agency Inspection Not Conducted or Records regarding Agency Inspection have not been uploaded.							
□ Lot Number □Expiration Date forg/ 210L □Alcohol Reference Solution □Dry Gas Standard is □Incorrect □Expired.							
<ul> <li>□ FDLE/ATP Form 39 states in part, "If a tell Remarks section of FDLE/ATP Form 40 test(s); OR the □ Possible Cause and □ Alcohol Free Subject Test □ 0.05 g/210L Test □</li> </ul>	Agency Inspection R Corrective Action Tal Mouth Alcohol Tes	Report – Intoxilyzer 8000. The ⊠F ken on the following test(s) was no	EASON for repeating the following ot recorded: Interferent Detect Test				
□ FDLE/ATP Form 39 states in part, "If the instrument does not comply with the requirements of Chapter 11D-8, FAC, remove the instrument from service and notify the Department Inspector." □ The Department Inspector was not notified. However, the issue was satisfactorily corrected and the repeated Agency Inspection complies with the requirements of Chapter 11D-8, FAC. □ The Department Inspector was not notified. However, the repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was correctly removed from evidentiary use. □ The Department Inspector was not notified. The repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was not removed from evidentiary use.							
☐ The Agency Inspection is noted as "Con	mplies" when it does	not comply with the requirements	of Chapter 11D-8, FAC.				
□ Other:							
OTHER ELECTRONIC DATA REVIEW							
☐ Login Records	Comments:						
			20 g/210 L Test were repeated. The				
☐ Cylinder Change Records  Date:	reason for repeating the tests, as well as the corrective action taken prior to repeating them, must be included. Please see below for the corrective action and return to your						
☐ Control Test Records Date:	Department Inspector by 10/11/2020.						
☐ Diagnostic Check Records  Date:							
CORRECTIVE ACTION							
Record hand-written amendments on the FDLE/ATP Form 40, Agency Inspection Report, initial and date the amendments, mark the report "AMENDED", and forward a copy to the Department Inspector by 10/11/2020 (Date).							
<ul> <li>□ Provide a written explanation regarding the referenced item(s) to the Department Inspector by (Date).</li> <li>□ Upload the Agency Inspection(s).</li> <li>□ Remove the instrument from evidentiary use until otherwise directed by the Department.</li> </ul>							
□ No action required □ Other:							
Signature of Alcohol Testing Program Staf	H Member	<u>9/11</u> <b>Da</b>	<u>/2020</u> te				