

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FWC
Time of Inspection: 20:29

Date of Inspection: 09/22/2020

Serial Number: 80-000891
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted	Yes	
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#: 201908 B NB Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#: 201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 1060911 Exp: 12/11/2021
0.000	0.049	0.079	0.198	0.082
0.000	0.049	0.079	0.199	0.082
0.000	0.049	0.079	0.199	0.082

Number of Simulators Used: 5

Remarks:
Time-Date changed. Time off by a couple of minutes.
NB.

Lot # for .08 Alcohol solution missing the letter B.
Correct lot # 201908B.

Nicole M. Basford

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Nicole M. Basford NB102

NICOLE M BASFORD

Signature and Printed Name

09/22/2020

Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FWC

Time of Inspection: 00:52

Date of Inspection: 06/23/2020

Serial Number: 80-000891

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1060911 Exp: 12/11/2021
0.000	0.049	0.078	0.196	0.081
0.000	0.049	0.078	0.197	0.081
0.000	0.049	0.079	0.198	0.081

Number of Simulators Used: 5

Remarks:
NA

After this instrument failed the first inspection, I repeated the inspection using a different simulator. After changing simulators, the instrument passed inspection.

*— Nicole Basford (FW NW Agency Inspector)
Nicole Basford N362*

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Nicole Basford #N362

NICOLE M BASFORD

Signature and Printed Name

06/23/2020
Date