

JAX

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FWC
Time of Inspection: 10:42

Date of Inspection: 05/01/2020

Serial Number: 80-000393
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#: 201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#: 201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 00919080A1 Exp: 03/05/2021
0.000	0.045	0.069 / 0.076	0.198	0.078
5.000	0.048	0.074 / 0.077	0.198	0.078
0.000	0.048	0.076 / 0.077	0.199	0.078

Number of Simulators Used: 4

Remarks:

05: Control Outside Tolerance.

.08 Solution not at correct temp. Retest with a tolerance.

The above instrument complies () does not comply () with Chapter 11B-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11B-8, FAC.

RICHARD L GILMORE

Signature and Printed Name

05/01/2020
Date