

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PORT ST JOE PD
Time of Inspection: 15:41

Date of Inspection: 07/14/2020

Serial Number: 80-000908
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:_____ Exp:_____	0.20g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____
0.000	0.000			
0.000	0.000			
0.000	0.000			

Number of Simulators Used: 5

Remarks:

Int Det:CHANGED ACETONE SOLUTION / REHEATED. 05: Control Outside Tolerance. Non-compliance:OUTSIDETO LERANCE.

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Russell L Burch

RUSSELL L BURCH

Signature and Printed Name

07/14/2020
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PORT ST JOE PD
Time of Inspection: 15:33

Date of Inspection: 07/09/2020

Serial Number: 80-000908
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000	0.039 / 0.043			
0.000	0.040 / 0.043			
0.000	0.041 / 0.043			

Number of Simulators Used: 5

Remarks:

05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance:

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Russell L Burch

RUSSELL L BURCH

Signature and Printed Name

07/09/2020
Date

PORT ST JOE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000908
07/14/2020
Software: 8100.27

WET CONTROL TEST

Test	g/210L	Time
Air Blank	0.000	15:50
Control Test	0.000	15:51
Air Blank	0.000	15:51

.05

Sgt Russel + Burl
Operator's Signature

PORT ST JOE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000908
07/14/2020
Software: 8100.27

DRY CONTROL TEST

Test	g/210L	Time
Air Blank	0.000	15:59
Control Test	0.071	16:00
Air Blank	0.000	16:00

Russell J. Bond

Operator's Signature

.08

PORT ST JOE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000908
07/14/2020
Software: 8100.27

WET CONTROL TEST

Test	g/210L	Time
Air Blank	0.000	15:52
Control Test	0.195	15:52
Air Blank	0.000	15:53

Russell Beal

Operator's Signature

.20

PORT ST JOE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000908
07/14/2020
Software: 8100.27

WET CONTROL TEST

Test	g/210L	Time
Air Blank	0.000	15:44
Control Test	0.071	15:44
Air Blank	0.000	15:45

Russell A. Bunn
Operator's Signature

.08

PORT ST JOE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000908
07/14/2020
Software: 8100.27

WET CONTROL TEST

Test	g/210L	Time
Air Blank	0.000	15:48
Control Test	0.071	15:48
Air Blank	0.000	15:49

Russell I Berril
Operator's Signature

.08

PORT ST JOE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000908
07/14/2020
Software: 8100.27

WET CONTROL TEST

Test	g/210L	Time
Air Blank	0.000	15:57
Control Test	0.071	15:58
Air Blank	0.000	15:58

Russell Bunt

Operator's Signature

.08