

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: SANIBEL ISLAND PD  
Time of Inspection: 12:56

2020 <sup>5</sup>  
Date of Inspection: 07/07/~~2010~~

Serial Number: 80-000937  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:28719080A1 Exp: 12/05/2021
0.000	0.048	0.079	0.201	0.079
0.000	0.049	0.080	0.202	0.079
0.000	0.048	0.081	0.202	0.079

Number of Simulators Used: 5

Remarks:

Date should read 07/07/2020 not 07/07/2040 <sup>RS</sup>

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

ROMAN SERRANO

Signature and Printed Name

07/07/2040  
Date