

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA CORRECTIONS

Time of Inspection: 10:51

Date of Inspection: 11/23/2020

Serial Number: 80-000968

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1190213 Exp: 08/28/2022
0.000	0.052	0.082	0.195	0.083
0.000	0.052	0.081	0.197	0.084
0.000	0.052	0.080	0.199	0.084

Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



 CHARLES OWEN MCCONNELL GAYLE

Signature and Printed Name

11/23/2020
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA CORRECTIONS

Time of Inspection: 14:43

Date of Inspection: 10/20/2020

Serial Number: 80-000968

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1190213 Exp: 08/28/2022
0.000	0.050	0.084	0.200	0.082
0.000	0.049	0.084	0.203	0.082
0.000	0.050	0.084	0.203	0.081

Number of Simulators Used: 4

Remarks:

ADDITIONAL TEST DUE TO TRANSMISSION ERROR

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

  CHARLES OWEN MCCONNELL GAYLE
Signature and Printed Name

10/20/2020
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA CORRECTIONS

Time of Inspection: 13:50

Date of Inspection: 10/20/2020

Serial Number: 80-000968

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: 4

Remarks:

A F / M A: Sequence AbortedNo Sample Provided. MISSED M/A TEST HUMAN ERROR.

*Mouth Alcohol was provided within the 3 minutes time frame.
Answer phone*

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Charles Owen McConnell

Charles Owen McConnell

Charles Owen McConnell

CHARLES OWEN MCCONNELL GAYLE

Signature and Printed Name

10/20/2020
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA CORRECTIONS
Time of Inspection: 13:24

Date of Inspection: 10/20/2020

Serial Number: 80-000968
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1190213 Exp: 08/28/2022
0.000	0.050	0.084	0.197	0.080
0.000	0.050	0.084	0.200	0.081
0.000	0.050	0.083	0.201	0.081

Number of Simulators Used: 4

Remarks:

SECOND INSPECTION DUE TO CONTROL OUTSIDE TOLERANCE 0.08

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.





CHARLES OWEN MCCONNELL GAYLE

Signature and Printed Name

10/20/2020
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA CORRECTIONS

Time of Inspection: 10:51

Date of Inspection: 10/20/2020

Serial Number: 80-000968

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____
0.000	0.050	0.086 / 0.086		
0.000	0.050	0.086 / 0.086		
0.000	0.050	0.085 / 0.085		

Number of Simulators Used: 4

Remarks:

OUTSIDE TOLERENCE08 SOLUTION OUTSIDE TOLERANCE.
EPEAT TEST, Control .

. SIMULATOR LID WAS LOOSE TIGHTENED AND R

Test restart

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Charles Owen McConnell Gayle

Charles Owen McConnell Gayle

Charles Owen McConnell Gayle

CHARLES OWEN MCCONNELL GAYLE

Signature and Printed Name

10/20/2020
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA CORRECTIONS

Time of Inspection: 10:54

Date of Inspection: 09/01/2020

Serial Number: 80-000968

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: 5

Remarks:

RESTART. COMP M/A, HIT WRONG KEY / CLOSED OUT INSPECTION. RE-STARTED INSPECTION

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



OWEN CHARLES GAYLE MCCONNELL

Signature and Printed Name

09/01/2020
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA CORRECTIONS

Time of Inspection: 11:51

Date of Inspection: 09/01/2020

Serial Number: 80-000968

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1060911 Exp: 12/11/2021
0.000	0.049	0.080	0.200	0.080
0.000	0.050	0.080	0.201	0.081
0.000	0.049	0.080	0.201	0.080

Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 OWEN CHARLES GAYLE MCCONNELL
Signature and Printed Name

09/01/2020
Date



Florida Department of
Law Enforcement

AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Agency: Osceola Corrections

Instrument Serial Number: 80-000968

AGENCY INSPECTION DATA REVIEW

Agency Inspector: Daniel Lyons

Date of Inspection: 07/14/2020

Time of Inspection: 16:56:00

Agency Inspection Discrepancy: ☐ Incomplete ☐ Untimely/Not Received ☐ Erroneous Information
☐ Procedural ☒ Other (Required Information Missing)

☐ Agency Inspection Not Conducted or Records regarding Agency Inspection have not been uploaded.

☐ Lot Number ☐ Expiration Date for _____ g/ 210L ☐ Alcohol Reference Solution ☐ Dry Gas Standard is ☐ Incorrect ☐ Expired.

☒ FDLE/ATP Form 39 states in part, "If a test must be repeated, the REASON must be entered when prompted and recorded in the Remarks section of FDLE/ATP Form 40 Agency Inspection Report – Intoxilyzer 8000. The ☒ REASON for repeating the following test(s); OR the ☒ Possible Cause and Corrective Action Taken on the following test(s) was not recorded:

☐ Alcohol Free Subject Test ☐ Mouth Alcohol Test ☒ Alcohol Free Test ☐ Interferent Detect Test

☐ 0.05 g/210L Test ☐ 0.08 g/210L Test ☐ 0.20 g/210L Test ☐ 0.08 g/210L Dry Gas Standard Test

☐ FDLE/ATP Form 39 states in part, "If the instrument does not comply with the requirements of Chapter 11D-8, FAC, remove the instrument from service and notify the Department Inspector."

☐ The Department Inspector was not notified. However, the issue was satisfactorily corrected and the repeated Agency Inspection complies with the requirements of Chapter 11D-8, FAC.

☐ The Department Inspector was not notified. However, the repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was correctly removed from evidentiary use.

☐ The Department Inspector was not notified. The repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was not removed from evidentiary use.

☐ The Agency Inspection is noted as "Complies" when it does not comply with the requirements of Chapter 11D-8, FAC.

☐ Other: _____

OTHER ELECTRONIC DATA REVIEW

☐ Login Records
Date: _____

☐ Cylinder Change Records
Date: _____

☐ Control Test Records
Date: _____

☐ Diagnostic Check Records
Date: _____

Comments:

The Alcohol Free Test was repeated. The reason for the repeated test and the corrective action taken prior to repeating the test must be recorded.

Please see below for the corrective action required and send to the Department Inspector by 09/20/2020.

CORRECTIVE ACTION

☒ Record hand-written amendments on the FDLE/ATP Form 40, Agency Inspection Report, initial and date the amendments, mark the report "AMENDED", and forward a copy to the Department Inspector by 09/20/2020 (Date).

☐ Provide a written explanation regarding the referenced item(s) to the Department Inspector by _____ (Date).

☐ Upload the Agency Inspection(s).

☐ Remove the instrument from evidentiary use until otherwise directed by the Department.

☐ No action required

☐ Other: _____

Signature of Alcohol Testing Program Staff Member

8/20/2020
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA CORRECTIONS

Time of Inspection: 16:56

Date of Inspection: 07/14/2020

Serial Number: 80-000968

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1060911 Exp: 12/11/2021
0.000 / 0.000	0.048	0.079	0.197	0.080
RFI / 0.000	0.047	0.079	0.198	0.080
/ 0.000	0.048	0.079	0.199	0.080

Number of Simulators Used: 5


Remarks:

00: RFI Detect.

D/S WALKED IN ROOM W/PORTABLE RADIO, ON. GUIDANCE WAS PROVIDED TO THE DEPT. TEST WAS RETESTED, NOT RE-STARTED.

The above instrument complies (☒) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 DANIEL E LYONS
Signature and Printed Name

07/14/2020
Date

Florida Department of Law Enforcement

Alcohol Testing Program

17754

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA CORRECTIONS

Time of Inspection: 15:36

Date of Inspection: 08/03/2020

Serial Number: 80-000968

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1060911 Exp: 12/11/2021
0.000	0.050	0.079	0.200	0.079
0.000	0.050	0.079	0.202	0.080
0.000	0.050	0.079	0.202	0.080

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Daniel E Lyons
DANIEL E LYONS

Signature and Printed Name

08/03/2020
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA CORRECTIONS
Time of Inspection: 16:56

Date of Inspection: 07/14/2020

Serial Number: 80-000968
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1060911 Exp: 12/11/2021
0.000 / 0.000	0.048	0.079	0.197	0.080
RFI / 0.000	0.047	0.079	0.198	0.080
/ 0.000	0.048	0.079	0.199	0.080

Number of Simulators Used: 5

Remarks:

00: RFI Detect.

*DIS WALKED INTO ROOM WITH PORTABLE RADIO ON.
GUIDANCE WAS PROVIDED TO THE DEPUTY. TEST WAS RETESTED
NOT RESTARTED.*

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

DANIEL E LYONS

Signature and Printed Name

07/14/2020
Date

Florida Department of Law Enforcement Alcohol Testing Program

80-000968
BK

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA CORRECTIONS

Time of Inspection: 14:09

Date of Inspection: 02/26/2020

Serial Number: 80-000968

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2020	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1060911 Exp: 12/11/2021
0.000	0.049	0.076	0.195	0.080
0.000	0.049	0.077	0.196	0.079
0.000	0.048	0.077	0.194	0.080

Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



DANIEL E LYONS
Signature and Printed Name

02/26/2020
Date

Florida Department of Law Enforcement Alcohol Testing Program

80-000968
BK

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA CORRECTIONS

Time of Inspection: 17:52

Date of Inspection: 02/07/2020

Serial Number: 80-000968

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1060911 Exp: 12/11/2021
0.000	0.050	0.078	0.197	0.079
0.000	0.050	0.079	0.198	0.080
0.000	0.050	0.079	0.198	0.080

Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



DANIEL E LYONS

Signature and Printed Name

02/07/2020

Date

Florida Department of Law Enforcement Alcohol Testing Program

80-000968
13K

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA CORRECTIONS

Time of Inspection: 14:49

Date of Inspection: 01/24/2020

Serial Number: 80-000968

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1060911 Exp: 12/11/2021
0.000	0.050	0.077	0.194	0.080
0.000	0.050	0.077	0.195	0.080
0.000	0.050	0.077	0.195	0.080

Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



DANIEL E LYONS

Signature and Printed Name

01/24/2020
Date

Florida Department of Law Enforcement Alcohol Testing Program

80-000968
BK

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA CORRECTIONS

Time of Inspection: 10:33

Date of Inspection: 01/03/2020

Serial Number: 80-000968

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1060911 Exp: 12/11/2021
0.000	0.051	0.076	0.195	0.079
0.000	0.050	0.077	0.195	0.079
0.000	0.050	0.077	0.194	0.079

Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



DANIEL E LYONS

Signature and Printed Name

01/03/2020
Date