

1049
SP

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PASCO COUNTY SO
Time of Inspection: 14:42

Date of Inspection: 03/22/2020

Serial Number: 80-001049
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____
/ 0.000	0.050	0.080	0.171 / 0.187	
/ 0.000	0.050	0.080	0.182 / 0.192	
/ 0.000	0.049	0.079	0.186 / 0.193	

Number of Simulators Used: 4

Remarks:

A F / M A: Interferent Detect. 00: Ambient Fail. 20: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

On 0.20g/210L Test I did not let the solution warm up enough

Taken out of service due to

- Interferent Detect
- Ambient Fail

- unable to upload inspection - First read "no dil test" error thought ok & inst. uploaded from

BS 3/22/2020

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

BARRY NIXON

Signature and Printed Name

03/22/2020
Date

1049
SP

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PASCO COUNTY SO
Time of Inspection: 14:14

Date of Inspection: 01/15/2020

Serial Number: 80-001049
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:14219080A2 Exp: 08/05/2021
0.000	0.050	0.080	0.201	0.080
0.000	0.050	0.080	0.203	0.080
0.000	0.050	0.081	0.203	0.081

Number of Simulators Used: 4

Remarks:

A F / M A: No Sample Provided. REPEATED TEST

*Did not provide sample in time.
By 3932 1/15/2020*

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

BARRY NIXON
Signature and Printed Name

01/15/2020
Date