

# Florida Department of Law Enforcement Alcohol Testing Program

80-001054  
TSK

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO  
Time of Inspection: 02:51

Date of Inspection: 02/13/2020

Serial Number: 80-001054  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:17919080A1 Exp: 08/05/2021
/ 0.000	0.045	0.077	0.191	0.079
/ 0.000	0.046	0.077	0.193	0.079
/ 0.000	0.046	0.077	0.195	0.079

Number of Simulators Used: 5

**Remarks:**

00: Ambient Fail.

Re did test due to order in Rwm,

A TRUE COPY  
PAUL BLACKMAN, SHERIFF  
HIGHLANDS COUNTY, FLORIDA  
BY \_\_\_\_\_ D.S.

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

\_\_\_\_\_  
Signature and Printed Name

GREGORY S GLASCO

02/13/2020  
Date