

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PASCO COUNTY SO
Time of Inspection: 15:25

Date of Inspection: 10/22/2020

Serial Number: 80-001074
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:14020080A1 Exp: 07/05/2022
0.000	0.049	0.080	0.200	0.081
0.000	0.049	0.081	0.201	0.081
0.000	0.050	0.081	0.201	0.081

Number of Simulators Used: 4

Remarks:

INST NOT RESPONDING

At the last "Review Remarks" the inst. would not respond to the keyboard. Pushed "Green Start button" and inst. responded. Inst. will clearly finish with inspection and was in compliance.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

3932

BARRY NIXON

Signature and Printed Name

10/22/2020
Date

1074
SR

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PASCO COUNTY SO
Time of Inspection: 13:59

Date of Inspection: 01/15/2020

Serial Number: 80-001074
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:14219080A2 Exp: 08/05/2021
0.000	0.048	0.078	0.191	0.079
0.000	0.049	0.078	0.199	0.079
0.000	0.049	0.078	0.200	0.079

Number of Simulators Used: 4

Remarks:

A F / M A: No Sample Provided. REPEATED TEST

*DL not found sig in hr
DW 397
1/15/20*

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

BARRY NIXON
Signature and Printed Name

01/15/2020
Date