

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PASCO COUNTY SO
Time of Inspection: 14:44

Date of Inspection: 10/22/2020

Serial Number: 80-001081
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:22419080A3 Exp: 10/05/2021
0.000	0.034 / 0.048	0.078	0.191	0.082
0.000	0.038 / 0.048	0.079	0.199	0.081
0.000	0.046 / 0.048	0.078	0.200	0.081

Number of Simulators Used: 4

Remarks:

05: Control Outside Tolerance.

*Tightest simulator thru repeated
3932
10/22/2020*

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

3932

BARRY NIXON
Signature and Printed Name

10/22/2020
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PASCO COUNTY SO
Time of Inspection: 14:44

Date of Inspection: 10/22/2020

Serial Number: 80-001081
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:22419080A3 Exp: 10/05/2021
0.000	0.034 / 0.048	0.078	0.191	0.082
0.000	0.038 / 0.048	0.079	0.199	0.081
0.000	0.046 / 0.048	0.078	0.200	0.081

Number of Simulators Used: 4

Remarks:

05: Control Outside Tolerance.

*Tightened simulator then reported.
80 3932
10/22/2020*

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

BARRY NIXON BARRY NIXON
Signature and Printed Name

10/22/2020
Date

1081
SP

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PASCO COUNTY SO
Time of Inspection: 14:14

Date of Inspection: 01/15/2020

Serial Number: 80-001081
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:22419080A3 Exp: 10/05/2021
0.000	0.048	0.079	0.202	0.082
0.000	0.048	0.080	0.202	0.081
0.000	0.049	0.079	0.202	0.081

Number of Simulators Used: 4

Remarks:

A F / M A: No Sample Provided. REPEATED TEST

*Do not provide sample in time
Per 3920
1/15/2020*

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

BARRY NIXON

Signature and Printed Name

01/15/2020
Date