

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: VOLUSIA COUNTY S.O.
Time of Inspection: 14:43

Date of Inspection: 08/31/2020

Serial Number: 80-001130
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1206004 Exp: 09/20/2022
0.000	0.046	0.079	0.199	0.000 / 0.079
0.000	0.047	0.079	0.199	0.000 / 0.079
0.000	0.047	0.078	0.199	0.000 / 0.079

Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance.

Dry gas standard was not connected during first test. Connected D.G.S. and re conducted test.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



JIMMIE N STONE

Signature and Printed Name

08/31/2020
Date