AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO

Serial Number: 80-001139

Time of Inspection: 13:40

Date of Inspection: 12/10/2020 Software: 8100.27

Check or Test	YES	МО
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:970670 Exp: 11/18/2020
0.000	0.048	0.078	0.198	0.081 06/07/2021
0.000	0.049	0.080	0.202	0.081
0.000	0.049	0.080	0.203	0.081

Number of Simulators Used: 5

Remarks:

B ARMSTRONG PRESENT AND PARTICIPATED.

The above	instrument complies (X) does not comply () with Chapter 11D-8, F	AC.
I certify performed	that I hold a valid Florida Department of Law Enforcement Agency Inspection in accordance with the provisions of Chapter 11D-8, FAC.	ctor Permit and that I
	JOSHUA R SAPP	

S gnature and Printed Name

12/10/2020 Date

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agend	cy:	Broward
Time	of	Inspection:

Broward County SO

Date of Inspection:

Serial Number: 80-001139

Software: 8100.27

CHECK OR TEST	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		
Alcohol Free Subject Test: 0.000		
Mouth Alcohol Test: Slope Not Met		
Interferent Detect Test: Interferent Detect		
Diagnostic Check (Post-Inspection): OK		

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
			E	

Number	of	Simulators	Used:	
--------	----	------------	-------	--

Remarks: On 02-19-2020 I signed into the instruments (80-001139 & 80-006924) multiple times in an attempt to upload the agency inspections. I tried different analog lines in our building to see if we were having an issue with our lines. I also tried a different instrument (80-006924) to ensure there was no issue with the instrument. Ultimately ITD was contacted and they responded to check our lines. The lines in our office were working properly as a result I contacted our department inspector, Michael Haughty. Mr. Haughty advised he would send an email to Tallahassee and request they reset the server. An attempt was made 02-20-2020 on instrument 80-001139 and the agency inspection was successfully uploaded. The remaining instruments were uploaded to FDLE on 02-21-2020.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.
I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I
performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.
11/4/1/802/0
Signature and Printed Name
July 1111 confidence
6/00/0A
1 2/28/20
The state of the s
Date

AGENCY INSPECTION REPORT - INTOXILYZER 8000

7 ~~~~~		
	7	_

Broward County SO

Time of Inspection: Date of Inspection:

Serial Number: 80-006924

Software: 8100.27

CHECK OR TEST	YES	NO
Date and/or Time Adjusted		1
Diagnostic Check (Pre-Inspection): OK		
Alcohol Free Subject Test: 0.000		+
Mouth Alcohol Test: Slope Not Met		
Interferent Detect Test: Interferent Detect		+
Diagnostic Check (Post-Inspection): OK		+

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

•	_		
Number	of	Simulators	Used:

Remarks: On 02-19-2020 I signed into the instruments (80-001139 & 80-006924) multiple times in an attempt to upload the agency inspections. I tried different analog lines in our building to see if we were having an issue with our lines. I also tried a different instrument (80-006924) to ensure there was no issue with the instrument. Ultimately ITD was contacted and they responded to check our lines. The lines in our office were working properly as a result I contacted our department inspector, Michael Haughty. Mr. Haughty advised he would send an email to Tallahassee and request they reset the server. An attempt was made 02-20-2020 on instrument 80-001139 and the agency inspection was successfully uploaded. The remaining instruments were uploaded to FDLE on 02-21-2020.

The above instrument complies (\(\) does not comply () with Chapter 11D-8, FAC.	
I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and the performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.	hat I
Signature and Printed Name 2/28/20 Date	

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO Time of Inspection: 16:16

Date of Inspection: 01/16/2020

Serial Number: 80-001139

Software: 8100.27

Check or Test	YES	NO.
Date and/or Time Adjusted		
		No
Diagnostic Check (Pre-Inspection): OK		
	Yes	
Alcohol Free Subject Test: 0.000		
	Yes	
Mouth Alcohol Test: Slope Not Met		
	Yes	1
Interferent Detect Test: Interferent Detect		
	Yes	
Diagnostic Check (Post-Inspection): OK		
	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:970670 Exp: 06/07/2021
0.000	0.049	0.079	0.198	0.082
0.000	0.049	0.079	0.200	0.082
0.000	0.049	0.079	0.199	0.081

Number of Simulators Used: 5

A FRAZIER PRESENT & PARTICIPATED.

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name

01/16/2020 Date