

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: Broward County SO

Serial Number: 80-001139

Time of Inspection:

Date of Inspection:

Software: 8100.27

CHECK OR TEST	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		
Alcohol Free Subject Test: 0.000		
Mouth Alcohol Test: Slope Not Met		
Interferent Detect Test: Interferent Detect		
Diagnostic Check (Post-Inspection): OK		


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: _____

Remarks: On 02-19-2020 I signed into the instruments (80-001139 & 80-006924) multiple times in an attempt to upload the agency inspections. I tried different analog lines in our building to see if we were having an issue with our lines. I also tried a different instrument (80-006924) to ensure there was no issue with the instrument. Ultimately ITD was contacted and they responded to check our lines. The lines in our office were working properly as a result I contacted our department inspector, Michael Haugthy. Mr. Haugthy advised he would send an email to Tallahassee and request they reset the server. An attempt was made 02-20-2020 on instrument 80-001139 and the agency inspection was successfully uploaded. The remaining instruments were uploaded to FDLE on 02-21-2020.

The above instrument complies () does not comply (____) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


 Signature and Printed Name
 Date: 2/28/20

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: Broward County SO

Serial Number: 80-006924

Time of Inspection:

Date of Inspection:

Software: 8100.27

CHECK OR TEST	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		
Alcohol Free Subject Test: 0.000		
Mouth Alcohol Test: Slope Not Met		
Interferent Detect Test: Interferent Detect		
Diagnostic Check (Post-Inspection): OK		

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: _____

Remarks: On 02-19-2020 I signed into the instruments (80-001139 & 80-006924) multiple times in an attempt to upload the agency inspections. I tried different analog lines in our building to see if we were having an issue with our lines. I also tried a different instrument (80-006924) to ensure there was no issue with the instrument. Ultimately ITD was contacted and they responded to check our lines. The lines in our office were working properly as a result I contacted our department inspector, Michael Haughty. Mr. Haughty advised he would send an email to Tallahassee and request they reset the server. An attempt was made 02-20-2020 on instrument 80-001139 and the agency inspection was successfully uploaded. The remaining instruments were uploaded to FDLE on 02-21-2020.

The above instrument complies () does not comply (____) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



Signature and Printed Name

2/28/20

Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO
Time of Inspection: 16:16

Date of Inspection: 01/16/2020

Serial Number: 80-001139
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:970670 Exp: 06/07/2021
0.000	0.049	0.079	0.198	0.082
0.000	0.049	0.079	0.200	0.082
0.000	0.049	0.079	0.199	0.081

Number of Simulators Used: 5

Remarks:

A FRAZIER

Present & PARTICIPATED.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

J. Rapp 17593

 Signature and Printed Name JOSHUA R SAPP

01/16/2020
Date