

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI PD  
Time of Inspection: 21:24

Date of Inspection: 06/30/2020

Serial Number: 80-001179  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted	Yes	
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:867139 Exp: 11/15/2020 <i>SEE REMARKS</i>	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:867139 Exp: 11/15/2020
0.000	0.049	0.079	0.199	0.079
0.000	0.050	0.080	0.198	0.080
0.000	0.050	0.080	0.199	0.080

Number of Simulators Used: 5

Remarks:  
Time-Date changed.

*INSPECTION COMPLETE.*  
*INTOX IN COMPLIANCE.*  
*LOT AND EXP. DATE FOR 0.20g/210L TEST SHOULD READ 201902C/02/20/2021*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*J. S. Blanco* N827 JUAN S BLANCO  
\_\_\_\_\_  
Signature and Printed Name

06/30/2020  
Date