

125e
SP

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ORANGE COUNTY S.O.
Time of Inspection: 14:54

Date of Inspection: 01/07/2020

Serial Number: 80-001256
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:13518080A5 Exp: 08/05/2020
0.000	0.051	0.081	0.204	0.000 / 0.079
0.000	0.050	0.081	0.205	0.080 / 0.080
0.000	0.050	0.081	0.204	0.079 / 0.079

Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance ATTACHED DGS TUBE ^{1st} after sample REPEATED DGS STANDARD TEST AFTER ATTACHING DGS TUBE
dm

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name

KELLY M MELVILLE

01/07/2020
Date