

1257  
SP

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: Orange County Sheriff's Office

Serial Number: 80-001257

Time of Inspection: N/A

Date of Inspection: 04/02/2020

Software:

CHECK OR TEST	YES	NO
Date and/or Time Adjusted	N/A	N/A
Diagnostic Check (Pre-Inspection): OK	N/A	N/A
Alcohol Free Subject Test: 0.000	N/A	N/A
Mouth Alcohol Test: Slope Not Met	N/A	N/A
Interferent Detect Test: Interferent Detect	N/A	N/A
Diagnostic Check (Post-Inspection): OK	N/A	N/A

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: N/A Exp: N/A	0.08g/210L Test (g/210L) Lot#: N/A Exp: N/A	0.20g/210L Test (g/210L) Lot#: N/A Exp: N/A	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: N/A Exp: N/A
N/A	N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A	N/A


Number of Simulators Used:     N/A    

Remarks: Not able to perform an Agency Inspection due to the instrument not staying powered on when the green Start Test button was pushed. Additionally, prior to the power issue of 04/02/2020, when the instrument was powered on, the instrument had displayed a "DISABLED MODE MEMORY FULL" message, despite an upload of the instrument's data on 03/19/2020. I notified Department Inspector Shayla Platt via telephone of these issues. I am sending the instrument to CMI for analysis.

(All of the information entered on this Agency Inspection Report was hand-typed by Agency Inspector K. Melville)

The above instrument complies (   N/A   ) does not comply (   N/A   ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



Kelly M. Melville

Signature and Printed Name

04/02/2020  
Date