

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BAKER COUNTY SO
Time of Inspection: 01:38

Date of Inspection: 09/21/2020

Serial Number: 80-001287
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24818080A3 Exp: 10/05/2020
0.000	0.042 / 0.046	0.075	0.199	0.078
0.000	/ 0.045	0.076	0.200	0.078
0.000	/ 0.045	0.075	0.200	0.078

Number of Simulators Used: 5

Remarks:
05: Improper Sample.

RECEIVED

SEP 21 2020

Records Dept.
Baker County Sheriffs Office

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Sg CM 409

CALEB A COLLINS

Signature and Printed Name

09/21/2020
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BAKER COUNTY SO
Time of Inspection: 20:22

Date of Inspection: 05/31/2020

Serial Number: 80-001287
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Gas Std Test (g/210L) Lot#:27818080A3 Exp: 05/05/2020
0.000	0.044 / 0.048	0.077	0.199	0.079
0.000	0.044 / 0.048	0.077	0.198	0.079
0.000	0.043 / 0.047	0.076	0.199	0.079

Number of Simulators Used: 5

Remarks:

05: Control Outside Tolerance.

RECEIVED

JUN 15 2020

Records Dept.
Baker County Sheriff's Office

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Sof C CA 409

CALEB A COLLINS

Signature and Printed Name

05/31/2020
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BAKER COUNTY SO
Time of Inspection: 19:23

Date of Inspection: 05/31/2020

Serial Number: 80-001287
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____
0.000 / 0.000				

Number of Simulators Used: 5

Remarks:
00: Ambient Fail, Ambient Fail. Non-compliance:


RECEIVED

JUN 15 2020

Records Dept.
Baker County Sheriffs Office

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that performed this inspection in accordance with the provisions of Chapter 11D-8, FAC:



 Signature and Printed Name CALEB A COLLINS

05/31/2020
 Date

SP 1287

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT -- INTOXILYZER 8000

Agency: BAKER COUNTY SO
Time of Inspection: 16:35

Date of Inspection: 04/29/2020

Serial Number: 80-0012
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.05g/210L Dry Gas Std Test (g/210L) Lot#:24818080A3 Exp: 05/05/2020
0.000	0.046	0.077	0.198	0.079
0.000	0.046	0.077	0.200	0.079
0.000	0.046	0.077	0.200	0.079

Number of Simulators Used: 5

Remarks:

RECEIVED

APR 30 2020

Records Dept.
Baker County Sheriffs Office

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Sgt. Will Yous *Derek Watson* 382

Signature and Printed Name

CALEB COLLINS / DEREK WATSON

04/29/2020
Date

Florida Department of Law Enforcement Alcohol Testing Program

SP
1287

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BAKER COUNTY SO
Time of Inspection: 02:36

Date of Inspection: 04/29/2020

Serial Number: 80-001207
Software: 8100.27

Check or Test	YES	
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000	0.042			
0.000	0.042			
0.000	0.042			

Number of Simulators Used: 5

Remarks:

05: Control Outside Tolerance. Non-compliance: OUTSIDE .045 TO .055.

*Failed due to older Alcohol Reference Solution.
Used new solution & passed*

RECEIVED

APR 30 2020

Records Dept.
Baker County Sheriffs Office

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Spencer M 409

Signature and Printed Name

CALEB A COLLINS

04/29/2020
Date

Florida Department of Law Enforcement Alcohol Testing Program

80-001287
TBL

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BAKER COUNTY SO
Time of Inspection: 15:33

Date of Inspection: 02/28/2020

Serial Number: 80-001287
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24818080A3 Exp: 10/05/2020
0.000	0.049	0.079	0.199	0.081
0.000	0.049	0.080	0.200	0.081
0.000	0.049	0.080	0.200	0.081

Number of Simulators Used: 5

Remarks:
Int Det:NO INT DET.

RECEIVED

FEB 28 2020

Records Dept.
Baker County Sheriffs Office

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

[Signature]

Sgt. C. M. 109

WATSON, DEREK / COLLINS, CALEB

Signature and Printed Name

02/28/2020
Date