

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: GULF BREEZE PD
Time of Inspection: 09:45

Date of Inspection: 11/09/2020

Serial Number: 80-001306
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | Yes | |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020 | 0.08g/210L Test (g/210L) Lot#:202007A Exp: 07/21/2022 | 0.20g/210L Test (g/210L) Lot#:202005A Exp: 05/12/2022 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1094494 Exp: 02/19/2022 |
|----------------------------|---|---|---|--|
| 0.000 | 0.046 | 0.079 | 0.081 / 0.198 | 0.077 |
| 0.000 | 0.047 | 0.080 | 0.080 / 0.200 | 0.077 |
| 0.000 | 0.047 | 0.080 | 0.080 / 0.200 | 0.077 |

Number of Simulators Used: 4

Remarks:

Time-Date changed. 20: Control Outside Tolerance.

Time adjusted for the end of Daylight Savings Time - JLR
Inadvertently started .2 test before changing simulator. - JLR

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Thomas C. Roberson

THOMAS C ROBERSON
Signature and Printed Name

11/09/2020
Date

▶CCCCCCCC

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: GULF BREEZE PD
Time of Inspection: 11:21

Date of Inspection: 09/05/2020

Serial Number: 80-001306
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020 | 0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021 | 0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1298240 Exp: 03/19/2023 |
|----------------------------|---|---|---|--|
| 0.000 | 0.047 | 0.078 | 0.197 | 0.000 / 0.081 |
| 0.000 | 0.047 | 0.080 | 0.198 | 0.000 / 0.081 |
| 0.000 | 0.047 | 0.080 | 0.198 | 0.000 / 0.081 |

Number of Simulators Used: 4

Remarks:

Int Det: .08: Control Outside Tolerance.

Two bottles not connected properly x 2. - JKL

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Thomas C. Roberson

THOMAS C ROBERSON

Signature and Printed Name

09/05/2020

Date