

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: GAINESVILLE PD
Time of Inspection: 20:34

Date of Inspection: 11/25/2020

Serial Number: 80-001364
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:30219080A4 Exp: 01/05/2022
0.000	0.047	0.078	0.199	0.079
0.000	0.047	0.078	0.199	0.079
0.000	0.047	0.078	0.199	0.079

Number of Simulators Used: 4

Remarks:
IN COMPLIANCE IN COMPLIANCE.

The above instrument complies (^{as} X) does not comply (~~as~~) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



AARON M STEMEN

Signature and Printed Name

11/25/2020
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: GAINESVILLE PD
Time of Inspection: 21:32

Date of Inspection: 06/29/2020

Serial Number: 80-001364
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000	0.042 / 0.042			
0.000	0.042 / 0.042			
0.000	0.042 / 0.042			

Number of Simulators Used: 4

Remarks:

05: Control Outside Tolerance, Control Outside Tolerance. .05 OUTSIDE TOL.

- Relocated to new room, verified connections, seals, and hoses on simulators.
- Conducted a new inspection which was in compliance.
- Contacted Brett Kirkland at FDLE ATP and was advised to keep instrument in service since subsequent was in compliance after verifying the issue was not with the instrument.

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



AARON M STEMEN

Signature and Printed Name

06/29/2020
Date

GAINESVILLE PD

Intoxilyzer - Alcohol Analyzer


Model 8000

SN 80-001364

06/29/2020

Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	21:33
Control Test	0.071	21:34
Air Blank	0.000	21:35
Control Test	0.071	21:35
Air Blank	0.000	21:36
Control Test Stats		
Average	0.0710	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	


Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT – INTOXILYZER 8000

Agency: GAINESVILLE PD
Time of Inspection: 23:31

Date of Inspection: 06/29/2020

Serial Number: 80-001364
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24818080A2 Exp: 10/05/2020
0.000	0.047	0.079	0.202	0.079
0.000	0.048	0.079	0.202	0.080
0.000	0.048	0.079	0.202	0.080

Number of Simulators Used: 4

Remarks:
IN COMPLIANCE

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



AARON M STEMEN

Signature and Printed Name

06/29/2020
Date

INTOXILYZER 8000
 Instrument Initialization
 22:15 06/29/2020

GAINESVILLE PD
 Intoxilyzer - Alcohol Analyzer
 Model: 8000 SN 80-001364
 06/29/2020
 Software: 8100.27


Test	g/210L	Time
Air Blank	0.000	22:28
Control Test	0.048	22:29
Air Blank	0.000	22:29
Control Test	0.048	22:30
Air Blank	0.000	22:30
Control Test Stats		
Average	0.0480	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	



 Operator's Signature

GAINESVILLE PD
 Intoxilyzer - Alcohol Analyzer
 Model: 8000 SN 80-001364
 06/29/2020
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	22:34
Control Test	0.078	22:35
Air Blank	0.000	22:35
Control Test	0.077	22:36
Air Blank	0.000	22:37
Control Test Stats		
Average	0.0775	
Std Dev	0.0007	
Rel Std Dev(%)	0.9124	



 Operator's Signature

GAINESVILLE PD
 Intoxilyzer - Alcohol Analyzer
 Model: 8000 SN 80-001364
 06/29/2020
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	22:38
Control Test	0.201	22:39
Air Blank	0.000	22:39
Control Test	0.199	22:40
Air Blank	0.000	22:40
Control Test Stats		
Average	0.2000	
Std Dev	0.0014	
Rel Std Dev(%)	0.7071	



 Operator's Signature

GAINESVILLE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001364
06/30/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	22:05
Control Test	0.078	22:06
Air Blank	0.000	22:06
Control Test	0.079	22:06
Air Blank	0.000	22:07
Control Test Stats		
Average	0.0785	
Std Dev	0.0007	
Rel Std Dev(%)	0.9008	



Operator's Signature