

SP 1413

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: Orange County Sheriff's Office

Serial Number: 80-001418

Time of Inspection: N/A

Date of Inspection: 08/03/2020

Software:

CHECK OR TEST	YES	NO
Date and/or Time Adjusted	N/A	N/A
Diagnostic Check (Pre-Inspection): OK	N/A	N/A
Alcohol Free Subject Test: 0.000	N/A	N/A
Mouth Alcohol Test: Slope Not Met	N/A	N/A
Interferent Detect Test: Interferent Detect	N/A	N/A
Diagnostic Check (Post-Inspection): OK	N/A	N/A

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: N/A Exp: N/A	0.08g/210L Test (g/210L) Lot#: N/A Exp: N/A	0.20g/210L Test (g/210L) Lot#: N/A Exp: N/A	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: N/A Exp: N/A
N/A	N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A	N/A

Number of Simulators Used: N/A

Remarks: Did not attempt to perform an Agency Inspection due to the reported issue on 07/24/2020 where the instrument was reported to have shut itself down during a breath test and begun emitting black smoke. Department Inspector Shayla Platt was notified via telephone of these issues, and of the fact that I am not attempting to power the instrument on today. I am sending the instrument to CMI for analysis.

OK

(All of the information entered on this Agency Inspection Report was hand-typed by Agency Inspector K. Melville)

The above instrument complies (N/A) does not comply (N/A) with Chapter 11D-8, FAC.

OK

OK I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



Signature and Printed Name

Kelly M. Melville

08/03/2020
Date

1418
SP

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ORANGE COUNTY S.O.
Time of Inspection: 14:40

Date of Inspection: 01/07/2020

Serial Number: 80-001418
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) 4 4A Lot#:06/19080A1 Exp: 05/05/2021
0.000	0.048	0.080	0.202	0.079
0.000	0.049	0.080	0.203	0.079
0.000	0.049	0.080	0.202	0.079

Number of Simulators Used: 5

Remarks:

The DGS Lot number is "06419080A1", Expiration Date "05/05/2021"

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name

KELLY M MELVILLE

01/07/2020
Date