

AMENDED

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PINELLAS PARK P.D.

Time of Inspection: 19:18

Date of Inspection: 07/27/2020

Serial Number: 80-001732

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:28719080A1 Exp: 12/05/2021
0.000	0.048	0.079	0.197	0.081
0.000	0.048	0.079	0.199	0.080
0.000	0.048	0.080	0.199	0.080

Number of Simulators Used: 5

Remarks:

A F / M A: .


Mouth alcohol repeated due to too much mouth alcohol initially being used. Mouth was rinsed out and test repeated without issue.

② SJR

08-20-20

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



JACOB M ROLLESTON

Signature and Printed Name

07/27/2020
Date

AMENDED

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PINELLAS PARK P.D.
Time of Inspection: 18:22

Date of Inspection: 05/04/2020

Serial Number: 80-001732
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#: 201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#: 201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 28719080A1 Exp: 12/05/2021
0.000	0.048	0.049 / 0.078	0.196	0.080
0.000	0.048	0.049 / 0.079	0.199	0.080
0.000	0.049	0.049 / 0.080	0.199	0.080

Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance.

Lid of the simulator was not sealed properly. Lid was properly re-attached and allowed time to reach appropriate standards.

*JS19
08-20-20*

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JS19

JACOB M ROLLESTON
Signature and Printed Name

05/04/2020
Date