AMENDED

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIALEAH PD

Serial Number: 80-002462

Time of Inspection: 14:21

Date of Inspection: 06/25/2020

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
		No
Diagnostic Check (Pre-Inspection): OK		
	Yes	
Alcohol Free Subject Test: 0.000		
		No
Mouth Alcohol Test: Slope Not Met		
		No
Interferent Detect Test: Interferent Detect		
		No
Diagnostic Check (Post-Inspection): OK		
		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number	of	Simulators	Used:	5
	~-	D-TWGTGCCT D		_

Remarks:

A F / M A: Range Exceeded. Non-compliance: .

TOO MUCH MOUTH ACCOHOL STILL RESIDUAL # 06-25-2020

The above instrumer	at complies () does not comp	ly (X) with Chapte	er 11D-8, FAC.	
1 1	hold a valid Florida Department section in accordance with the pro-	of Law Enforcement Agovisions of Chapter 11D-	ency Inspector Permit -8, FAC.	and that I
	#1614	CESAR	CORREA	
	Signature	and Printed Name		
•		25 (222		

06/25/2020 Date



AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Instrument Serial Number: 80-002462

Agency: Hialeah Police Department

AGE	ENCY INSPECTION DATA REVIEW
Age	ncy Inspector: Cesar Correa Date of Inspection: 06/25/2020 Time of Inspection: 14:21:25
Age	ncy Inspection Discrepancy: Incomplete Untimely/Not Received Erroneous Information Procedural Other
	Agency Inspection Not Conducted or Records regarding Agency Inspection have not been uploaded.
	Lot Number □Expiration Date forg/ 210L □Alcohol Reference Solution □Dry Gas Standard is □Incorrect □Expired.
×	FDLE/ATP Form 39 states in part, "If a test must be repeated, the REASON must be entered when prompted and recorded in the Remarks section of FDLE/ATP Form 40 Agency Inspection Report – Intoxilyzer 8000. The REASON for repeating the following test(s); OR the Responsible Cause and Corrective Action Taken on the following test(s) was not recorded: Alcohol Free Subject Test Routh Alcohol Test Routh Alcohol Free Test Routh Interferent Detect Test 0.05 g/210L Test 0.08 g/210L Test
	FDLE/ATP Form 39 states in part, "If the instrument does not comply with the requirements of Chapter 11D-8, FAC, remove the instrument from service and notify the Department Inspector." The Department Inspector was not notified. However, the issue was satisfactorily corrected and the repeated Agency Inspection complies with the requirements of Chapter 11D-8, FAC. The Department Inspector was not notified. However, the repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was correctly removed from evidentiary use. The Department Inspector was not notified. The repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was not removed from evidentiary use.
	The Agency Inspection is noted as "Complies" when it does not comply with the requirements of Chapter 11D-8, FAC.
	Other:
OTH	HER ELECTRONIC DATA REVIEW
	Login Records Comments:
	Cylinder Change Records Date:
	Control Test Records Date:
	Diagnostic Check Records Date:
COI	RRECTIVE ACTION Record hand-written amendments on the FDLE/ATP Form 40, Agency Inspection Report, initial and date the amendments, mark the
	report "AMENDED", and forward a copy to the Department Inspector by 8/20/2020 (Date). Provide a written explanation regarding the referenced item(s) to the Department Inspector by (Date). Upload the Agency Inspection(s). Remove the instrument from evidentiary use until otherwise directed by the Department. No action required Other:
	David Eliezer Reyes Rivera Date: 2020.07.15 12:01:59 -04'00' Signature of Alcohol Testing Program Staff Member Date: 2020.07.15 12:01:59 -04'00' Date: 2020.07.15 12:01:59 -04'00' Date: 2020.07.15 12:01:59 -04'00'