

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HILLSBOROUGH CO SO
Time of Inspection: 18:43

Date of Inspection: 07/08/2020

Serial Number: 80-003388
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:22419080A Exp: 10/05/2021
0.000	0.048	0.079	0.199	0.079
0.000	0.049	0.079	0.200	0.080
0.000	0.049	0.078	0.200	0.080

Number of Simulators Used: 5

Remarks:

AI J. DAIL #254236 PRESENT FOR INSPECTION.

J. Dail 254236

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

[Signature]

Signature and Printed Name

ROGER G SKIPPER

07/08/2020
Date

*Reviewed
7-8-20
1/5/21.21. Bleep
4/12*