

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HENDRY COUNTY SO
Time of Inspection: 19:30

Date of Inspection: 08/31/2020

Serial Number: 80-006234
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021 08/07/2021	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1012059 Exp: 09/05/2021
0.000	0.052	0.083	0.205	0.080
0.000	0.052	0.083	0.205	0.080
0.000	0.052	0.082	0.205	0.080

Number of Simulators Used: 5

Remarks: **Amended**
[Signature] 517

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

[Signature] 517
Signature and Printed Name

JAMES R HESTER

08/31/2020
Date



Florida Department of Law Enforcement

AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Agency: Hendry County Sheriff's Office

Instrument Serial Number: 80-006234

AGENCY INSPECTION DATA REVIEW		
Agency Inspector: James Hester	Date of Inspection: 08/31/2020	Time of Inspection: 19:30:21
Agency Inspection Discrepancy: <input type="checkbox"/> Incomplete <input type="checkbox"/> Untimely/Not Received <input checked="" type="checkbox"/> Erroneous Information <input type="checkbox"/> Procedural <input type="checkbox"/> Other _____		
<input type="checkbox"/> Agency Inspection Not Conducted or Records regarding Agency Inspection have not been uploaded.		
<input type="checkbox"/> Lot Number <input checked="" type="checkbox"/> Expiration Date for <u>0.08 g/ 210L</u> <input checked="" type="checkbox"/> Alcohol Reference Solution <input type="checkbox"/> Dry Gas Standard is <input checked="" type="checkbox"/> Incorrect <input type="checkbox"/> Expired.		
<input type="checkbox"/> FDLE/ATP Form 39 states in part, "If a test must be repeated, the REASON must be entered when prompted and recorded in the Remarks section of FDLE/ATP Form 40 Agency Inspection Report – Intoxilyzer 8000. The <input type="checkbox"/> REASON for repeating the following test(s); OR the <input type="checkbox"/> Possible Cause and Corrective Action Taken on the following test(s) was not recorded: <input type="checkbox"/> Alcohol Free Subject Test <input type="checkbox"/> Mouth Alcohol Test <input type="checkbox"/> Alcohol Free Test <input type="checkbox"/> Interferent Detect Test <input type="checkbox"/> 0.05 g/210L Test <input type="checkbox"/> 0.08 g/210L Test <input type="checkbox"/> 0.20 g/210L Test <input type="checkbox"/> 0.08 g/210L Dry Gas Standard Test		
<input type="checkbox"/> FDLE/ATP Form 39 states in part, "If the instrument does not comply with the requirements of Chapter 11D-8, FAC, remove the instrument from service and notify the Department Inspector." <input type="checkbox"/> The Department Inspector was not notified. However, the issue was satisfactorily corrected and the repeated Agency Inspection complies with the requirements of Chapter 11D-8, FAC. <input type="checkbox"/> The Department Inspector was not notified. However, the repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was correctly removed from evidentiary use. <input type="checkbox"/> The Department Inspector was not notified. The repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was not removed from evidentiary use.		
<input type="checkbox"/> The Agency Inspection is noted as "Complies" when it does not comply with the requirements of Chapter 11D-8, FAC.		
<input type="checkbox"/> Other: _____		

OTHER ELECTRONIC DATA REVIEW	
<input type="checkbox"/> Login Records Date: _____	Comments: The expiration date (year) for the 0.08 g/210 L Alcohol Reference Solution is incorrect. Please see below for required corrective action and return to the Department Inspector by <u>10/01/2020</u> .
<input type="checkbox"/> Cylinder Change Records Date: _____	
<input type="checkbox"/> Control Test Records Date: _____	
<input type="checkbox"/> Diagnostic Check Records Date: _____	

CORRECTIVE ACTION
<input checked="" type="checkbox"/> Record hand-written amendments on the FDLE/ATP Form 40, Agency Inspection Report, initial and date the amendments, mark the report "AMENDED", and forward a copy to the Department Inspector by <u>10/01/2020</u> (Date).
<input type="checkbox"/> Provide a written explanation regarding the referenced item(s) to the Department Inspector by _____ (Date).
<input type="checkbox"/> Upload the Agency Inspection(s).
<input type="checkbox"/> Remove the instrument from evidentiary use until otherwise directed by the Department.
<input type="checkbox"/> No action required
<input type="checkbox"/> Other: _____

Signature of Alcohol Testing Program Staff Member

9/1/2020
Date



AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Agency: Hendry County Sheriff's Office

Instrument Serial Number: 80-006234

AGENCY INSPECTION DATA REVIEW
Agency Inspector: James Hester
Date of Inspection: 06/23/2020
Time of Inspection: 15:17:35
Agency Inspection Discrepancy: [] Incomplete [] Untimely/Not Received [x] Erroneous Information
[] Agency Inspection Not Conducted or Records regarding Agency Inspection have not been uploaded.
[x] Lot Number [] Expiration Date for 0.05 g/ 210L [x] Alcohol Reference Solution [] Dry Gas Standard is [x] Incorrect [] Expired.
[] FDLE/ATP Form 39 states in part, "If a test must be repeated, the REASON must be entered when prompted and recorded in the Remarks section of FDLE/ATP Form 40 Agency Inspection Report - Intoxilyzer 8000. The [] REASON for repeating the following test(s); OR the [] Possible Cause and Corrective Action Taken on the following test(s) was not recorded:
[] Alcohol Free Subject Test [] Mouth Alcohol Test [] Alcohol Free Test [] Interferent Detect Test
[] 0.05 g/210L Test [] 0.08 g/210L Test [] 0.20 g/210L Test [] 0.08 g/210L Dry Gas Standard Test
[] FDLE/ATP Form 39 states in part, "If the instrument does not comply with the requirements of Chapter 11D-8, FAC, remove the instrument from service and notify the Department Inspector."
[] The Department Inspector was not notified. However, the issue was satisfactorily corrected and the repeated Agency Inspection complies with the requirements of Chapter 11D-8, FAC.
[] The Department Inspector was not notified. However, the repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was correctly removed from evidentiary use.
[] The Department Inspector was not notified. The repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was not removed from evidentiary use.
[] The Agency Inspection is noted as "Complies" when it does not comply with the requirements of Chapter 11D-8, FAC.
[] Other: _____

OTHER ELECTRONIC DATA REVIEW
[] Login Records Date: _____
[] Cylinder Change Records Date: _____
[] Control Test Records Date: _____
[] Diagnostic Check Records Date: _____
Comments:
The lot number for the 0.05 g/210 L alcohol reference solution was mistyped. Please see below for the required corrective action.

CORRECTIVE ACTION
[x] Record hand-written amendments on the FDLE/ATP Form 40, Agency Inspection Report, initial and date the amendments, mark the report "AMENDED", and forward a copy to the Department Inspector by 08/15/2020 (Date).
[] Provide a written explanation regarding the referenced item(s) to the Department Inspector by _____ (Date).
[] Upload the Agency Inspection(s).
[] Remove the instrument from evidentiary use until otherwise directed by the Department.
[] No action required
[] Other: _____

Signature of Alcohol Testing Program Staff Member

7/15/2020
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HENDRY COUNTY SC
Time of Inspection: 19:12

Date of Inspection: 03/12/2020

Serial Number: 80-006234
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201311C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1012059 Exp: 09/05/2021
0.000	0.047	0.074 / 0.076	0.197	0.080
0.000	0.047	0.075 / 0.076	0.198	0.081
0.000	0.048	0.075 / 0.076	0.199	0.080


Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



Signature and Printed Name JAMES R. HESTER

03/12/2020
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HENDRY COUNTY SO
Time of Inspection: 16:56

Date of Inspection: 03/12/2020

Serial Number: 80-006234
Software: 8100.27

Check or Test	YES	NO
Data and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#: 201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000	0.047	0.068 / 0.072		
0.000	0.047	0.069 / 0.072		
0.000	0.047	0.071 / 0.073		

Number of Simulators Used: 5

Remarks:

CC: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: 0.080 TEST NOT WITHIN TOL.

A/C was very cold in Room. Allowed Simulator to heat for longer before trying new inspection. I Also found a leak in simulator, jar was cracked. Simulator Replaced. Spoke with Department Inspector who said start New inspection.

JR 517

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JR 517

Signature and Printed Name

JAMES R HESTER

03/12/2020
Date